



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB123827

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	01/27/2023	09:50
2	O1320-01		1	12	100.00	>212.0	>212.0	01/27/2023	10:20
3	O1321-01		1	14	100.00	>212.0	>212.0	01/27/2023	10:50
4	O1321-02		1	14	100.00	>212.0	>212.0	01/27/2023	11:20
5	O1327-01		1	13	100.00	>212.0	>212.0	01/27/2023	11:50
6	O1327-01DUP		1	14	100.00	>212.0	>212.0	01/27/2023	12:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16123827

WorkList Name : fp1-27

WorkList ID : 166903

Department : Wet-Chemistry

Date : 01-27-2023 08:26:46

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/31/2023	Water	O1320-01	Flash Point	Cool 4 deg C	PSEG03	D12	TOTE-COMP-0126	01/26/2023	1010B
02/02/2023	Water	O1321-01	Flash Point	Cool 4 deg C	PSEG03	D12	33902	01/26/2023	1010B
02/02/2023	Water	O1321-02	Flash Point	Cool 4 deg C	PSEG03	D12	33903	01/26/2023	1010B
02/02/2023	Water	O1327-01	Flash Point	Cool 4 deg C	PSEG03	D12	N48965	01/26/2023	1010B

Date/Time 01/27/2023 09:30
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: JACOB

Date/Time 01/27/2023 13:00
 Raw Sample Received by: JACOB
 Raw Sample Relinquished by: RM CWC