

Clear table


Instrument ID: IC-2										Analyst : IZ		Method: 300.0 / 9056A	
ident	concentra tion F-	concentratio n CL-	concentratio on NO2	concentratio on BR-	concentratio on NO3	concentratio on HPO4	concentratio on SO4	file name	date time	Initial wt/ Final Vol	Analyst		
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000_2023-01-19_	1/19/23 11:00		IZ/MA		
STD2	0.4880	0.6840	0.7060	2.3300	0.5760	1.0560	3.4740	2023-01-19_	1/19/23 11:31		IZ/MA		
STD3	0.7760	1.2150	1.2130	4.0420	1.0070	2.0000	6.0020	2023-01-19_	1/19/23 12:01		IZ/MA		
STD4	0.9450	1.5030	1.5180	5.0520	1.2780	2.5430	7.4510	2023-01-19_	1/19/23 12:32		IZ/MA		
STD5	1.8550	2.8760	2.8370	9.5070	2.3720	4.8860	14.5380	2023-01-19_	1/19/23 13:02		IZ/MA		
STD6	3.9270	5.9350	5.9150	19.6950	4.9230	9.9270	29.5160	2023-01-19_	1/19/23 13:33		IZ/MA		
STD7	5.0990	7.5870	7.6110	25.3740	6.3440	12.5880	38.0180	2023-01-19_	1/19/23 14:04		IZ/MA		
ICV	1.8290	2.9990	3.0090	9.9010	2.4740	4.7390	14.6300	2023-01-19_	1/19/23 15:07		IZ/MA		
ICB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2023-01-19_	1/19/23 15:38		IZ/MA		
CCV	1.8670	2.8420	2.8460	9.4520	2.3280	5.4670	13.9380	2023-01-26_	1/26/23 14:30		IZ/MA		
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2023-01-26_	1/26/23 15:02		IZ/MA		
LB123829BLW	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2023-01-26_	1/26/23 15:32		IZ/MA		
LB123829BSW	1.8220	2.8160	2.8210	9.1480	2.3230	5.0180	13.8790	2023-01-26_	1/26/23 16:03		IZ/MA		
O1323-01	0.4100	244.2150	0.0000	0.0000	0.1940	0.5580	22.8130	2023-01-26_	1/26/23 16:34		IZ/MA		
O1323-03	0.5850	119.0520	0.0000	0.0000	0.5690	0.0000	17.8690	2023-01-26_	1/26/23 17:05		IZ/MA		
O1323-05	34.6110	330.4650	0.0000	0.0000	0.0000	0.0000	13.7400	2023-01-26_	1/26/23 17:36		IZ/MA		
O1323-07	32.3750	327.0310	0.0000	0.0000	0.0000	0.0000	11.9630	2023-01-26_	1/26/23 18:06		IZ/MA		
O1323-09	0.4400	193.8190	0.0000	0.0000	0.8310	0.6290	27.2640	2023-01-26_	1/26/23 18:37		IZ/MA		
O1320-01	1.1550	273.9570	0.4470	0.0000	1.5930	0.0000	81.4100	2023-01-26_	1/26/23 19:07		IZ/MA		
O1327-01	0.5370	258.0760	0.0000	1.5110	1.1350	0.0000	112.1560	2023-01-26_	1/26/23 19:38		IZ/MA		
CCV	1.8970	2.8780	2.9200	9.5860	2.3590	6.6210	14.2170	2023-01-26_	1/26/23 20:39		IZ/MA		
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2023-01-26_	1/26/23 21:10		IZ/MA		
O1323-11	0.0000	457.0520	0.0000	0.0000	0.2090	0.0000	26.8430	2023-01-26_	1/26/23 21:41		IZ/MA		
O1323-13	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2023-01-26_	1/26/23 22:12		IZ/MA		
O1323-13MS	2.0350	3.2190	3.0910	10.0100	2.5180	6.9050	14.8420	2023-01-26_	1/26/23 22:42		IZ/MA		
O1323-13MSD	1.8940	2.9290	2.8820	9.4940	2.3680	6.6100	14.2310	2023-01-26_	1/26/23 23:13		IZ/MA		
CCV	1.9230	2.9040	2.9180	9.5990	2.3720	6.6710	14.2100	2023-01-26_	1/26/23 23:44		IZ/MA		
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2023-01-27_	1/27/23 0:14		IZ/MA		

Clear table

Instrument ID: IC-2									
Analyst : IZ				Method:		300.0 / 9056A			
ident	concentra tion F-	concentratio n CL-	concentratio on NO2	concentratio on BR-	concentratio on NO3	concentratio on HPO4	concentratio on SO4	file name	date time
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2023-01-19_	1/19/23 11:00
STD2	0.4880	0.6840	0.7060	2.3300	0.5760	1.0560	3.4740	2023-01-19_	1/19/23 11:31
STD3	0.7760	1.2150	1.2130	4.0420	1.0070	2.0000	6.0020	2023-01-19_	1/19/23 12:01
STD4	0.9450	1.5030	1.5180	5.0520	1.2780	2.5430	7.4510	2023-01-19_	1/19/23 12:32
STD5	1.8550	2.8760	2.8370	9.5070	2.3720	4.8860	14.5380	2023-01-19_	1/19/23 13:02
STD6	3.9270	5.9350	5.9150	19.6950	4.9230	9.9270	29.5160	2023-01-19_	1/19/23 13:33
STD7	5.0990	7.5870	7.6110	25.3740	6.3440	12.5880	38.0180	2023-01-19_	1/19/23 14:04
ident	True Value F-	True Value CL-	True Value NO2	True Value BR-	True Value NO3	True Value HPO4	True Value SO4		
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000		
STD2	0.4000	0.6000	0.6000	2.0000	0.5000	1.0000	3.0000		
STD3	0.8000	1.2000	1.2000	4.0000	1.0000	2.0000	6.0000		
STD4	1.0000	1.5000	1.5000	5.0000	1.2500	2.5000	7.5000		
STD5	2.0000	3.0000	3.0000	10.0000	2.5000	5.0000	15.0000		
STD6	4.0000	6.0000	6.0000	20.0000	5.0000	10.0000	30.0000		
STD7	5.0000	7.5000	7.5000	25.0000	6.2500	12.5000	37.0000		
ident	Relative Error F-	Relative Error CL-	Relative Error NO2	Relative Error BR-	Relative Error NO3	Relative Error HPO4	Relative Error SO4	Initial wt/ Final Vol	Analyst
STD1	22.0000	14.0000	17.6667	16.5000	15.2000	5.6000	15.8000		IZ/MA
STD2	-3.0000	1.2500	1.0833	1.0500	0.7000	0.0000	0.0333		
STD3	-5.5000	0.2000	1.2000	1.0400	2.2400	1.7200	-0.6533		
STD4	-7.2500	-4.1333	-5.4333	-4.9300	-5.1200	-2.2800	-3.0800		
STD5	-1.8250	-1.0833	-1.4167	-1.5250	-1.5400	-0.7300	-1.6133		
STD6	1.9800	1.1600	1.4800	1.4960	1.5040	0.7040	2.7514		
STD7									

CALIBRATION OF COMPONENT F-

Method: AnionsIC2-011923.mtw
 Equation: $Q = 0.0718827 \cdot A + 3.79033$
 RSD: 5.705 %
 Correlation coefficient: 0.998224




K3 = 0 K2 = 0 K1 = 0.0718827 K0 = 3.79033
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	1.665	101.3	0.4	20	5.877		
3	3.591	175.3	0.8	20	5.877		
4	5.046	210.3	1	20	5.877		
5	10.25	463.5	2	20	5.877		
6	24.4	1040	4	20	5.877		
7	33.78	1366	5	20	5.877		

CALIBRATION OF COMPONENT CL-

Method: AnionsIC2-011923.mtw
 Equation: $Q = 0.122731 \cdot A + 3.35655$
 RSD: 2.808 %
 Correlation coefficient: 0.999570




K3 = 0 K2 = 0 K1 = 0.122731 K0 = 3.35655
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	5.645	84.08	0.6	20	9.035		
3	11.46	170.6	1.2	20	9.035		
4	14.93	217.5	1.5	20	9.035		
5	30.2	441.3	3	20	9.035		
6	65.55	939.9	6	20	9.035		
7	84.31	1209	7.5	20	9.035		

5?□

CALIBRATION OF COMPONENT NO2

Method: AnionsIC2-011923.mtw
 Equation: $Q = 0.0538346 \cdot A + 5.16968$
 RSD: 3.640 %
 Correlation coefficient: 0.999277




K3 = 0 K2 = 0 K1 = 0.0538346 K0 = 5.16968
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	8.896	166.1	0.6	20	10.93		
3	18.77	354.5	1.2	20	10.93		
4	24.82	468.1	1.5	20	10.93		
5	50.96	958	3	20	10.93		
6	109.4	2102	6	20	10.93		
7	138.2	2732	7.5	20	10.93		

CALIBRATION OF COMPONENT BR-

Method: AnionsIC2-011923.mtw
 Equation: $Q = 0.271166 \cdot A + 15.2217$
 RSD: 3.488 %
 Correlation coefficient: 0.999337




K3 = 0 K2 = 0 K1 = 0.271166 K0 = 15.2217
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	6.174	115.7	2	20	13.41		
3	13.1	242	4	20	13.41		
4	17.12	316.5	5	20	13.41		
5	35.22	645.1	10	20	13.41		
6	78.35	1397	20	20	13.41		
7	102.9	1815	25	20	13.41		

±□□□

CALIBRATION OF COMPONENT NO3

Method: AnionsIC2-011923.mtw
 Equation: $Q = 0.0475204 \cdot A + 3.28798$
 RSD: 3.530 %
 Correlation coefficient: 0.999321




K3 = 0 K2 = 0 K1 = 0.0475204 K0 = 3.28798
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	7.292	173.3	0.5	20	16.64		
3	15.21	354.6	1	20	16.64		
4	19.8	468.5	1.25	20	16.64		
5	40.31	929.2	2.5	20	16.64		
6	88.71	2003	5	20	16.64		
7	117	2601	6.25	20	16.64		

CALIBRATION OF COMPONENT HPO4

Method: AnionsIC2-011923.mtw
 Equation: $Q = 0.0494757 \cdot A + 8.69571$
 RSD: 1.605 %
 Correlation coefficient: 0.999860




K3 = 0 K2 = 0 K1 = 0.0494757 K0 = 8.69571
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	5.091	251.2	1	20	21.97		
3	13.33	632.6	2	20	21.97		
4	18.27	852.2	2.5	20	21.97		
5	40.12	1799	5	20	21.97		
6	92.02	3837	10	20	21.97		
7	120.8	4913	12.5	20	21.97		

±□;□

CALIBRATION OF COMPONENT SO4

Method: AnionsIC2-011923.mtw
 Equation: $Q = 0.154808 \cdot A + 18.7638$
 RSD: 2.942 %
 Correlation coefficient: 0.999528



K3 = 0 K2 = 0 K1 = 0.154808 K0 = 18.7638
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	10.52	327.6	3	20	25.5		
3	21.61	654.2	6	20	25.5		
4	27.95	841.5	7.5	20	25.5		
5	56.94	1757	15	20	25.5		
6	126.6	3692	30	20	25.5		
7	165.8	4790	37.5	20	25.5		

_?□

Report date: 1/19/2023 4:18:28 PM
Printed by: wet


Ident: STD1
Analysis from: 1/19/2023 11:00:25 AM
File: _2023-01-19_

Last save: 1/19/2023 2:36:27 PM

Method: AnionsIC2-011923.mtw
Run operator: wet
Analysis number: 33045

Last save: 1/19/2023 2:01:

SAMPLE:
: IZ/MA
Vial number: 2
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 1/19/2023 4:19:28 PM
Printed by: wet


Ident: STD2
Analysis from: 1/19/2023 11:31:03 AM
File: _2023-01-19_

Last save: 1/19/2023 2:36:27 PM

Method: AnionsIC2-011923.mtw
Run operator: wet
Analysis number: 33046

Last save: 1/19/2023 2:01:

SAMPLE:
: IZ/MA
Vial number: 3
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.12	0.652	1.54	3.40	82.927	6.91
2	9.19	0.212	5.64	12.50	84.078	7.00
3	11.11	0.258	8.89	19.70	166.075	13.83
4	14.37	0.282	6.18	13.67	115.707	9.64
5	16.43	0.348	7.29	16.15	173.319	14.43
6	22.44	0.635	5.09	11.27	251.163	20.92
7	25.64	0.465	10.52	23.30	327.566	27.28
7	28.00	0.407	45.16	99.99	1200.836	100.00

This report has been created by IC Net
METROHM LTD

Report date: 1/19/2023 4:19:36 PM
Printed by: wet


Ident: STD3
Analysis from: 1/19/2023 12:01:40 PM
File: _2023-01-19_

Last save: 1/19/2023 2:36:27 PM

Method: AnionsIC2-011923.mtw
Run operator: wet
Analysis number: 33047

Last save: 1/19/2023 2:01:

SAMPLE:
: IZ/MA
Vial number: 4
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.02	0.584	3.50	3.61	163.093	6.34
2	9.07	0.212	11.46	11.82	170.642	6.64
3	10.96	0.259	18.77	19.35	354.456	13.78
4	14.18	0.276	13.10	13.50	241.952	9.41
5	16.19	0.342	15.21	15.68	354.642	13.79
6	22.24	0.609	13.33	13.74	632.560	24.60
7	25.43	0.455	21.61	22.28	654.237	25.44
7	28.00	0.391	96.99	100.00	2571.582	100.00

This report has been created by IC Net
METROHM LTD

Report date: 1/19/2023 4:19:44 PM
Printed by: wet


Ident: STD4
Analysis from: 1/19/2023 12:32:18 PM
File: _2023-01-19_

Last save: 1/19/2023 2:36:27 PM

Method: AnionsIC2-011923.mtw
Run operator: wet
Analysis number: 33048

Last save: 1/19/2023 2:01:

SAMPLE:
: IZ/MA
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.00	0.527	5.05	3.94	210.322	6.23
2	9.06	0.209	14.93	11.67	217.524	6.45
3	10.95	0.259	24.82	19.40	468.089	13.87
4	14.16	0.275	17.12	13.38	316.508	9.38
5	16.18	0.348	19.80	15.47	468.497	13.88
6	22.29	0.598	18.28	14.28	852.213	25.25
7	25.49	0.453	27.95	21.85	841.462	24.94
7	28.00	0.381	127.94	100.00	3374.616	100.00

This report has been created by IC Net
METROHM LTD

Report date: 1/19/2023 4:19:51 PM
Printed by: wet


Ident: STD5
Analysis from: 1/19/2023 1:02:56 PM
File: _2023-01-19_

Last save: 1/19/2023 2:36:27 PM

Method: AnionsIC2-011923.mtw
Run operator: wet
Analysis number: 33049

Last save: 1/19/2023 2:01:

SAMPLE:
: IZ/MA
Vial number: 6
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.07	0.593	10.25	3.88	463.485	6.63
2	9.11	0.208	30.20	11.44	441.349	6.31
3	11.00	0.259	50.95	19.30	957.967	13.70
4	14.23	0.271	35.22	13.34	645.058	9.22
5	16.21	0.337	40.31	15.27	929.238	13.29
6	22.34	0.572	40.12	15.20	1799.193	25.73
7	25.68	0.464	56.94	21.57	1757.051	25.12
7	28.00	0.386	264.00	100.00	6993.341	100.00

This report has been created by IC Net
METROHM LTD

Report date: 1/19/2023 4:19:58 PM
Printed by: wet


Ident: STD6
Analysis from: 1/19/2023 1:33:34 PM
File: _2023-01-19_

Last save: 1/19/2023 2:36:27 PM

Method: AnionsIC2-011923.mtw
Run operator: wet
Analysis number: 33050

Last save: 1/19/2023 2:01:

SAMPLE:
: IZ/MA
Vial number: 7
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.03	0.566	24.40	4.17	1039.904	6.93
2	9.05	0.203	65.55	11.20	939.853	6.26
3	10.91	0.264	109.43	18.70	2101.519	14.00
4	14.14	0.262	78.35	13.39	1396.507	9.30
5	16.06	0.329	88.71	15.16	2002.740	13.34
6	22.16	0.538	92.02	15.73	3837.261	25.57
7	25.41	0.438	126.62	21.64	3692.017	24.60
7	28.00	0.372	585.07	100.00	15009.802	100.00

This report has been created by IC Net
METROHM LTD

Report date: 1/19/2023 4:20:05 PM
Printed by: wet


Ident: STD7
Analysis from: 1/19/2023 2:04:11 PM
File: _2023-01-19_

Last save: 1/19/2023 2:36:27 PM

Method: AnionsIC2-011923.mtw
Run operator: wet
Analysis number: 33051

Last save: 1/19/2023 2:01:

SAMPLE:
: IZ/MA
Vial number: 8
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.08	0.530	33.78	4.43	1366.005	7.03
2	9.11	0.207	84.31	11.05	1209.022	6.22
3	10.98	0.277	138.16	18.11	2731.577	14.06
4	14.22	0.259	102.91	13.49	1815.329	9.34
5	16.12	0.325	117.04	15.34	2600.803	13.39
6	22.26	0.526	120.80	15.84	4912.956	25.29
7	25.54	0.435	165.82	21.74	4790.440	24.66
7	28.00	0.366	762.82	100.00	19426.131	100.00

This report has been created by IC Net
METROHM LTD

Report date: 1/19/2023 4:20:12 PM
Printed by: wet


Ident: ICV
Analysis from: 1/19/2023 3:07:25 PM
File: _2023-01-19_

Last save: 1/19/2023 3:35:25 PM

Method: AnionsIC2-011923.mtw
Run operator: wet
Analysis number: 33053

Last save: 1/19/2023 2:32:

SAMPLE:
: IZ/MA
Vial number: 15
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.02	0.556	10.63	3.91	456.108	6.43
2	9.06	0.211	31.17	11.45	461.385	6.50
3	10.94	0.262	53.72	19.74	1021.822	14.40
4	14.16	0.273	36.51	13.41	674.090	9.50
5	16.14	0.341	41.74	15.34	971.854	13.70
6	22.23	0.572	39.19	14.40	1739.894	24.53
7	25.49	0.449	59.20	21.75	1768.900	24.93
7	28.00	0.380	272.17	100.00	7094.053	100.00

This report has been created by IC Net
METROHM LTD

Report date: 1/19/2023 4:20:20 PM
Printed by: wet


Ident: ICB
Analysis from: 1/19/2023 3:38:45 PM
File: _2023-01-19_

Last save: 1/19/2023 4:06:45 PM

Method: AnionsIC2-011923.mtw
Run operator: wet
Analysis number: 33054

Last save: 1/19/2023 2:32:

SAMPLE:
: IZ/MA
Vial number: 10
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 1/27/2023 10:44:52 AM
Printed by: wet


Ident: CCV
Analysis from: 1/26/2023 2:30:56 PM
File: _2023-01-26_

Last save: 1/26/2023 2:58:56 PM

Method: AnionsIC2-011923.mtw
Run operator: wet
Analysis number: 33126

Last save: 1/19/2023 2:32:

SAMPLE:
: IZ/MA
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.08	0.572	10.53	4.04	466.822	6.55
2	9.12	0.211	29.47	11.31	435.740	6.11
3	10.97	0.267	49.37	18.94	961.377	13.49
4	14.12	0.281	33.76	12.95	640.998	8.99
5	16.09	0.340	39.36	15.10	910.591	12.77
6	22.45	0.578	44.45	17.05	2034.032	28.53
7	26.21	0.470	53.70	20.60	1679.483	23.56
7	28.00	0.389	260.64	100.00	7129.043	100.00

This report has been created by IC Net
METROHM LTD

Report date: 1/27/2023 10:45:03 AM
Printed by: wet


Ident: CCB
Analysis from: 1/26/2023 3:02:14 PM
File: _2023-01-26_

Last save: 1/26/2023 3:56:41 PM

Method: AnionsIC2-011923.mtw
Run operator: wet
Analysis number: 33127

Last save: 1/19/2023 2:32:

SAMPLE:
: IZ/MA
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 1/27/2023 10:45:23 AM
Printed by: wet


Ident: LB123829BLW
Analysis from: 1/26/2023 3:32:52 PM
File: _2023-01-26_

Last save: 1/27/2023 10:45:19 AM

Method: AnionsIC2-011923.mtw
Run operator: wet
Analysis number: 33128

Last save: 1/19/2023 2:32:

SAMPLE:
: IZ/MA
Vial number: 13
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 1/27/2023 10:45:33 AM
Printed by: wet


Ident: LB123829BSW
Analysis from: 1/26/2023 4:03:30 PM
File: _2023-01-26_

Last save: 1/27/2023 10:44:42 AM

Method: AnionsIC2-011923.mtw
Run operator: wet
Analysis number: 33129

Last save: 1/19/2023 2:32:

SAMPLE:
: IZ/MA
Vial number: 14
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.06	0.553	10.61	4.11	454.167	6.59
2	9.07	0.212	29.50	11.42	431.579	6.26
3	10.92	0.270	49.15	19.03	951.889	13.82
4	14.05	0.274	33.95	13.15	618.572	8.98
5	16.01	0.335	40.05	15.51	908.334	13.19
6	22.42	0.565	41.20	15.95	1852.571	26.89
7	26.19	0.468	53.79	20.83	1671.812	24.27
7	28.00	0.382	258.26	100.00	6888.924	100.00

This report has been created by IC Net
METROHM LTD

Report date: 1/27/2023 10:45:55 AM
Printed by: wet


Ident: 01323-01
Analysis from: 1/26/2023 4:34:46 PM
File: _2023-01-26_

Last save: 1/27/2023 10:45:50 AM

Method: AnionsIC2-011923.mtw
Run operator: wet
Analysis number: 33130

Last save: 1/19/2023 2:32:

SAMPLE:
: IZ/MA
Vial number: 31
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.23	0.952	0.91	0.06	61.429	0.14
2	9.32	0.408	1550.44	94.33	39769.417	93.09
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	16.38	0.341	0.55	0.03	12.482	0.03
6	22.77	0.837	0.79	0.05	49.916	0.12
7	26.47	0.467	90.82	5.53	2826.038	6.62
7	28.00	0.429	1643.51	100.00	42719.283	100.00

This report has been created by IC Net
METROHM LTD

Report date: 1/27/2023 10:46:11 AM
Printed by: wet


Ident: 01323-03
Analysis from: 1/26/2023 5:05:24 PM
File: _2023-01-26_

Last save: 1/27/2023 10:46:06 AM

Method: AnionsIC2-011923.mtw
Run operator: wet
Analysis number: 33131

Last save: 1/19/2023 2:32:

SAMPLE:
: IZ/MA
Vial number: 32
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.19	0.899	1.66	0.13	110.091	0.50
2	9.14	0.228	1242.50	94.01	19373.133	88.70
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	16.19	0.333	7.56	0.57	170.331	0.78
6	0.00	0.000	0.00	0.00	0.000	0.00
7	26.34	0.469	69.92	5.29	2187.330	10.01
7	28.00	0.276	1321.64	100.00	21840.886	100.00

This report has been created by IC Net
METROHM LTD

Report date: 1/27/2023 10:48:08 AM
Printed by: wet


Ident: 01323-05
Analysis from: 1/26/2023 5:36:02 PM
File: _2023-01-26_

Last save: 1/27/2023 10:48:06 AM

Method: AnionsIC2-011923.mtw
Run operator: wet
Analysis number: 33132

Last save: 1/19/2023 2:32:

SAMPLE:
: IZ/MA
Vial number: 33
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.90	0.780	217.01	11.57	9577.101	14.68
2	9.37	0.532	1599.86	85.30	53824.457	82.49
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	26.43	0.482	51.03	2.72	1653.835	2.53
7	28.00	0.256	1867.89	99.59	65055.393	99.71

This report has been created by IC Net
METROHM LTD

Report date: 1/27/2023 10:48:55 AM
Printed by: wet


Ident: 01323-07
Analysis from: 1/26/2023 6:06:40 PM
File: _2023-01-26_

Last save: 1/27/2023 10:48:49 AM

Method: AnionsIC2-011923.mtw
Run operator: wet
Analysis number: 33133

Last save: 1/19/2023 2:32:

SAMPLE:
: IZ/MA
Vial number: 34
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.89	0.770	206.07	11.12	8955.135	14.03
2	9.37	0.522	1596.55	86.14	53264.870	83.46
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	26.46	0.487	43.52	2.35	1424.275	2.23
7	28.00	0.254	1846.14	99.61	63644.281	99.72

This report has been created by IC Net
METROHM LTD

Report date: 1/27/2023 10:49:18 AM
Printed by: wet


Ident: O1323-09
Analysis from: 1/26/2023 6:37:17 PM
File: _2023-01-26_

Last save: 1/27/2023 10:49:17 AM

Method: AnionsIC2-011923.mtw
Run operator: wet
Analysis number: 33134

Last save: 1/19/2023 2:32:

SAMPLE:
: IZ/MA
Vial number: 35
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.21	0.832	1.24	0.08	69.636	0.20
2	9.20	0.324	1487.77	92.29	31556.928	89.18
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	16.16	0.328	12.72	0.79	280.477	0.79
6	22.61	1.790	0.73	0.05	78.677	0.22
7	26.33	0.464	109.69	6.80	3401.110	9.61
7	28.00	0.534	1612.15	100.00	35386.829	100.00

This report has been created by IC Net
METROHM LTD

Report date: 1/27/2023 10:49:39 AM
Printed by: wet


Ident: O1320-01
Analysis from: 1/26/2023 7:07:57 PM
File: _2023-01-26_

Last save: 1/26/2023 7:35:58 PM

Method: AnionsIC2-011923.mtw
Run operator: wet
Analysis number: 33135

Last save: 1/19/2023 2:32:

SAMPLE:
: IZ/MA
Vial number: 40
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.12	0.699	5.43	0.28	268.699	0.48
2	9.30	0.458	1579.26	80.86	44616.033	79.74
3	10.94	0.262	3.88	0.20	70.193	0.13
4	0.00	0.000	0.00	0.00	0.000	0.00
5	16.10	0.326	27.30	1.40	601.444	1.07
6	0.00	0.000	0.00	0.00	0.000	0.00
7	26.26	0.467	337.03	17.26	10396.260	18.58
7	28.00	0.316	1952.90	100.00	55952.629	100.00

This report has been created by IC Net
METROHM LTD

Report date: 1/27/2023 10:51:07 AM
Printed by: wet


Ident: 01327-01
Analysis from: 1/26/2023 7:38:35 PM
File: _2023-01-26_

Last save: 1/27/2023 10:51:06 AM

Method: AnionsIC2-011923.mtw
Run operator: wet
Analysis number: 33136

Last save: 1/19/2023 2:32:

SAMPLE:
: IZ/MA
Vial number: 41
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.21	0.953	1.39	0.07	96.750	0.17
2	9.29	0.429	1553.28	76.33	42028.017	73.79
3	0.00	0.000	0.00	0.00	0.000	0.00
4	14.30	0.262	3.22	0.16	55.289	0.10
5	16.19	0.324	18.68	0.92	408.467	0.72
6	0.00	0.000	0.00	0.00	0.000	0.00
7	26.37	0.477	458.28	22.52	14368.521	25.23
7	28.00	0.349	2034.84	100.00	56957.044	100.00

This report has been created by IC Net
METROHM LTD

Report date: 1/27/2023 10:51:40 AM
Printed by: wet


Ident: CCV
Analysis from: 1/26/2023 8:39:59 PM
File: _2023-01-26_

Last save: 1/27/2023 10:51:38 AM

Method: AnionsIC2-011923.mtw
Run operator: wet
Analysis number: 33138

Last save: 1/19/2023 2:32:

SAMPLE:
: IZ/MA
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.05	0.545	11.28	4.09	475.153	6.17
2	9.05	0.207	31.08	11.26	441.694	5.74
3	10.87	0.265	52.86	19.15	988.942	12.85
4	13.96	0.274	35.15	12.74	650.855	8.46
5	15.90	0.330	41.36	14.99	923.753	12.00
6	22.44	0.633	48.96	17.74	2500.669	32.49
7	26.26	0.468	55.27	20.03	1715.567	22.29
7	28.00	0.389	275.96	100.00	7696.633	100.00

This report has been created by IC Net
METROHM LTD

Report date: 1/27/2023 10:51:58 AM
Printed by: wet


Ident: CCB
Analysis from: 1/26/2023 9:10:37 PM
File: _2023-01-26_

Last save: 1/27/2023 10:51:55 AM

Method: AnionsIC2-011923.mtw
Run operator: wet
Analysis number: 33139

Last save: 1/19/2023 2:32:

SAMPLE:
: IZ/MA
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 1/27/2023 10:52:21 AM
Printed by: wet


Ident: 01323-11
Analysis from: 1/26/2023 9:41:24 PM
File: _2023-01-26_

Last save: 1/27/2023 10:52:16 AM

Method: AnionsIC2-011923.mtw
Run operator: wet
Analysis number: 33140

Last save: 1/19/2023 2:32:

SAMPLE:
: IZ/MA
Vial number: 36
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	9.49	0.752	1650.63	93.81	74452.831	95.68
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	16.11	0.326	0.84	0.05	18.604	0.02
6	0.00	0.000	0.00	0.00	0.000	0.00
7	26.26	0.463	108.10	6.14	3346.717	4.30
7	28.00	0.220	1759.57	100.00	77818.152	100.00

This report has been created by IC Net
METROHM LTD

Report date: 1/27/2023 10:53:07 AM
Printed by: wet


Ident: 01323-13
Analysis from: 1/26/2023 10:12:01 PM
File: _2023-01-26_

Last save: 1/27/2023 10:53:01 AM

Method: AnionsIC2-011923.mtw
Run operator: wet
Analysis number: 33141

Last save: 1/19/2023 2:32:

SAMPLE:
:
Vial number: 37
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 1/27/2023 10:53:19 AM
Printed by: wet


Ident: O1323-13MS
Analysis from: 1/26/2023 10:42:39 PM
File: _2023-01-26_

Last save: 1/26/2023 11:10:39 PM

Method: AnionsIC2-011923.mtw
Run operator: wet
Analysis number: 33142

Last save: 1/19/2023 2:32:

SAMPLE:
: IZ/MA
Vial number: 38
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.14	0.612	11.15	3.86	513.591	6.30
2	9.14	0.213	33.88	11.72	497.282	6.10
3	11.00	0.274	54.11	18.72	1052.217	12.91
4	14.18	0.276	36.62	12.67	682.193	8.37
5	16.16	0.339	43.16	14.93	990.595	12.16
6	22.81	0.621	52.28	18.09	2615.456	32.10
7	26.66	0.469	57.86	20.02	1796.218	22.05
7	28.00	0.401	289.05	100.00	8147.552	100.00

This report has been created by IC Net
METROHM LTD

Report date: 1/27/2023 10:53:41 AM
Printed by: wet


Ident: O1323-13MSD
Analysis from: 1/26/2023 11:13:17 PM
File: _2023-01-26_

Last save: 1/27/2023 10:53:38 AM

Method: AnionsIC2-011923.mtw
Run operator: wet
Analysis number: 33143

Last save: 1/19/2023 2:32:

SAMPLE:
: IZ/MA
Vial number: 39
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.07	0.591	10.55	3.82	474.315	6.17
2	9.04	0.207	31.49	11.39	450.021	5.86
3	10.85	0.263	52.23	18.89	974.627	12.68
4	13.96	0.271	35.27	12.76	644.076	8.38
5	15.88	0.328	41.72	15.09	927.408	12.07
6	22.45	0.618	50.11	18.13	2496.090	32.48
7	26.29	0.470	55.08	19.92	1717.326	22.35
7	28.00	0.393	276.46	100.00	7683.864	100.00

This report has been created by IC Net
METROHM LTD

Report date: 1/27/2023 10:55:23 AM
Printed by: wet


Ident: CCV
Analysis from: 1/26/2023 11:44:03 PM
File: _2023-01-26_

Last save: 1/27/2023 10:55:20 AM

Method: AnionsIC2-011923.mtw
Run operator: wet
Analysis number: 33144

Last save: 1/19/2023 2:32:

SAMPLE:
: IZ/MA
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	6.18	0.556	11.24	4.05	482.300	6.24
2	9.17	0.208	31.09	11.20	445.830	5.77
3	11.00	0.266	52.08	18.76	988.050	12.78
4	14.11	0.271	35.64	12.84	651.827	8.43
5	16.04	0.328	41.83	15.07	929.024	12.01
6	22.64	0.624	50.34	18.13	2521.003	32.60
7	26.50	0.467	55.40	19.95	1714.631	22.17
7	28.00	0.389	277.63	100.00	7732.663	100.00

This report has been created by IC Net
METROHM LTD

Report date: 1/27/2023 10:55:49 AM
Printed by: wet


Ident: CCB
Analysis from: 1/27/2023 12:14:42 AM
File: _2023-01-27_

Last save: 1/27/2023 10:55:45 AM

Method: AnionsIC2-011923.mtw
Run operator: wet
Analysis number: 33145

Last save: 1/19/2023 2:32:

SAMPLE:
: IZ/MA
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD