

LB123.830


SFAM01.1

INSTRUMENT ID: 01

Linear

μ Abs.:

47958



A= 0.0000e+000

B= 2.0984e-004 *3 b12*

C= -3.0410e-003 *3.0410e-003*

Rho= 0.9999279

Accept= Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.025	0.025	132	0.000	132					
0.05	0.050					185					
0.20	0.200	0.203	0.003	983	0.0 %	983	0				
2.50	2.500	2.526	0.026	12052	0.0 %	12052					
5.00	5.000	4.930	-0.070	23508	0.0 %	23508					
7.50	7.500	7.456	-0.044	35547	0.0 %	35547					
10.0	10.000	10.060	0.060	47958	0.0 %	47958					

0/0 D
=
2
1
-1
-1
1

LB123830

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	132	-		0 SFAM01.1	PPB	1/27/2023 11:22	Std
	0.2 S01	983	-		0.2 SFAM01.1	PPB	1/27/2023 11:24	Std
	2.5 S02	12052	-		2.5 SFAM01.1	PPB	1/27/2023 11:26	Std
	5 S03	23508	-		5 SFAM01.1	PPB	1/27/2023 11:28	Std
	7.5 S04	35547	-		7.5 SFAM01.1	PPB	1/27/2023 11:31	Std
	10 S05	47958	-		10 SFAM01.1	PPB	1/27/2023 11:33	Std
ICV022	ICV022	18062	3.787 -		SFAM01.1	PPB	1/27/2023 11:36	SMPL
ICB022	ICB022	185	0.0358 -		SFAM01.1	PPB	1/27/2023 11:38	SMPL
CCV034	CCV034	23667	4.9632 -		SFAM01.1	PPB	1/27/2023 11:40	SMPL
CCB034	CCB034	139	0.0261 -		SFAM01.1	PPB	1/27/2023 11:43	SMPL
PB150516BL	PBS516	62	0.01 -		SFAM01.1	PPB	1/27/2023 11:45	SMPL
O1298-01	MCOAQ7	645	0.1323 -		SFAM01.1	PPB	1/27/2023 11:47	SMPL
O1298-02	MCOAQ8	617	0.1264 -		SFAM01.1	PPB	1/27/2023 11:49	SMPL
O1298-03	MCOAQ9	975	0.2016 -		SFAM01.1	PPB	1/27/2023 11:52	SMPL
O1298-04	MCOAR0	7702	1.6131 -		SFAM01.1	PPB	1/27/2023 11:54	SMPL
O1298-05	MCOAR1	1003	0.2074 -		SFAM01.1	PPB	1/27/2023 11:56	SMPL
O1298-06	MCOAR2	713	0.1466 -		SFAM01.1	PPB	1/27/2023 11:59	SMPL
O1298-07	MCOAR3	735	0.1512 -		SFAM01.1	PPB	1/27/2023 12:01	SMPL
O1298-08	MCOAR4	667	0.1369 -		SFAM01.1	PPB	1/27/2023 12:03	SMPL
O1298-09	MCOAR5	895	0.1848 -		SFAM01.1	PPB	1/27/2023 12:05	SMPL
O1298-10	MCOAR7	794	0.1636 -		SFAM01.1	PPB	1/27/2023 12:08	SMPL
O1298-11	MCOAR8	740	0.1522 -		SFAM01.1	PPB	1/27/2023 12:10	SMPL
O1298-12	MCOAR9	559	0.1143 -		SFAM01.1	PPB	1/27/2023 12:12	SMPL
O1298-13	MCOAS0	874	0.1804 -		SFAM01.1	PPB	1/27/2023 12:14	SMPL
O1298-14	MCOAS2	1284	0.2664 -		SFAM01.1	PPB	1/27/2023 12:17	SMPL
O1298-15	MCOAS4	1069	0.2213 -		SFAM01.1	PPB	1/27/2023 12:19	SMPL
O1298-16	MCOAS5	582	0.1191 -		SFAM01.1	PPB	1/27/2023 12:21	SMPL
O1298-17	MCOAS6	1706	0.3549 -		SFAM01.1	PPB	1/27/2023 12:23	SMPL
O1298-18	MCOAS8	920	0.19 -		SFAM01.1	PPB	1/27/2023 12:26	SMPL
O1298-19	MCOAS9	1544	0.3209 -		SFAM01.1	PPB	1/27/2023 12:28	SMPL
CCV035	CCV035	23741	4.9787 -		SFAM01.1	PPB	1/27/2023 12:30	SMPL
CCB035	CCB035	138	0.0259 -		SFAM01.1	PPB	1/27/2023 12:33	SMPL
O1298-20	MCOAT0	943	0.1948 -		SFAM01.1	PPB	1/27/2023 12:35	SMPL
O1298-21	MCOAT0D	891	0.1839 -		SFAM01.1	PPB	1/27/2023 12:37	SMPL
O1298-22	MCOAT0S	13841	2.9013 -		SFAM01.1	PPB	1/27/2023 12:39	SMPL
PB150517BL	PBS517	157	0.0299 -		SFAM01.1	PPB	1/27/2023 12:42	SMPL
O1299-01	MCOAB8	737	0.1516 -		SFAM01.1	PPB	1/27/2023 12:44	SMPL
O1299-02	MCOAB9	1138	0.2358 -		SFAM01.1	PPB	1/27/2023 12:46	SMPL
O1299-03	MCOAC0	978	0.2022 -		SFAM01.1	PPB	1/27/2023 12:48	SMPL
O1299-04	MCOAT1	754	0.1552 -		SFAM01.1	PPB	1/27/2023 12:51	SMPL
O1299-05	MCOAT2	813	0.1676 -		SFAM01.1	PPB	1/27/2023 12:53	SMPL
O1299-06	MCOAT3	710	0.1459 -		SFAM01.1	PPB	1/27/2023 12:55	SMPL
O1299-07	MCOAT4	885	0.1827 -		SFAM01.1	PPB	1/27/2023 12:57	SMPL
O1299-08	MCOAT5	856	0.1766 -		SFAM01.1	PPB	1/27/2023 13:00	SMPL
O1299-09	MCOAT6	789	0.1625 -		SFAM01.1	PPB	1/27/2023 13:02	SMPL
O1299-10	MCOAT7	735	0.1512 -		SFAM01.1	PPB	1/27/2023 13:04	SMPL

LB123830

INSTRUMENT ID : CV1

O1299-11	MC0AT8	895	0.1848 -	SFAM01.1	PPB	1/27/2023 13:07 SMPL
O1299-12	MC0AT9	1252	0.2597 -	SFAM01.1	PPB	1/27/2023 13:09 SMPL
O1299-13	MC0AW0	1586	0.3298 -	SFAM01.1	PPB	1/27/2023 13:11 SMPL
O1299-14	MC0AW1	4330	0.9056 -	SFAM01.1	PPB	1/27/2023 13:13 SMPL
O1299-15	MC0AW2	1005	0.2078 -	SFAM01.1	PPB	1/27/2023 13:16 SMPL
O1299-16	MC0AW3	1408	0.2924 -	SFAM01.1	PPB	1/27/2023 13:18 SMPL
CCV036	CCV036	23894	5.0108 -	SFAM01.1	PPB	1/27/2023 13:20 SMPL
CCB036	CCB036	146	0.0276 -	SFAM01.1	PPB	1/27/2023 13:22 SMPL
O1299-17	MC0AW4	907	0.1873 -	SFAM01.1	PPB	1/27/2023 13:25 SMPL
O1299-18	MC0AW5	988	0.2043 -	SFAM01.1	PPB	1/27/2023 13:27 SMPL
O1299-19	MC0AW6	763	0.1571 -	SFAM01.1	PPB	1/27/2023 13:29 SMPL
O1299-20	MC0AW7	913	0.1885 -	SFAM01.1	PPB	1/27/2023 13:31 SMPL
O1299-21	MC0AW7D	957	0.1978 -	SFAM01.1	PPB	1/27/2023 13:36 SMPL
O1299-22	MC0AW7S	12917	2.7074 -	SFAM01.1	PPB	1/27/2023 13:40 SMPL
CCV037	CCV037	23365	4.8998 -	SFAM01.1	PPB	1/27/2023 13:43 SMPL
CCB037	CCB037	68	0.0112 -	SFAM01.1	PPB	1/27/2023 13:45 SMPL