

63123842


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

41884



A= 0.0000e+000

B= 2.4034e-004

C= -9.9899e-002

Rho= 0.9998779

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%11
0.0	0.000	0.007	0.007	444	0.000	444					-15
0.2	0.200	0.170	-0.030	1124	0.0 %	1124					-2
2.5	2.500	2.458	-0.042	10644	0.0 %	10644					2
5.0	5.000	5.124	0.124	21737	0.0 %	21737					0
7.5	7.500	7.474	-0.026	31512	0.0 %	31512					0
10.0	10.000	9.967	-0.033	41884	0.0 %	41884					0

11237

LB123842 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Jan 2023 15:54:01

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	444	-	0.0000	7470A	PPB	27 Jan 2023 15:55:42	S	Std
0.2 - 1	50.2	1124	-	0.2000	7470A	PPB	27 Jan 2023 15:58:00	S	Std
2.5 - 1	52.5	10644	-	2.5000	7470A	PPB	27 Jan 2023 16:00:17	S	Std
5.0 - 1	55	21737	-	5.0000	7470A	PPB	27 Jan 2023 16:02:34	S	Std
7.5 - 1	57.5	31512	-	7.5000	7470A	PPB	27 Jan 2023 16:04:50	S	Std
10.0 - 1	510	41884	-	10.0000	7470A	PPB	27 Jan 2023 16:09:48	S	Std
ICV23 - 1	ICV23	16498	3.8652	-	7470A	PPB	27 Jan 2023 16:12:37	U	SMPL
ICB23 - 1	ICB23	219	-0.0473	-	7470A	PPB	27 Jan 2023 16:14:51	U	SMPL
CCV38 - 1	CCV38	21299	5.0191	-	7470A	PPB	27 Jan 2023 16:17:07	U	SMPL
CCB38 - 1	CCB38	162	-0.0610	-	7470A	PPB	27 Jan 2023 16:19:23	U	SMPL
CRA - 1	CRA	1095	0.1633	-	7470A	PPB	27 Jan 2023 16:21:38	U	SMPL
HighStd - 1	HighStd	42258	10.0564	-	7470A	PPB	27 Jan 2023 16:23:53	U	SMPL
ChkStd - 1	ChkStd	27864	6.5970	-	7470A	PPB	27 Jan 2023 16:26:09	U	SMPL
PB150521BL - 1	PBW	213	-0.0487	-	7470A	PPB	27 Jan 2023 16:28:24	U	SMPL
PB150521BS - 1	LCSW	16335	3.8261	-	7470A	PPB	27 Jan 2023 16:30:40	U	SMPL
O1289-02 - 1	VNJ209-STOCK	277	-0.0333	-	7470A	PPB	27 Jan 2023 16:32:57	U	SMPL
O1318-02 - 1	DR-1-012523	371	-0.0107	-	7470A	PPB	27 Jan 2023 16:35:13	U	SMPL
O1318-05 - 1	RJ-1-012523	330	-0.0206	-	7470A	PPB	27 Jan 2023 16:37:30	U	SMPL
O1318-05DUP - 1	RJ-1-012523DUP	337	-0.0189	-	7470A	PPB	27 Jan 2023 16:42:32	U	SMPL
O1318-05MS - 1	RJ-1-012523MS	17013	3.9890	-	7470A	PPB	27 Jan 2023 16:44:47	U	SMPL
CCV39 - 1	CCV39	21528	5.0742	-	7470A	PPB	27 Jan 2023 16:47:02	U	SMPL
CCB39 - 1	CCB39	133	-0.0679	-	7470A	PPB	27 Jan 2023 16:49:17	U	SMPL
O1318-05MSD - 1	RJ-1-012523MSD	16187	3.7905	-	7470A	PPB	27 Jan 2023 16:51:33	U	SMPL
PB150493TB - 1		287	-0.0309	-	7470A	PPB	27 Jan 2023 16:53:49	U	SMPL
O1318-05LX5 - 1		58	-0.0860	-	7470A	PPB	27 Jan 2023 16:56:04	U	SMPL
O1318-05A - 1		17144	4.0205	-	7470A	PPB	27 Jan 2023 16:58:20	U	SMPL
CCV40 - 1	CCV40	21428	5.0501	-	7470A	PPB	27 Jan 2023 17:00:36	U	SMPL
CCB40 - 1	CCB40	158	-0.0619	-	7470A	PPB	27 Jan 2023 17:02:53	U	SMPL