



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB123854

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	01/27/2023	18:20
2	CAL2	1	Water	NA	NA	20.8	7.01	01/27/2023	18:21
3	CAL3	1	Water	NA	NA	20.7	10.03	01/27/2023	18:22
4	ICV	1	Water	NA	NA	20.7	6.99	01/27/2023	18:25
5	CCV1	1	Water	NA	NA	20.7	2.01	01/27/2023	18:26
6	O1328-02	1	Solid	20.02	20	22.8	9.52	01/27/2023	18:30
7	O1328-07	1	Solid	20.01	20	22.3	8.75	01/27/2023	18:32
8	O1328-12	1	Solid	20.04	20	22.5	8.40	01/27/2023	18:35
9	O1328-17	1	Solid	20.02	20	22.4	8.04	01/27/2023	18:36
10	O1328-22	1	Solid	20.03	20	22.6	9.10	01/27/2023	18:37
11	O1329-02	1	Solid	20.02	20	22.4	8.26	01/27/2023	18:40
12	O1329-07	1	Solid	20.03	20	22.8	8.25	01/27/2023	18:41
13	O1329-12	1	Solid	20.04	20	23.0	8.19	01/27/2023	18:43
14	O1329-17	1	Solid	20.02	20	22.9	8.47	01/27/2023	18:45
15	O1333-02	1	Solid	20.03	20	22.3	8.25	01/27/2023	18:47
16	CCV2	1	Water	NA	NA	20.7	12.03	01/27/2023	18:50
17	O1333-05	1	Solid	20.03	20	23.6	9.07	01/27/2023	18:52
18	O1335-01	1	Solid	20.04	20	22.2	6.66	01/27/2023	18:55
19	O1352-02	1	Solid	20.04	20	22.6	7.86	01/27/2023	19:00
20	O1356-02	1	Solid	20.02	20	23.3	9.20	01/27/2023	19:02

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	O1356-02DUP	1	Solid	20.03	20	23.3	9.21	01/27/2023	19:03
22	CCV3	1	Water	NA	NA	20.7	2.01	01/27/2023	19:05

WORKLIST(Hardcopy Internal Chain)

123854

WorkList Name : CORROSIIVITY O1333 WorkList ID : 166922 Department : Wet-Chemistry Date : 01-27-2023 11:45:18

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/26/2023	Solid	O1328-02	Corrosivity	Cool 4 deg C	SOUN03	D12	COMP-1-BORING-1-0-7	01/26/2023	9045D
01/26/2023	Solid	O1328-07	Corrosivity	Cool 4 deg C	SOUN03	D12	COMP-2-BORING-1-7-14	01/26/2023	9045D
01/26/2023	Solid	O1328-12	Corrosivity	Cool 4 deg C	SOUN03	D12	COMP-3-BORING-1-14-23	01/26/2023	9045D
01/26/2023	Solid	O1328-17	Corrosivity	Cool 4 deg C	SOUN03	D12	COMP-4-BORING-1-23-25	01/26/2023	9045D
01/26/2023	Solid	O1328-22	Corrosivity	Cool 4 deg C	SOUN03	D12	COMP-5-BORING-2-0-7	01/26/2023	9045D
01/26/2023	Solid	O1329-02	Corrosivity	Cool 4 deg C	SOUN03	D12	COMP-6-BORING-2-7-14	01/26/2023	9045D
01/26/2023	Solid	O1329-07	Corrosivity	Cool 4 deg C	SOUN03	D12	COMP-7-BORING-2-14-21	01/26/2023	9045D
01/26/2023	Solid	O1329-12	Corrosivity	Cool 4 deg C	SOUN03	D12	COMP-8-BORING-2-21-28	01/26/2023	9045D
01/26/2023	Solid	O1329-17	Corrosivity	Cool 4 deg C	SOUN03	D12	COMP-9-BORING-2-28-3	01/26/2023	9045D
01/26/2023	Solid	O1333-02	Corrosivity	Cool 4 deg C	PSEG03	D12	TP-J	01/26/2023	9045D
01/26/2023	Solid	O1333-05	Corrosivity	Cool 4 deg C	PSEG03	D12	MH-11	01/26/2023	9045D
01/27/2023	Solid	O1335-01	Corrosivity	Cool 4 deg C	PSEG03	D21	334-337	01/27/2023	9045D
01/27/2023	Solid	O1352-02	Corrosivity	Cool 4 deg C	PSEG03	D21	TP-2-WC	01/27/2023	9045D
01/27/2023	Solid	O1356-02	Corrosivity	Cool 4 deg C	PSEG03	D21	TP-J	01/27/2023	9045D

Date/Time 01.27.23 18:15
 Raw Sample Received by: JP CACI
 Raw Sample Relinquished by: JP CACI

Date/Time 01.27.23 19:30
 Raw Sample Received by: JP CACI
 Raw Sample Relinquished by: JP CACI