

LB123856


7471B

INSTRUMENT ID: CV2

Linear

μ Abs.:

55229



A= 0.0000e+000

B= 1.8166e-004

C= 5.8023e-002

Rho= 0.9998216

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.068	0.068	57	0.000	57					1
0.2	0.200	0.216	0.016	871	0.0 %	871					8
2.5	2.500	2.484	-0.016	13356	0.0 %	13356					1
5.0	5.000	4.881	-0.119	26552	0.0 %	26552					2
7.5	7.500	7.459	-0.041	40743	0.0 %	40743					1
10.0	10.000	10.091	0.091	55229	0.0 %	55229					1

LB123856 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 30 Jan 2023 13:05:46

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	57	-	0.0000	7471B	PPB	30 Jan 2023 13:10:11	S	Std
0.2 - 1	50.2	871	-	0.2000	7471B	PPB	30 Jan 2023 13:12:28	S	Std
2.5 - 1	52.5	13356	-	2.5000	7471B	PPB	30 Jan 2023 13:14:45	S	Std
5.0 - 1	55	26552	-	5.0000	7471B	PPB	30 Jan 2023 13:17:02	S	Std
7.5 - 1	57.5	40743	-	7.5000	7471B	PPB	30 Jan 2023 13:19:19	S	Std
10.0 - 1	59	55229	-	10.0000	7471B	PPB	30 Jan 2023 13:24:06	S	Std
ICV24 - 1	ICV24	20667	3.8123	-	7471B	PPB	30 Jan 2023 13:26:46	U	SMPL
ICB24 - 1	ICB24	22	0.0620	-	7471B	PPB	30 Jan 2023 13:29:01	U	SMPL
CCV41 - 1	CCV41	26054	4.7909	-	7471B	PPB	30 Jan 2023 13:31:15	U	SMPL
CCB41 - 1	CCB41	-44	0.0500	-	7471B	PPB	30 Jan 2023 13:33:31	U	SMPL
CRA - 1	CRA	868	0.2157	-	7471B	PPB	30 Jan 2023 13:35:46	U	SMPL
HighStd - 1	HighStd	55675	10.1717	-	7471B	PPB	30 Jan 2023 13:38:01	U	SMPL
ChkStd - 1	ChkStd	37801	6.9248	-	7471B	PPB	30 Jan 2023 13:40:16	U	SMPL
PB150533BL - 1	PBS	-85	0.0426	-	7471B	PPB	30 Jan 2023 13:42:32	U	SMPL
PB150533BS - 1	LCSS	18308	3.3838	-	7471B	PPB	30 Jan 2023 13:44:48	U	SMPL
O1318-01 - 1	DR-1-012523	64915	11.8502	-	7471B	PPB	30 Jan 2023 13:47:05	U	SMPL
O1318-04 - 1	RJ-1-012523	22153	4.0822	-	7471B	PPB	30 Jan 2023 13:49:21	U	SMPL
O1322-01 - 1	EO-01-1-26-2023	1073	0.2529	-	7471B	PPB	30 Jan 2023 13:51:38	U	SMPL
O1328-01 - 1	COMP-1-BORING-1-0-7	2905	0.5857	-	7471B	PPB	30 Jan 2023 13:53:55	U	SMPL
CCV42 - 1	CCV42	26150	4.8083	-	7471B	PPB	30 Jan 2023 13:56:26	U	SMPL
CCB42 - 1	CCB42	-19	0.0546	-	7471B	PPB	30 Jan 2023 14:00:41	U	SMPL
O1328-01MS - 1	COMP-1-BORING-1-0-7MS	21789	4.0161	-	7471B	PPB	30 Jan 2023 14:02:57	U	SMPL
O1328-01MSD - 1	COMP-1-BORING-1-0-7MSD	22841	4.2072	-	7471B	PPB	30 Jan 2023 14:05:12	U	SMPL
O1328-06 - 1	COMP-2-BORING-1-7-14	478	0.1449	-	7471B	PPB	30 Jan 2023 14:07:27	U	SMPL
O1328-11 - 1	COMP-3-BORING-1-14-23	242	0.1020	-	7471B	PPB	30 Jan 2023 14:09:43	U	SMPL
O1328-16 - 1	COMP-4-BORING-1-23-25	204	0.0951	-	7471B	PPB	30 Jan 2023 14:11:59	U	SMPL
O1328-21 - 1	COMP-5-BORING-2-0-7	39961	7.3171	-	7471B	PPB	30 Jan 2023 14:14:16	U	SMPL
O1329-01 - 1	COMP-6-BORING-2-7-14	12527	2.3336	-	7471B	PPB	30 Jan 2023 14:16:33	U	SMPL
O1329-06 - 1	COMP-7-BORING-2-14-21	218	0.0976	-	7471B	PPB	30 Jan 2023 14:18:50	U	SMPL
O1329-11 - 1	COMP-8-BORING-2-21-28	309	0.1142	-	7471B	PPB	30 Jan 2023 14:21:06	U	SMPL
O1329-16 - 1	COMP-9-BORING-2-28-35	628	0.1721	-	7471B	PPB	30 Jan 2023 14:23:22	U	SMPL
CCV43 - 1	CCV43	25915	4.7656	-	7471B	PPB	30 Jan 2023 14:25:38	U	SMPL
CCB43 - 1	CCB43	-15	0.0553	-	7471B	PPB	30 Jan 2023 14:27:53	U	SMPL
O1337-01 - 1	SU-02-012723	1629	0.3539	-	7471B	PPB	30 Jan 2023 14:30:09	U	SMPL
O1345-01 - 1	RBR200007	1097	0.2573	-	7471B	PPB	30 Jan 2023 14:32:25	U	SMPL
O1345-03 - 1	ETGI-346	8026	1.5160	-	7471B	PPB	30 Jan 2023 14:34:40	U	SMPL
O1353-01 - 1	72-11985	1558	0.3410	-	7471B	PPB	30 Jan 2023 14:36:56	U	SMPL
O1356-01 - 1	TP-J	444	0.1387	-	7471B	PPB	30 Jan 2023 14:39:12	U	SMPL
O1357-01 - 1	VNJ-205	3130	0.6266	-	7471B	PPB	30 Jan 2023 14:41:28	U	SMPL
O1328-01DUP - 1	COMP-1-BORING-1-0-7DUP	2714	0.5510	-	7471B	PPB	30 Jan 2023 14:59:55	U	SMPL
O1328-01LX5 - 1		465	0.1425	-	7471B	PPB	30 Jan 2023 15:05:08	U	SMPL
O1328-01A - 1		24039	4.4248	-	7471B	PPB	30 Jan 2023 15:07:25	U	SMPL
O1318-01DLX2 - 1		29146	5.3525	-	7471B	PPB	30 Jan 2023 15:09:40	U	SMPL
CCV44 - 1	CCV44	25449	4.6810	-	7471B	PPB	30 Jan 2023 15:11:56	U	SMPL
CCB44 - 1	CCB44	1	0.0582	-	7471B	PPB	30 Jan 2023 15:14:11	U	SMPL