



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB123866

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	01/30/2023	13:30
2	CAL2	1	Water	NA	NA	20.7	7.00	01/30/2023	13:31
3	CAL3	1	Water	NA	NA	20.7	10.03	01/30/2023	13:33
4	ICV	1	Water	NA	NA	20.8	6.98	01/30/2023	13:35
5	CCV1	1	Water	NA	NA	20.8	2.01	01/30/2023	13:37
6	O1358-03	1	Solid	20.02	20	22.4	8.53	01/30/2023	13:40
7	O1358-06	1	Solid	20.03	20	21.6	9.68	01/30/2023	13:42
8	O1358-09	1	Solid	20.04	20	21.6	9.44	01/30/2023	13:45
9	O1358-12	1	Solid	20.02	20	21.9	9.62	01/30/2023	13:50
10	O1358-15	1	Solid	20.03	20	22.2	10.34	01/30/2023	14:00
11	O1360-02	1	Solid	20.03	20	22.1	8.86	01/30/2023	14:02
12	O1363-02	1	Solid	20.04	20	22.3	7.93	01/30/2023	14:10
13	O1363-02DUP	1	Solid	20.03	20	22.4	7.94	01/30/2023	14:11
14	CCV2	1	Water	NA	NA	20.7	12.03	01/30/2023	14:15

VB 122866

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIIVITY O1363 WorkList ID : 166973 Department : Wet-Chemistry Date : 01-30-2023 12:53:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/27/2023	Solid	O1358-03	Corrosivity	Cool 4 deg C	RUTW01	D22	JPP-11-012723	01/27/2023	9045D
01/27/2023	Solid	O1358-06	Corrosivity	Cool 4 deg C	RUTW01	D22	DR-2-012723	01/27/2023	9045D
01/27/2023	Solid	O1358-09	Corrosivity	Cool 4 deg C	RUTW01	D22	JC-2-012723	01/27/2023	9045D
01/27/2023	Solid	O1358-12	Corrosivity	Cool 4 deg C	RUTW01	D22	JC-3-012723	01/27/2023	9045D
01/27/2023	Solid	O1358-15	Corrosivity	Cool 4 deg C	RUTW01	D22	JPP-5-012723	01/27/2023	9045D
01/28/2023	Solid	O1360-02	Corrosivity	Cool 4 deg C	PSEG03	D22	TP-3-WC	01/28/2023	9045D
01/30/2023	Solid	O1363-02	Corrosivity	Cool 4 deg C	PSEG03	D22	TP-1-WC	01/30/2023	9045D

Date/Time 01-30-23 13:10
Raw Sample Received by: JACWC
Raw Sample Relinquished by: JDCSM

Date/Time 01-30-23 17:00
Raw Sample Received by: JDCSM
Raw Sample Relinquished by: JPLAC