

LB123872


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

50282



A= 0.0000e+000

B= 1.9696e-004

C= 1.0238e-002

Rho= 0.9996101

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	P/O
0.0	0.000	0.018	0.018	41	0.000	41					
0.2	0.200	0.181	-0.019	865	0.0 %	865					-10
2.5	2.500	2.372	-0.128	11991	0.0 %	11991					-5
5.0	5.000	5.197	0.197	26334	0.0 %	26334					4
7.5	7.500	7.519	0.019	38123	0.0 %	38123					0
10.0	10.000	9.914	-0.086	50282	0.0 %	50282					1

LB123872 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 31 Jan 2023 16:01:22

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	41	-	0.00007471B	PPB	31 Jan 2023 16:01:32	S	Std
0.2 - 1	50.2	865	-	0.20007471B	PPB	31 Jan 2023 16:03:49	S	Std
2.5 - 1	52.5	11991	-	2.50007471B	PPB	31 Jan 2023 16:06:06	S	Std
5.0 - 1	55	26334	-	5.00007471B	PPB	31 Jan 2023 16:08:24	S	Std
7.5 - 1	57.5	38123	-	7.50007471B	PPB	31 Jan 2023 16:10:40	S	Std
10.0 - 1	59.0	50282	-	10.00007471B	PPB	31 Jan 2023 16:12:56	S	Std
ICV26 - 1	ICV26	22077	4.3584	-7471B	PPB	31 Jan 2023 16:15:41	U	SMPL
ICB26 - 1	ICB26	12	0.0126	-7471B	PPB	31 Jan 2023 16:17:56	U	SMPL
CCV50 - 1	CCV50	26550	5.2394	-7471B	PPB	31 Jan 2023 16:20:11	U	SMPL
CCB50 - 1	CCB50	3	0.0108	-7471B	PPB	31 Jan 2023 16:22:26	U	SMPL
CRA - 1	CRA	876	0.1828	-7471B	PPB	31 Jan 2023 16:24:41	U	SMPL
HighStd - 1	HighStd	50995	10.0540	-7471B	PPB	31 Jan 2023 16:26:56	U	SMPL
ChkStd - 1	ChkStd	36619	7.2225	-7471B	PPB	31 Jan 2023 16:29:12	U	SMPL
PB150582BL - 1	PBS	-58	-0.0012	-7471B	PPB	31 Jan 2023 16:31:28	U	SMPL
PB150582BS - 1	LCSS	18964	3.7453	-7471B	PPB	31 Jan 2023 16:33:44	U	SMPL
O1333-01 - 1	TP-1	1147	0.2361	-7471B	PPB	31 Jan 2023 16:36:01	U	SMPL
O1333-04 - 1	MH-11	557	0.1199	-7471B	PPB	31 Jan 2023 16:38:17	U	SMPL
O1348-01 - 1	N-H5(32-34)	3846	0.7677	-7471B	PPB	31 Jan 2023 16:40:34	U	SMPL
O1348-02 - 1	E-H5(31.5-33.5)	253	0.0601	-7471B	PPB	31 Jan 2023 16:42:51	U	SMPL
O1348-02DUP - 1	E-H5(31.5-33.5)DUP	230	0.0555	-7471B	PPB	31 Jan 2023 16:45:06	U	SMPL
CCV51 - 1	CCV51	26690	5.2670	-7471B	PPB	31 Jan 2023 16:47:21	U	SMPL
CCB51 - 1	CCB51	-41	0.0022	-7471B	PPB	31 Jan 2023 16:49:36	U	SMPL
O1348-02MS - 1	E-H5(31.5-33.5)MS	19269	3.8054	-7471B	PPB	31 Jan 2023 16:51:52	U	SMPL
O1348-02MSD - 1	E-H5(31.5-33.5)MSD	20582	4.0640	-7471B	PPB	31 Jan 2023 16:54:07	U	SMPL
O1348-03 - 1	W-H5(31-32)	4252	0.8477	-7471B	PPB	31 Jan 2023 16:56:23	U	SMPL
O1348-04 - 1	MID1-H5(21-22)	179	0.0455	-7471B	PPB	31 Jan 2023 16:58:38	U	SMPL
O1348-05 - 1	MID2-H5(32-34)	153	0.0404	-7471B	PPB	31 Jan 2023 17:00:54	U	SMPL
O1348-06 - 1	S-H5(30.5-32)	185	0.0467	-7471B	PPB	31 Jan 2023 17:03:10	U	SMPL
O1352-01 - 1	TP-2-WC	1385	0.2830	-7471B	PPB	31 Jan 2023 17:05:27	U	SMPL
O1358-02 - 1	JPP-11-012723	17883	3.5324	-7471B	PPB	31 Jan 2023 17:07:44	U	SMPL
O1358-05 - 1	DR-2-012723	8757	1.7350	-7471B	PPB	31 Jan 2023 17:10:01	U	SMPL
O1358-08 - 1	JC-2-012723	146207	28.8065	-7471B	PPB	31 Jan 2023 17:12:17	U	SMPL
CCV52 - 1	CCV52	26547	5.2388	-7471B	PPB	31 Jan 2023 17:14:33	U	SMPL
CCB52 - 1	CCB52	-34	0.0035	-7471B	PPB	31 Jan 2023 17:16:50	U	SMPL
O1358-11 - 1	JC-3-012723	47634	9.3920	-7471B	PPB	31 Jan 2023 17:19:06	U	SMPL
O1358-14 - 1	JPP-5-012723	14346	2.8358	-7471B	PPB	31 Jan 2023 17:21:21	U	SMPL
O1360-01 - 1	TP-3-WC	1728	0.3506	-7471B	PPB	31 Jan 2023 17:23:37	U	SMPL
O1363-01 - 1	TP-1-WC	1120	0.2308	-7471B	PPB	31 Jan 2023 17:25:53	U	SMPL
O1348-02LX5 - 1		54	0.0209	-7471B	PPB	31 Jan 2023 17:28:09	U	SMPL
O1348-02A - 1		21280	4.2014	-7471B	PPB	31 Jan 2023 17:30:25	U	SMPL
O1358-08DLX10 - 1		16133	3.1877	-7471B	PPB	31 Jan 2023 17:32:42	U	SMPL
CCV53 - 1	CCV53	26733	5.2754	-7471B	PPB	31 Jan 2023 17:34:59	U	SMPL
CCB53 - 1	CCB53	-40	0.0024	-7471B	PPB	31 Jan 2023 17:37:17	U	SMPL