



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB123878

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	02/01/2023	10:00
2	O1335-02		1	12	100.00	>212.0	>212.0	02/01/2023	10:30
3	O1335-02DUP		1	14	100.00	>212.0	>212.0	02/01/2023	11:00
4	O1369-07		1	12	100.00	>212.0	>212.0	02/01/2023	11:30
5	O1369-08		1	13	100.00	>212.0	>212.0	02/01/2023	12:00
6	O1383-05		1	14	100.00	>212.0	>212.0	02/01/2023	12:30
7	O1383-06		1	13	100.00	>212.0	>212.0	02/01/2023	13:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66123878

WorkList Name : fp1-30

WorkList ID : 166961

Department : Wet-Chemistry

Date : 01-30-2023 10:23:16

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/03/2023	Water	O1335-02	Flash Point	Cool 4 deg C	PSEG03	D21	338	01/27/2023	1010B
02/07/2023	Water	O1369-07	Flash Point	Cool 4 deg C	PSEG03	D31	COMP1-0131	01/31/2023	1010B
02/07/2023	Water	O1369-08	Flash Point	Cool 4 deg C	PSEG03	D31	17213	01/31/2023	1010B
02/07/2023	Water	O1383-05	Flash Point	Cool 4 deg C	PSEG03	D31	31603	01/31/2023	1010B
02/07/2023	Water	O1383-06	Flash Point	Cool 4 deg C	PSEG03	D31	31970	01/31/2023	1010B

Date/Time 02/01/2023 09:55
 Raw Sample Received by: RMA CWT
 Raw Sample Relinquished by: 10 (CWT)

Date/Time 02/01/2023 13:40
 Raw Sample Received by: RMA CWT
 Raw Sample Relinquished by: RMA CWT