



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB123880

Slope : 99.0

pH Meter ID : WC pH meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	02/01/2023	08:25
2	CAL2	1	Water	NA	NA	20.8	6.99	02/01/2023	08:26
3	CAL3	1	Water	NA	NA	20.7	10.01	02/01/2023	08:27
4	ICV	1	Water	NA	NA	20.8	7.00	02/01/2023	08:30
5	CCV1	1	Water	NA	NA	20.7	2.01	02/01/2023	08:35
6	O1335-02	1	Water	NA	NA	21.6	7.79	02/01/2023	08:36
7	O1369-07	1	Water	NA	NA	23.5	7.12	02/01/2023	08:40
8	O1369-08	1	Water	NA	NA	24.0	7.69	02/01/2023	08:43
9	O1369-08DUP	1	Water	NA	NA	24.1	7.70	02/01/2023	08:44
10	CCV2	1	Water	NA	NA	20.8	12.03	02/01/2023	08:45

123880

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity o1369 w WorkList ID : 167012 Department : Wet-Chemistry Date : 02-01-2023 08:19:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/27/2023	Water	O1335-02	Corrosivity	Cool 4 deg C	PSEG03	D21 338		01/27/2023	9040C
01/31/2023	Water	O1369-07	Corrosivity	Cool 4 deg C	PSEG03	D31 COMP1-0131		01/31/2023	9040C
01/31/2023	Water	O1369-08	Corrosivity	Cool 4 deg C	PSEG03	D31 17213		01/31/2023	9040C

Date/Time 02.01.23 08:20
 Raw Sample Received by: JP (wc)
 Raw Sample Relinquished by: CP SR

Date/Time 02.01.23 11:00
 Raw Sample Received by: CP SR
 Raw Sample Relinquished by: JP (wc)