



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB123882

Slope : 99.0

pH Meter ID : WC pH meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	02/01/2023	08:25
2	CAL2	1	Water	NA	NA	20.8	6.99	02/01/2023	08:26
3	CAL3	1	Water	NA	NA	20.7	10.01	02/01/2023	08:27
4	ICV	1	Water	NA	NA	20.8	7.00	02/01/2023	08:30
5	CCV1	1	Water	NA	NA	20.7	2.01	02/01/2023	09:05
6	O1379-01	1	Water	NA	NA	20.3	4.73	02/01/2023	09:10
7	O1379-01DUP	1	Water	NA	NA	20.4	4.74	02/01/2023	09:11
8	O1383-05	1	Water	NA	NA	23.4	8.17	02/01/2023	09:15
9	O1383-06	1	Water	NA	NA	23.5	6.68	02/01/2023	09:20
10	CCV2	1	Water	NA	NA	20.6	12.03	02/01/2023	09:22

WORKLIST(Hardcopy Internal Chain)

123842

WorkList Name : ph o1379

WorkList ID : 167013

Department : Wet-Chemistry

Date : 02-01-2023 08:20:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/30/2023	Water	O1379-01	pH	Cool 4 deg C	STER10	N11	CELL-A-POOL-WATER(DE-IONIZED)	01/30/2023	9040C
01/31/2023	Water	O1383-05	pH	Cool 4 deg C	PSEG03	D31	31603	01/31/2023	9040C
01/31/2023	Water	O1383-06	pH	Cool 4 deg C	PSEG03	D31	31970	01/31/2023	9040C

02-01-23 08:22

Date/Time

Raw Sample Received by:

JP (WC)

Raw Sample Relinquished by:

CP SM

Date/Time 02-01-23

Raw Sample Received by:

CP SM

Raw Sample Relinquished by:

JP (WC)