

LB123885


SFAM01.1

INSTRUMENT ID: 1V1

Linear

 μ Abs.:

41532



A= 0.0000e+000

B= 2.4050e-004 *slope*C= -2.3868e-003 *-100 percent*

Rho= 0.9999111

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.00	0.000	0.007	0.007	41	0.000	41					0/00
0.05	0.050					185	0				=
0.20	0.200	0.166	-0.034	699	0.0 %	699					-17
2.50	2.500	2.483	-0.017	10334	0.0 %	10334					-1
5.00	5.000	5.104	0.104	21231	0.0 %	21231					2
7.50	7.500	7.454	-0.046	31006	0.0 %	31006					-1
10.0	10.000	9.986	-0.014	41532	0.0 %	41532					0

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	41	-		0 SFAM01.1	PPB	2/1/2023 12:12	Std
	0.2 S01	699	-		0.2 SFAM01.1	PPB	2/1/2023 12:14	Std
	2.5 S02	10334	-		2.5 SFAM01.1	PPB	2/1/2023 12:16	Std
	5 S03	21231	-		5 SFAM01.1	PPB	2/1/2023 12:18	Std
	7.5 S04	31006	-		7.5 SFAM01.1	PPB	2/1/2023 12:23	Std
	10 S05	41532	-		10 SFAM01.1	PPB	2/1/2023 12:26	Std
ICV028	ICV028	17895	4.3013 -		SFAM01.1	PPB	2/1/2023 12:29	SMPL
ICB028	ICB028	84	0.0178 -		SFAM01.1	PPB	2/1/2023 12:31	SMPL
CCV058	CCV058	21439	5.1536 -		SFAM01.1	PPB	2/1/2023 12:33	SMPL
CCB058	CCB058	44	0.0082 -		SFAM01.1	PPB	2/1/2023 12:35	SMPL
PB150612BL	PBS612	35	0.006 -		SFAM01.1	PPB	2/1/2023 12:38	SMPL
O1300-01	MCOAC1	365	0.0854 -		SFAM01.1	PPB	2/1/2023 12:40	SMPL
O1300-02	MCOAC2	872	0.2073 -		SFAM01.1	PPB	2/1/2023 12:42	SMPL
O1300-03	MCOAC3	362	0.0847 -		SFAM01.1	PPB	2/1/2023 12:44	SMPL
O1300-04	MCOAC4	646	0.153 -		SFAM01.1	PPB	2/1/2023 12:47	SMPL
O1300-05	MCOAC5	535	0.1263 -		SFAM01.1	PPB	2/1/2023 12:49	SMPL
O1300-06	MCOAC6	589	0.1393 -		SFAM01.1	PPB	2/1/2023 12:51	SMPL
O1300-07	MCOAC7	737	0.1749 -		SFAM01.1	PPB	2/1/2023 12:53	SMPL
O1300-08	MCOAC8	1266	0.3021 -		SFAM01.1	PPB	2/1/2023 12:56	SMPL
O1300-09	MCOAC9	1542	0.3685 -		SFAM01.1	PPB	2/1/2023 12:58	SMPL
O1300-10	MCOAD0	1591	0.3802 -		SFAM01.1	PPB	2/1/2023 13:00	SMPL
O1300-11	MCOAD1	1330	0.3175 -		SFAM01.1	PPB	2/1/2023 13:02	SMPL
O1300-12	MCOAD2	1610	0.3848 -		SFAM01.1	PPB	2/1/2023 13:05	SMPL
O1300-13	MCOAD3	814	0.1934 -		SFAM01.1	PPB	2/1/2023 13:07	SMPL
O1300-14	MCOAD4	2382	0.5705 -		SFAM01.1	PPB	2/1/2023 13:09	SMPL
O1300-15	MCOAD5	1496	0.3574 -		SFAM01.1	PPB	2/1/2023 13:11	SMPL
O1300-16	MCOAD6	846	0.2011 -		SFAM01.1	PPB	2/1/2023 13:14	SMPL
O1300-17	MCOAD7	1946	0.4656 -		SFAM01.1	PPB	2/1/2023 13:16	SMPL
O1300-18	MCOAD8	1867	0.4466 -		SFAM01.1	PPB	2/1/2023 13:18	SMPL
O1300-19	MCOAD9	1061	0.2528 -		SFAM01.1	PPB	2/1/2023 13:21	SMPL
CCV059	CCV059	21394	5.1428 -		SFAM01.1	PPB	2/1/2023 13:23	SMPL
CCB059	CCB059	5	-0.0012 -		SFAM01.1	PPB	2/1/2023 13:25	SMPL
O1300-20	MCOAE0	1461	0.349 -		SFAM01.1	PPB	2/1/2023 13:27	SMPL
O1300-21	MCOAE0D	1667	0.3985 -		SFAM01.1	PPB	2/1/2023 13:30	SMPL
O1300-22	MCOAE0S	11907	2.8612 -		SFAM01.1	PPB	2/1/2023 13:32	SMPL
PB150613BL	PBS613	0	-0.0024 -		SFAM01.1	PPB	2/1/2023 13:34	SMPL
O1301-01	MCOAE1	1307	0.3119 -		SFAM01.1	PPB	2/1/2023 13:36	SMPL
O1301-02	MCOAE2	1205	0.2874 -		SFAM01.1	PPB	2/1/2023 13:39	SMPL
O1301-03	MCOAE4	983	0.234 -		SFAM01.1	PPB	2/1/2023 13:41	SMPL
O1301-04	MCOAE5	667	0.158 -		SFAM01.1	PPB	2/1/2023 13:43	SMPL
O1301-05	MCOAE6	874	0.2078 -		SFAM01.1	PPB	2/1/2023 13:46	SMPL
O1301-06	MCOAE7	926	0.2203 -		SFAM01.1	PPB	2/1/2023 13:48	SMPL
O1301-07	MCOAE8	875	0.208 -		SFAM01.1	PPB	2/1/2023 13:50	SMPL
O1301-08	MCOAE9	831	0.1975 -		SFAM01.1	PPB	2/1/2023 13:52	SMPL
O1301-09	MCOAF0	907	0.2157 -		SFAM01.1	PPB	2/1/2023 13:55	SMPL
O1301-10	MCOAF3	861	0.2047 -		SFAM01.1	PPB	2/1/2023 13:57	SMPL

LB123885

INSTRUMENT ID : CV1

O1301-11	MC0AF4	815	0.1936 -	SFAM01.1	PPB	2/1/2023 13:59 SMPL
O1301-12	MC0AF5	1030	0.2453 -	SFAM01.1	PPB	2/1/2023 14:01 SMPL
O1301-13	MC0AF6	1027	0.2446 -	SFAM01.1	PPB	2/1/2023 14:04 SMPL
O1301-14	MC0AF7	1158	0.2761 -	SFAM01.1	PPB	2/1/2023 14:06 SMPL
O1301-15	MC0AF9	3347	0.8026 -	SFAM01.1	PPB	2/1/2023 14:08 SMPL
O1301-16	MC0AG1	754	0.1789 -	SFAM01.1	PPB	2/1/2023 14:10 SMPL
CCV060	CCV060	21605	5.1935 -	SFAM01.1	PPB	2/1/2023 14:13 SMPL
CCB060	CCB060	0	-0.0024 -	SFAM01.1	PPB	2/1/2023 14:15 SMPL
O1301-17	MC0AG2	811	0.1927 -	SFAM01.1	PPB	2/1/2023 14:17 SMPL
O1301-18	MC0AG3	1222	0.2915 -	SFAM01.1	PPB	2/1/2023 14:20 SMPL
O1301-19	MC0AG4	3282	0.7869 -	SFAM01.1	PPB	2/1/2023 14:22 SMPL
O1301-20	MC0AG4D	2976	0.7133 -	SFAM01.1	PPB	2/1/2023 14:24 SMPL
O1301-21	MC0AG4S	13521	3.2494 -	SFAM01.1	PPB	2/1/2023 14:26 SMPL
CCV061	CCV061	21699	5.2161 -	SFAM01.1	PPB	2/1/2023 14:29 SMPL
CCB061	CCB061	0	-0.0024 -	SFAM01.1	PPB	2/1/2023 14:31 SMPL