



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB123889

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	02/01/2023	14:30
2	CAL2	1	Water	NA	NA	20.7	7.00	02/01/2023	14:31
3	CAL3	1	Water	NA	NA	20.8	10.03	02/01/2023	14:32
4	ICV	1	Water	NA	NA	20.8	6.98	02/01/2023	14:35
5	CCV1	1	Water	NA	NA	20.6	2.01	02/01/2023	14:36
6	O1386-04	1	Solid	20.02	20	22.2	8.86	02/01/2023	14:38
7	O1386-04DUP	1	Solid	20.01	20	22.1	8.87	02/01/2023	14:39
8	O1386-08	1	Solid	20.03	20	21.5	8.61	02/01/2023	14:40
9	O1386-12	1	Solid	20.04	20	21.6	9.26	02/01/2023	14:45
10	O1386-16	1	Solid	20.02	20	22.4	11.02	02/01/2023	14:49
11	O1390-02	1	Solid	20.01	20	23.2	9.01	02/01/2023	14:50
12	O1390-05	1	Solid	20.02	20	22.9	8.56	02/01/2023	15:00
13	CCV2	1	Water	NA	NA	20.7	12.03	02/01/2023	15:01

WORKLIST(Hardcopy Internal Chain)

VB 123889

WorkList Name : CORROSIIVITY O1386 WorkList ID : 167033 Department : Wet-Chemistry Date : 02-01-2023 10:46:23

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/31/2023	Solid	O1386-04	Corrosivity	Cool 4 deg C	RUTW01	E51	JPP-9-013123	01/31/2023	9045D
01/31/2023	Solid	O1386-08	Corrosivity	Cool 4 deg C	RUTW01	E51	JPP-6-013123	01/31/2023	9045D
01/31/2023	Solid	O1386-12	Corrosivity	Cool 4 deg C	RUTW01	E51	JPP-14-013123	01/31/2023	9045D
01/31/2023	Solid	O1386-16	Corrosivity	Cool 4 deg C	RUTW01	E51	JPP-15-013123	01/31/2023	9045D
02/01/2023	Solid	O1390-02	Corrosivity	Cool 4 deg C	PSEG03	E51	STOCKPILE-1	02/01/2023	9045D
02/01/2023	Solid	O1390-05	Corrosivity	Cool 4 deg C	PSEG03	E51	STOCKPILE-2	02/01/2023	9045D

Date/Time 02-01-23 12/17
 Raw Sample Received by: 10 (WC) CJP SM
 Raw Sample Relinquished by: CJP SM

Date/Time 02-01-23 12/17
 Raw Sample Received by: CJP SM
 Raw Sample Relinquished by: 10 (WC)