



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB123893

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	02/02/2023	12:45
2	O1369-06		1	8	16.00	60.8	60.5	02/02/2023	13:15
3	O1369-06DUP		1	8	16.00	60.8	60.5	02/02/2023	13:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66123893

WorkList Name : fp2-2

WorkList ID : 167058

Department : Wet-Chemistry

Date : 02-02-2023 08:24:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/07/2023	Solid	O1369-06	Flash Point	Cool 4 deg C	PSEG03	D31	17221	01/31/2023	1010B

Date/Time 02/02/2023 12:35
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: JG (CWC)

Date/Time 02/02/2023 14:25
 Raw Sample Received by: JG (CWC)
 Raw Sample Relinquished by: RM CWC