

LB123894


7471B

INSTRUMENT ID: CY4

Linear

 μ Abs.:

50781



A= 0.0000e+000

B= 1.9297e-004

C= 4.3489e-003

Rho= 0.9992212

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.014	0.014	48	0.000	48					-
0.2	0.200	0.177	-0.023	894	0.0 %	894					-12
2.5	2.500	2.466	-0.034	12756	0.0 %	12756					-1
5.0	5.000	4.950	-0.050	25626	0.0 %	25626					-1
7.5	7.500	7.790	0.290	40347	0.0 %	40347					4
10.0	10.000	9.804	-0.196	50781	0.0 %	50781					-2

LB123894 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 02 Feb 2023 10:20:53

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	48	-	0.0000	7471B	PPB	02 Feb 2023 10:33:51	S	Std
0.2 - 1	SD	894	-	0.2000	7471B	PPB	02 Feb 2023 10:36:09	S	Std
2.5 - 1	SD	12756	-	2.5000	7471B	PPB	02 Feb 2023 10:38:26	S	Std
5.0 - 1	SD	25626	-	5.0000	7471B	PPB	02 Feb 2023 10:40:43	S	Std
7.5 - 1	SD	40347	-	7.5000	7471B	PPB	02 Feb 2023 10:42:59	S	Std
10.0 - 1	SD	50781	-	10.0000	7471B	PPB	02 Feb 2023 10:45:15	S	Std
ICV29 - 1	ICV29	22362	4.3196	-	7471B	PPB	02 Feb 2023 10:48:16	U	SMPL
ICB29 - 1	ICB29	36	0.0113	-	7471B	PPB	02 Feb 2023 10:50:31	U	SMPL
CCV62 - 1	CCV62	25581	4.9408	-	7471B	PPB	02 Feb 2023 10:52:46	U	SMPL
CCB62 - 1	CCB62	-3	0.0038	-	7471B	PPB	02 Feb 2023 10:55:01	U	SMPL
CRA - 1	CRA	889	0.1759	-	7471B	PPB	02 Feb 2023 10:57:16	U	SMPL
HighStd - 1	HighStd	50789	9.8053	-	7471B	PPB	02 Feb 2023 10:59:31	U	SMPL
ChkStd - 1	ChkStd	37266	7.1957	-	7471B	PPB	02 Feb 2023 11:01:47	U	SMPL
PB150619BL - 1	PBS	-54	-0.0061	-	7471B	PPB	02 Feb 2023 11:04:03	U	SMPL
PB150619BS - 1	LCSS	18501	3.5746	-	7471B	PPB	02 Feb 2023 11:06:19	U	SMPL
O1144-04 - 1	WASTE-ELIZABETH	2054	0.4007	-	7471B	PPB	02 Feb 2023 11:08:35	U	SMPL
O1366-01 - 1	NB-07-013023	1216	0.2390	-	7471B	PPB	02 Feb 2023 11:10:52	U	SMPL
O1371-01 - 1	1-BERKSHIRE-NW	5707	1.1057	-	7471B	PPB	02 Feb 2023 11:13:09	U	SMPL
O1371-02 - 1	2-RT-15-N	6311	1.2222	-	7471B	PPB	02 Feb 2023 11:15:26	U	SMPL
O1371-02DUP - 1	2-RT-15-NDUP	6119	1.1852	-	7471B	PPB	02 Feb 2023 11:17:42	U	SMPL
CCV63 - 1	CCV63	25652	4.9545	-	7471B	PPB	02 Feb 2023 11:19:58	U	SMPL
CCB63 - 1	CCB63	-53	-0.0059	-	7471B	PPB	02 Feb 2023 11:22:13	U	SMPL
O1371-02MS - 1	2-RT-15-NMS	24199	4.6741	-	7471B	PPB	02 Feb 2023 11:24:29	U	SMPL
O1371-02MSD - 1	2-RT-15-NMSD	24200	4.6743	-	7471B	PPB	02 Feb 2023 11:26:45	U	SMPL
O1371-03 - 1	3-RT-15-S	5152	0.9986	-	7471B	PPB	02 Feb 2023 11:29:00	U	SMPL
O1371-04 - 1	4-RAMP-E	25791	4.9814	-	7471B	PPB	02 Feb 2023 11:31:16	U	SMPL
O1374-01 - 1	JC-3-13123-A	31079	6.0018	-	7471B	PPB	02 Feb 2023 11:33:32	U	SMPL
O1374-03 - 1	JC-3-13123-B	59025	11.3947	-	7471B	PPB	02 Feb 2023 11:35:49	U	SMPL
O1382-01 - 1	OR-3-013123	5688	1.1020	-	7471B	PPB	02 Feb 2023 11:38:05	U	SMPL
O1383-01 - 1	0042	316	0.0653	-	7471B	PPB	02 Feb 2023 11:40:23	U	SMPL
O1383-04 - 1	2874-79	664	0.1325	-	7471B	PPB	02 Feb 2023 11:42:40	U	SMPL
O1385-01 - 1	HR-02-013123	2162	0.4216	-	7471B	PPB	02 Feb 2023 11:44:56	U	SMPL
CCV64 - 1	CCV64	25588	4.9422	-	7471B	PPB	02 Feb 2023 11:47:11	U	SMPL
CCB64 - 1	CCB64	-61	-0.0074	-	7471B	PPB	02 Feb 2023 11:49:27	U	SMPL
O1371-02LX5 - 1		1214	0.2386	-	7471B	PPB	02 Feb 2023 11:51:43	U	SMPL
O1371-02A - 1		25494	4.9240	-	7471B	PPB	02 Feb 2023 11:53:58	U	SMPL
O1374-03DLX2 - 1		29427	5.6830	-	7471B	PPB	02 Feb 2023 11:56:14	U	SMPL
CCV65 - 1	CCV65	25457	4.9169	-	7471B	PPB	02 Feb 2023 11:58:29	U	SMPL
CCB65 - 1	CCB65	3	0.0049	-	7471B	PPB	02 Feb 2023 12:00:45	U	SMPL