



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB123897

Slope : 99.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	02/02/2023	08:15
2	CAL2	1	Water	NA	NA	20.7	7.01	02/02/2023	08:16
3	CAL3	1	Water	NA	NA	20.7	10.03	02/02/2023	08:19
4	ICV	1	Water	NA	NA	20.7	6.98	02/02/2023	08:20
5	CCV1	1	Water	NA	NA	20.7	2.01	02/02/2023	08:25
6	O1404-01	1	Solid	20.02	20	23.1	7.60	02/02/2023	08:30
7	O1404-01DUP	1	Solid	20.03	20	23.2	7.61	02/02/2023	08:31
8	CCV2	1	Water	NA	NA	20.7	12.03	02/02/2023	08:35

13123897

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph o1404 WorkList ID : 167051 Department : Wet-Chemistry Date : 02-02-2023 08:06:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/01/2023	Solid	O1404-01	pH		Cool 4 deg C	NEWP02 E51	NPHACK-SOIL-0201	02/01/2023	9045D

Date/Time 02/02/23 08:10
Raw Sample Received by: JDCWC
Raw Sample Relinquished by: JDCWC

Date/Time 02/02/23 10:00
Raw Sample Received by: JDCWC
Raw Sample Relinquished by: JDCWC