

LB123905


SFAM01.1

INSTRUMENT ID: EV1

Linear

μ Abs.:

44984



A= 0.0000e+000

B= 2.2565e-004 slope

C= -7.7363e-002 y-intercept

Rho= 0.9999105

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	4/5/1
0.00	0.000	0.037	0.037	507	0.000	507					
0.05	0.050					185	0				
0.20	0.200	0.192	-0.008	1195	0.0 %	1195					-4
2.50	2.500	2.510	0.010	11467	0.0 %	11467					0
5.00	5.000	4.970	-0.030	22369	0.0 %	22369					-1
7.50	7.500	7.417	-0.083	33214	0.0 %	33214					-1
10.0	10.000	10.073	0.073	44984	0.0 %	44984					1

LB123905

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	507	-		0 SFAM01.1	PPB	2/2/2023 14:03	Std
	0.2 S01	1195	-		0.2 SFAM01.1	PPB	2/2/2023 14:05	Std
	2.5 S02	11467	-		2.5 SFAM01.1	PPB	2/2/2023 14:08	Std
	5 S03	22369	-		5 SFAM01.1	PPB	2/2/2023 14:10	Std
	7.5 S04	33214	-		7.5 SFAM01.1	PPB	2/2/2023 14:12	Std
	10 S05	44984	-		10 SFAM01.1	PPB	2/2/2023 14:14	Std
ICV030	ICV030	19305	4.2788 -		SFAM01.1	PPB	2/2/2023 14:17	SMPL
ICB030	ICB030	518	0.0395 -		SFAM01.1	PPB	2/2/2023 14:20	SMPL
CCV066	CCV066	22494	4.9983 -		SFAM01.1	PPB	2/2/2023 14:22	SMPL
CCB066	CCB066	468	0.0282 -		SFAM01.1	PPB	2/2/2023 14:24	SMPL
PB150647BL	PBS647	544	0.0454 -		SFAM01.1	PPB	2/2/2023 14:26	SMPL
O1302-01	MCOAG5	3780	0.7756 -		SFAM01.1	PPB	2/2/2023 14:29	SMPL
O1302-02	MCOAG6	1873	0.3453 -		SFAM01.1	PPB	2/2/2023 14:31	SMPL
O1302-03	MCOAG7	2164	0.4109 -		SFAM01.1	PPB	2/2/2023 14:33	SMPL
O1302-04	MCOAG8	1668	0.299 -		SFAM01.1	PPB	2/2/2023 14:36	SMPL
O1302-05	MCOAG9	1313	0.2189 -		SFAM01.1	PPB	2/2/2023 14:38	SMPL
O1302-06	MCOAH0	797	0.1025 -		SFAM01.1	PPB	2/2/2023 14:40	SMPL
O1302-07	MCOAH2	1150	0.1821 -		SFAM01.1	PPB	2/2/2023 14:42	SMPL
O1302-08	MCOAH3	1031	0.1553 -		SFAM01.1	PPB	2/2/2023 14:45	SMPL
O1302-09	MCOAH4	1088	0.1681 -		SFAM01.1	PPB	2/2/2023 14:47	SMPL
O1302-10	MCOAH5	1166	0.1857 -		SFAM01.1	PPB	2/2/2023 14:49	SMPL
O1302-11	MCOAH7	937	0.1341 -		SFAM01.1	PPB	2/2/2023 14:51	SMPL
O1302-12	MCOAH9	930	0.1325 -		SFAM01.1	PPB	2/2/2023 14:54	SMPL
O1302-13	MCOAJ0	3210	0.647 -		SFAM01.1	PPB	2/2/2023 14:56	SMPL
O1302-14	MCOAJ1	937	0.1341 -		SFAM01.1	PPB	2/2/2023 14:58	SMPL
O1302-15	MCOAJ3	753	0.0925 -		SFAM01.1	PPB	2/2/2023 15:01	SMPL
O1302-16	MCOAJ4	767	0.0957 -		SFAM01.1	PPB	2/2/2023 15:03	SMPL
O1302-17	MCOAJ5	987	0.1454 -		SFAM01.1	PPB	2/2/2023 15:05	SMPL
O1302-18	MCOAJ6	749	0.0916 -		SFAM01.1	PPB	2/2/2023 15:07	SMPL
O1302-19	MCOAJ7	1110	0.1731 -		SFAM01.1	PPB	2/2/2023 15:10	SMPL
CCV067	CCV067	22476	4.9943 -		SFAM01.1	PPB	2/2/2023 15:12	SMPL
CCB067	CCB067	411	0.0154 -		SFAM01.1	PPB	2/2/2023 15:14	SMPL
O1302-20	MCOAJ8	1697	0.3056 -		SFAM01.1	PPB	2/2/2023 15:16	SMPL
O1302-21	MCOAJ8D	1657	0.2965 -		SFAM01.1	PPB	2/2/2023 15:21	SMPL
O1302-22	MCOAJ8S	12002	2.6309 -		SFAM01.1	PPB	2/2/2023 15:24	SMPL
PB150648BL	PBS648	647	0.0686 -		SFAM01.1	PPB	2/2/2023 15:26	SMPL
O1303-01	MCOAJ9	1707	0.3078 -		SFAM01.1	PPB	2/2/2023 15:28	SMPL
O1303-02	MCOAK0	1729	0.3128 -		SFAM01.1	PPB	2/2/2023 15:30	SMPL
O1303-03	MCOAK1	1434	0.2462 -		SFAM01.1	PPB	2/2/2023 15:33	SMPL
O1303-04	MCOAK2	1337	0.2243 -		SFAM01.1	PPB	2/2/2023 15:35	SMPL
O1303-05	MCOAK3	2329	0.4482 -		SFAM01.1	PPB	2/2/2023 15:37	SMPL
O1303-06	MCOAK4	2993	0.598 -		SFAM01.1	PPB	2/2/2023 15:40	SMPL
O1303-07	MCOAK5	1531	0.2681 -		SFAM01.1	PPB	2/2/2023 15:42	SMPL
O1303-08	MCOAK6	2231	0.4261 -		SFAM01.1	PPB	2/2/2023 15:44	SMPL
O1303-09	MCOAK7	1436	0.2467 -		SFAM01.1	PPB	2/2/2023 15:46	SMPL
O1303-10	MCOAK8	1508	0.2629 -		SFAM01.1	PPB	2/2/2023 15:49	SMPL

LB123905

INSTRUMENT ID : CV1

O1303-11	MC0AK9	2062	0.3879 -	SFAM01.1	PPB	2/2/2023 15:51 SMPL
O1303-12	MC0AL0	1685	0.3029 -	SFAM01.1	PPB	2/2/2023 15:53 SMPL
O1303-13	MC0AL1	1927	0.3575 -	SFAM01.1	PPB	2/2/2023 15:55 SMPL
O1303-14	MC0AL2	1191	0.1914 -	SFAM01.1	PPB	2/2/2023 15:58 SMPL
O1303-15	MC0AW8	2454	0.4764 -	SFAM01.1	PPB	2/2/2023 16:00 SMPL
O1303-16	MC0AW9	2487	0.4838 -	SFAM01.1	PPB	2/2/2023 16:02 SMPL
CCV068	CCV068	22477	4.9945 -	SFAM01.1	PPB	2/2/2023 16:04 SMPL
CCB068	CCB068	407	0.0145 -	SFAM01.1	PPB	2/2/2023 16:07 SMPL
O1303-17	MC0AX0	4548	0.9489 -	SFAM01.1	PPB	2/2/2023 16:09 SMPL
O1303-18	MC0AX1	1340	0.225 -	SFAM01.1	PPB	2/2/2023 16:11 SMPL
O1303-19	MC0AX2	1643	0.2934 -	SFAM01.1	PPB	2/2/2023 16:14 SMPL
O1303-20	MC0AX3	3120	0.6267 -	SFAM01.1	PPB	2/2/2023 16:16 SMPL
O1303-21	MC0AX3D	3115	0.6255 -	SFAM01.1	PPB	2/2/2023 16:20 SMPL
O1303-22	MC0AX3S	12931	2.8405 -	SFAM01.1	PPB	2/2/2023 16:23 SMPL
CCV069	CCV069	20735	4.6014 -	SFAM01.1	PPB	2/2/2023 16:25 SMPL
CCB069	CCB069	399	0.0127 -	SFAM01.1	PPB	2/2/2023 16:27 SMPL