



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB123919

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: FLASH POINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	02/03/2023	08:40
2	O1391-03		1	12	100.00	>212.0	>212.0	02/03/2023	09:10
3	O1391-04		1	14	100.00	>212.0	>212.0	02/03/2023	09:40
4	O1406-02		1	8	17.00	62.6	62.3	02/03/2023	10:10
5	O1406-02DUP		1	8	17.00	62.6	62.3	02/03/2023	10:40
6	O1406-03		1	14	100.00	>212.0	>212.0	02/03/2023	11:10
7	O1407-05		1	13	100.00	>212.0	>212.0	02/03/2023	11:40
8	O1407-06		1	14	100.00	>212.0	>212.0	02/03/2023	12:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66123919

WorkList Name : fp2-3

WorkList ID : 167094

Department : Wet-Chemistry

Date : 02-03-2023 08:19:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/08/2023	Water	O1391-03	Flash Point	Cool 4 deg C	PSEG03	E51	8283	02/01/2023	1010B
02/08/2023	Water	O1391-04	Flash Point	Cool 4 deg C	PSEG03	E51	COMP2-0201	02/01/2023	1010B
02/09/2023	Water	O1406-02	Flash Point	Cool 4 deg C	PSEG03	E52	COMP1-0202	02/02/2023	1010B
02/09/2023	Water	O1406-03	Flash Point	Cool 4 deg C	PSEG03	E52	COMP2-0202	02/02/2023	1010B
02/07/2023	Water	O1407-05	Flash Point	Cool 4 deg C	PSEG03	E52	30985	02/02/2023	1010B
02/07/2023	Water	O1407-06	Flash Point	Cool 4 deg C	PSEG03	E52	30986	02/02/2023	1010B

Date/Time 02/03/2023 08130

Raw Sample Received by: RM Cwe

Raw Sample Relinquished by: Cwe

Date/Time 02/03/2023 12:50

Raw Sample Received by: Cwe

Raw Sample Relinquished by: RM Cwe