



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB123920

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: F;ASH POINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	02/03/2023	14:45
2	O1406-01		1	6	12.00	53.6	53.3	02/03/2023	15:15
3	O1406-01DUP		1	6	12.00	53.6	53.3	02/03/2023	15:45
4	O1407-08		1	8	14.00	57.2	56.9	02/03/2023	16:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6123920

WorkList Name : fp s2-3

WorkList ID : 167097

Department : Wet-Chemistry

Date : 02-03-2023 08:31:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/09/2023	Solid	O1406-01	Flash Point	Cool 4 deg C	PSEG03	E52	5715	02/02/2023	1010B
02/07/2023	Solid	O1407-08	Flash Point	Cool 4 deg C	PSEG03	E52	30983	02/02/2023	1010B

Date/Time 02/03/2023 14:30
Raw Sample Received by: RMA CWC
Raw Sample Relinquished by: RA (Est. 10.5)

Date/Time 02/03/2023 16:50
Raw Sample Received by: RA (Est. 10.5)
Raw Sample Relinquished by: RA CWC