

LB123921


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

36309



A= 0.0000e+000

B= 2.7743e-004

C= 2.8298e-002

Rho= 0.9997680

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/p D
0.0	0.000	0.060	0.060	115	0.000	115					
0.2	0.200	0.216	0.016	677	0.0 %	677					
2.5	2.500	2.517	0.017	8970	0.0 %	8970					
5.0	5.000	4.855	-0.145	17399	0.0 %	17399					
7.5	7.500	7.450	-0.050	26752	0.0 %	26752					
10.0	10.000	10.101	0.101	36309	0.0 %	36309					

o/p D
1
8
1
3
1

LB123921 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 03 Feb 2023 11:20:25

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	115	-	0.0000	7470A	PPB	03 Feb 2023 11:28:06	S	Std
0.2 - 1	50.2	677	-	0.2000	7470A	PPB	03 Feb 2023 11:30:23	S	Std
2.5 - 1	52.5	8970	-	2.5000	7470A	PPB	03 Feb 2023 11:32:40	S	Std
5.0 - 1	55	17399	-	5.0000	7470A	PPB	03 Feb 2023 11:34:58	S	Std
7.5 - 1	57.5	26752	-	7.5000	7470A	PPB	03 Feb 2023 11:37:14	S	Std
10.0 - 1	511	36309	-	10.0000	7470A	PPB	03 Feb 2023 11:42:09	S	Std
ICV31 - 1	ICV31	14148	3.9534	-	7470A	PPB	03 Feb 2023 11:48:09	U	SMPL
ICB31 - 1	ICB31	25	0.0352	-	7470A	PPB	03 Feb 2023 11:50:23	U	SMPL
CCV70 - 1	CCV70	16976	4.7379	-	7470A	PPB	03 Feb 2023 11:55:24	U	SMPL
CCB70 - 1	CCB70	-17	0.0236	-	7470A	PPB	03 Feb 2023 11:57:39	U	SMPL
CRA - 1	CRA	596	0.1936	-	7470A	PPB	03 Feb 2023 11:59:54	U	SMPL
HighStd - 1	HighStd	36562	10.1717	-	7470A	PPB	03 Feb 2023 12:02:09	U	SMPL
ChkStd - 1	ChkStd	24311	6.7729	-	7470A	PPB	03 Feb 2023 12:04:26	U	SMPL
PB150642BL - 1	PBW	-25	0.0214	-	7470A	PPB	03 Feb 2023 12:06:41	U	SMPL
PB150642BS - 1	LCSW	12625	3.5308	-	7470A	PPB	03 Feb 2023 12:08:57	U	SMPL
O1369-04 - 1	COMP2-0131	9	0.0308	-	7470A	PPB	03 Feb 2023 12:11:13	U	SMPL
O1369-05 - 1	COMP3-0131	27	0.0358	-	7470A	PPB	03 Feb 2023 12:13:29	U	SMPL
O1369-06 - 1	17221	762	0.2397	-	7470A	PPB	03 Feb 2023 12:15:46	U	SMPL
O1372-01 - 1	149TH-ST-DRUM-11-POLY	-160	-0.0161	-	7470A	PPB	03 Feb 2023 12:18:04	U	SMPL
O1372-02 - 1	149TH-ST-DRUM-12-PPE	-42	0.0166	-	7470A	PPB	03 Feb 2023 12:20:20	U	SMPL
CCV71 - 1	CCV71	17185	4.7959	-	7470A	PPB	03 Feb 2023 12:25:10	U	SMPL
CCB71 - 1	CCB71	-18	0.0233	-	7470A	PPB	03 Feb 2023 12:27:24	U	SMPL
O1386-04 - 1	JPP-9-013123	50	0.0422	-	7470A	PPB	03 Feb 2023 12:29:40	U	SMPL
O1386-08 - 1	JPP-6-013123	100	0.0560	-	7470A	PPB	03 Feb 2023 12:31:56	U	SMPL
O1386-12 - 1	JPP-14-013123	28	0.0361	-	7470A	PPB	03 Feb 2023 12:34:11	U	SMPL
O1386-16 - 1	JPP-15-013123	17	0.0330	-	7470A	PPB	03 Feb 2023 12:36:27	U	SMPL
O1390-02 - 1	STOCKPILE-1	22	0.0344	-	7470A	PPB	03 Feb 2023 12:38:43	U	SMPL
O1390-05 - 1	STOCKPILE-2	17	0.0330	-	7470A	PPB	03 Feb 2023 12:40:59	U	SMPL
O1390-05DUP - 1	STOCKPILE-2DUP	15	0.0325	-	7470A	PPB	03 Feb 2023 12:43:16	U	SMPL
O1390-05MSD - 1	STOCKPILE-2MS	13210	3.6931	-	7470A	PPB	03 Feb 2023 12:45:33	U	SMPL
O1390-05MSD - 1	STOCKPILE-2MSD	13887	3.8810	-	7470A	PPB	03 Feb 2023 12:47:50	U	SMPL
O1391-01 - 1	8288	-9	0.0258	-	7470A	PPB	03 Feb 2023 12:50:06	U	SMPL
CCV72 - 1	CCV72	17442	4.8672	-	7470A	PPB	03 Feb 2023 12:52:22	U	SMPL
CCB72 - 1	CCB72	5	0.0297	-	7470A	PPB	03 Feb 2023 12:54:37	U	SMPL
O1391-02 - 1	COMP1-0201	20	0.0338	-	7470A	PPB	03 Feb 2023 12:56:53	U	SMPL
PB150643BL - 1	PBW	-3	0.0275	-	7470A	PPB	03 Feb 2023 12:59:09	U	SMPL
PB150643BS - 1	LCSW	11558	3.2348	-	7470A	PPB	03 Feb 2023 13:01:24	U	SMPL
O1369-07 - 1	COMP1-0131	234	0.0932	-	7470A	PPB	03 Feb 2023 13:03:40	U	SMPL
O1369-08 - 1	17213	-66	0.0100	-	7470A	PPB	03 Feb 2023 13:05:57	U	SMPL
O1369-08DUP - 1	17213DUP	0	0.0283	-	7470A	PPB	03 Feb 2023 13:08:13	U	SMPL
O1369-08MS - 1	17213MS	12859	3.5958	-	7470A	PPB	03 Feb 2023 13:10:29	U	SMPL
O1391-03 - 1	8283	103	0.0569	-	7470A	PPB	03 Feb 2023 13:15:04	U	SMPL
O1391-04 - 1	COMP2-0201	-193	-0.0252	-	7470A	PPB	03 Feb 2023 13:17:19	U	SMPL
CCV73 - 1	CCV73	17272	4.8201	-	7470A	PPB	03 Feb 2023 13:19:35	U	SMPL
CCB73 - 1	CCB73	-21	0.0225	-	7470A	PPB	03 Feb 2023 13:21:50	U	SMPL
PB150611TB - 1	PB150611TB	21	0.0341	-	7470A	PPB	03 Feb 2023 13:24:06	U	SMPL
PB150615TB - 1	PB150615TB	33	0.0375	-	7470A	PPB	03 Feb 2023 13:26:22	U	SMPL
O1390-05LX5 - 1		-4	0.0272	-	7470A	PPB	03 Feb 2023 13:28:38	U	SMPL
O1390-05A - 1		14830	4.1426	-	7470A	PPB	03 Feb 2023 13:30:53	U	SMPL
O1369-08LX5 - 1		-47	0.0153	-	7470A	PPB	03 Feb 2023 13:33:09	U	SMPL
O1369-08A - 1		16199	4.5224	-	7470A	PPB	03 Feb 2023 13:35:26	U	SMPL
O1369-08MSD - 1	17213MSD	13245	3.7028	-	7470A	PPB	03 Feb 2023 13:55:32	U	SMPL
CCV74 - 1	CCV74	19572	5.4581	-	7470A	PPB	03 Feb 2023 13:58:07	U	SMPL
CCB74 - 1	CCB74	-15	0.0241	-	7470A	PPB	03 Feb 2023 14:00:23	U	SMPL