



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Sohil

Run Number: LB123931

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP100870
Turbidity Calibration std, 20NTU	WP100865
Turbidity Calibration std, 40NTU	WP100864
Turbidity Calibration std, 1NTU	WP100868
Turbidity Calibration std, 5NTU	WP100867
Turbidity Calibration - CCV std, 10 NTU	WP100866
10 NTU Standard 500 ml	W2971

Intercept: 0.054

Slope: 0.9292

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.152	0.11		02/03/2023	13:20
2	CAL2	1	1	1.150	1.18	-18	02/03/2023	13:24
3	CAL3	5	1	4.400	4.68	6.5	02/03/2023	13:27
4	CAL4	10	1	9.440	10.10	-1	02/03/2023	13:30
5	CAL5	20	1	18.500	19.85	0.7	02/03/2023	13:34
6	CAL6	40	1	37.300	40.08	-0.2	02/03/2023	13:37



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Sohil

Run Number: LB123931

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.430	10.090	02/03/2023	13:40
2	ICB		1	0.138	0.090	02/03/2023	13:44
3	CCV1	10	1	9.380	10.037	02/03/2023	13:47
4	CCB1		1	0.150	0.103	02/03/2023	13:50
5	LB123931BL		1	0.154	0.108	02/03/2023	13:53
6	O1245-02		1	4.040	4.290	02/03/2023	13:57
7	O1245-02DUP		1	4.020	4.268	02/03/2023	14:00
8	CCV2	10	1	9.280	9.929	02/03/2023	14:04
9	CCB2		1	0.156	0.110	02/03/2023	14:07

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TURB-012323

WorkList ID : 166769

Department : Wet-Chemistry

Date : 01-23-2023 13:03:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/19/2023	Water	O1245-01	Turbidity	Cool 4 deg C	CHEM02	QA Of	PT-TURB-WS LB 1230 31	01/17/2023	180.1
01/19/2023	Water	O1245-02	Turbidity	Cool 4 deg C	CHEM02	QA Of	PT-TURB-WS LB 1230 30	01/17/2023	SM2130 B

Date/Time 02/03/23 12:50
 Raw Sample Received by: 12 (QC)
 Raw Sample Relinquished by: SJ (QA Office)

Date/Time 02/03/23 14:20
 Raw Sample Received by: SJ (QA Office)
 Raw Sample Relinquished by: 12 (QC)