



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Rubina

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB123932

Slope : 99.2

BalanceID: WC PH METER-1

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	3.99	02/03/2023	14:30
2	CAL2	1	Water	NA	NA	20.7	6.99	02/03/2023	14:31
3	CAL3	1	Water	NA	NA	20.8	10.02	02/03/2023	14:33
4	ICV	1	Water	NA	NA	20.08	6.99	02/03/2023	14:40
5	CCV1	1	Water	NA	NA	20.08	2.02	02/03/2023	14:44
6	O1391-01	1	Solid	20	20	24.1	8.10	02/03/2023	14:47
7	O1391-01DUP	1	Solid	20	20	24.0	8.11	02/03/2023	14:50
8	O1391-02	1	Solid	20	20	22.7	7.13	02/03/2023	14:54
9	O1416-02	1	Solid	20	20	21.5	5.39	02/03/2023	15:00
10	O1418-04	1	Solid	20	20	21.1	8.77	02/03/2023	15:04
11	O1418-08	1	Solid	20	20	21.6	9.22	02/03/2023	15:06
12	O1418-12	1	Solid	20	20	21.9	8.77	02/03/2023	15:10
13	O1418-16	1	Solid	20	20	21.6	8.92	02/03/2023	15:13
14	CCV2	1	Water	NA	NA	20.8	12.3	02/03/2023	15:16

WORKLIST(Hardcopy Internal Chain)

LB123P32

WorkList Name : CORR-020323

WorkList ID : 167112

Department : Wet-Chemistry

Date : 02-03-2023 14:04:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/01/2023	Solid	O1391-01	Corrosivity	Cool 4 deg C	PSEG03	E51	8288	02/01/2023	9045D
02/01/2023	Solid	O1391-02	Corrosivity	Cool 4 deg C	PSEG03	E51	COMP1-0201	02/01/2023	9045D
02/02/2023	Solid	O1416-02	Corrosivity	Cool 4 deg C	PORT06	F32	B-01-SB01	02/02/2023	9045D
02/02/2023	Solid	O1418-04	Corrosivity	Cool 4 deg C	RUTW01	E52	JPP-16-020223	02/02/2023	9045D
02/02/2023	Solid	O1418-08	Corrosivity	Cool 4 deg C	RUTW01	E52	JPP-18A-020223	02/02/2023	9045D
02/02/2023	Solid	O1418-12	Corrosivity	Cool 4 deg C	RUTW01	E52	JPP-20-020223	02/02/2023	9045D
02/02/2023	Solid	O1418-16	Corrosivity	Cool 4 deg C	RUTW01	E52	JPP-22-020223	02/02/2023	9045D

Date/Time 02/03/23 14:10
 Raw Sample Received by: 12(51)
 Raw Sample Relinquished by: RM CWC

Date/Time 02/03/23 15:25
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: 12(51)