

LB123993


7471B

INSTRUMENT ID: CVL

Linear

 μ Abs.:

41091



A= 0.0000e+000

B= 2.4338e-004

C= -4.7233e-002

Rho= 0.9998818

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01)
0.0	0.000	-0.027	-0.027	85	0.000	85					-
0.2	0.200	0.146	-0.054	796	0.0 %	796					-27
2.5	2.500	2.546	0.046	10657	0.0 %	10657					2
5.0	5.000	5.104	0.104	21166	0.0 %	21166					2
7.5	7.500	7.476	-0.024	30911	0.0 %	30911					0
10.0	10.000	9.954	-0.046	41091	0.0 %	41091					0

LB123933 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 03 Feb 2023 15:00:30

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	85	-	0.0000	7471B	PPB	03 Feb 2023 15:03:17	S	Std
0.2 - 1	50.2	796	-	0.2000	7471B	PPB	03 Feb 2023 15:05:34	S	Std
2.5 - 1	52.5	10657	-	2.5000	7471B	PPB	03 Feb 2023 15:07:52	S	Std
5.0 - 1	55	21166	-	5.0000	7471B	PPB	03 Feb 2023 15:10:08	S	Std
7.5 - 1	57.5	30911	-	7.5000	7471B	PPB	03 Feb 2023 15:12:25	S	Std
10.0 - 1	510	41091	-	10.0000	7471B	PPB	03 Feb 2023 15:17:20	S	Std
ICV33 - 1	ICV33	17660	4.2509	-7471B	PPB	03 Feb 2023 15:23:51	U	SMPL	
ICB33 - 1	ICB33	55	-0.0338	-7471B	PPB	03 Feb 2023 15:26:06	U	SMPL	
CCV79 - 1	CCV79	19573	4.7165	-7471B	PPB	03 Feb 2023 15:30:57	U	SMPL	
CCB79 - 1	CCB79	25	-0.0411	-7471B	PPB	03 Feb 2023 15:33:12	U	SMPL	
CRA - 1	CRA	800	0.1475	-7471B	PPB	03 Feb 2023 15:35:27	U	SMPL	
HighStd - 1	HighStd	42041	10.1847	-7471B	PPB	03 Feb 2023 15:37:42	U	SMPL	
ChkStd - 1	ChkStd	29101	7.0354	-7471B	PPB	03 Feb 2023 15:39:58	U	SMPL	
PB150671BL - 1	PBS	-40	-0.0570	-7471B	PPB	03 Feb 2023 15:42:13	U	SMPL	
PB150671BS - 1	LCSS	16124	3.8770	-7471B	PPB	03 Feb 2023 15:44:30	U	SMPL	
O1386-03 - 1	JPP-9-013123	33679	8.1496	-7471B	PPB	03 Feb 2023 15:46:46	U	SMPL	
O1386-03DUP - 1	JPP-9-013123DUP	29419	7.1128	-7471B	PPB	03 Feb 2023 15:49:03	U	SMPL	
O1386-03MS - 1	JPP-9-013123MS	45270	10.9706	-7471B	PPB	03 Feb 2023 15:51:20	U	SMPL	
O1386-03MSD - 1	JPP-9-013123MSD	45889	11.1213	-7471B	PPB	03 Feb 2023 15:53:38	U	SMPL	
O1386-07 - 1	JPP-6-013123	113427	27.5587	-7471B	PPB	03 Feb 2023 15:55:53	U	SMPL	
CCV80 - 1	CCV80	21784	5.2546	-7471B	PPB	03 Feb 2023 15:58:09	U	SMPL	
CCB80 - 1	CCB80	41	-0.0373	-7471B	PPB	03 Feb 2023 16:00:26	U	SMPL	
O1386-11 - 1	JPP-14-013123	13054	3.1299	-7471B	PPB	03 Feb 2023 16:02:42	U	SMPL	
O1386-15 - 1	JPP-15-013123	1858	0.4050	-7471B	PPB	03 Feb 2023 16:04:57	U	SMPL	
O1390-01 - 1	STOCKPILE-1	4201	0.9752	-7471B	PPB	03 Feb 2023 16:07:13	U	SMPL	
O1390-04 - 1	STOCKPILE-2	7326	1.7358	-7471B	PPB	03 Feb 2023 16:09:28	U	SMPL	
O1401-01 - 1	OK-1-212023	13341	3.1997	-7471B	PPB	03 Feb 2023 16:11:46	U	SMPL	
O1403-01 - 1	NPWAYNE-SOIL-0201	51531	12.4944	-7471B	PPB	03 Feb 2023 16:14:03	U	SMPL	
O1404-01 - 1	NPHACK-SOIL-0201	5145	1.2050	-7471B	PPB	03 Feb 2023 16:16:19	U	SMPL	
O1407-01 - 1	30984	211	0.0041	-7471B	PPB	03 Feb 2023 16:18:36	U	SMPL	
O1410-01 - 1	72-12013	13214	3.1688	-7471B	PPB	03 Feb 2023 16:20:53	U	SMPL	
O1410-03 - 1	VNJ-225	5198	1.2179	-7471B	PPB	03 Feb 2023 16:23:09	U	SMPL	
CCV81 - 1	CCV81	22195	5.3546	-7471B	PPB	03 Feb 2023 16:25:25	U	SMPL	
CCB81 - 1	CCB81	10	-0.0448	-7471B	PPB	03 Feb 2023 16:27:40	U	SMPL	
O1413-01 - 1	TR-5-020223	3776	0.8718	-7471B	PPB	03 Feb 2023 16:29:56	U	SMPL	
PB150672BL - 1	PBS	65	-0.0314	-7471B	PPB	03 Feb 2023 16:32:12	U	SMPL	
PB150672BS - 1	LCSS	18402	4.4315	-7471B	PPB	03 Feb 2023 16:34:28	U	SMPL	
O1409-01 - 1	38414	3385	0.7766	-7471B	PPB	03 Feb 2023 16:36:44	U	SMPL	
O1414-01 - 1	SASP2-123-2(0-5)	4527	1.0546	-7471B	PPB	03 Feb 2023 16:39:00	U	SMPL	
O1414-02 - 1	SASP2-123-2(20-30)	125	-0.0168	-7471B	PPB	03 Feb 2023 16:41:17	U	SMPL	
O1414-03 - 1	SASP2-123-2(30-40)	105	-0.0217	-7471B	PPB	03 Feb 2023 16:43:33	U	SMPL	
O1414-04 - 1	SASP2-123-2(55-62)	223	0.0070	-7471B	PPB	03 Feb 2023 16:45:50	U	SMPL	
O1414-05 - 1	SASP2-123-2(65-72)	484	0.0706	-7471B	PPB	03 Feb 2023 16:48:08	U	SMPL	
O1414-06 - 1	SASP2-123-2(75-80)	257	0.0153	-7471B	PPB	03 Feb 2023 16:50:23	U	SMPL	
CCV82 - 1	CCV82	21275	5.1307	-7471B	PPB	03 Feb 2023 16:52:41	U	SMPL	
CCB82 - 1	CCB82	15	-0.0436	-7471B	PPB	03 Feb 2023 16:54:57	U	SMPL	
O1416-01 - 1	B-01-SB01	7562	1.7932	-7471B	PPB	03 Feb 2023 16:57:12	U	SMPL	
O1418-03 - 1	JPP-16-020223	97460	23.6726	-7471B	PPB	03 Feb 2023 16:59:28	U	SMPL	
O1418-03DUP - 1	JPP-16-020223DUP	110517	26.8504	-7471B	PPB	03 Feb 2023 17:01:43	U	SMPL	
O1418-03MS - 1	JPP-16-020223MS	123217	29.9414	-7471B	PPB	03 Feb 2023 17:04:00	U	SMPL	
O1418-03MSD - 1	JPP-16-020223MSD	145288	35.3130	-7471B	PPB	03 Feb 2023 17:06:17	U	SMPL	
O1418-07 - 1	JPP-18A-020223	0	-0.0472	-7471B	PPB	03 Feb 2023 17:08:35	U	SMPL	
O1418-03MSD - 1	JPP-16-020223MSD	123281	29.9570	-7471B	PPB	03 Feb 2023 17:08:53	U	SMPL	
O1418-07 - 1	JPP-18A-020223	76827	18.6510	-7471B	PPB	03 Feb 2023 17:11:09	U	SMPL	
O1418-11 - 1	JPP-20-020223	182975	44.4853	-7471B	PPB	03 Feb 2023 17:13:28	U	SMPL	
O1418-15 - 1	JPP-22-020223	70108	17.0157	-7471B	PPB	03 Feb 2023 17:15:45	U	SMPL	
CCV83 - 1	CCV83	20421	4.9228	-7471B	PPB	03 Feb 2023 17:18:04	U	SMPL	
CCB83 - 1	CCB83	102	-0.0224	-7471B	PPB	03 Feb 2023 17:20:20	U	SMPL	
O1386-03LX5 - 1		6455	1.5238	-7471B	PPB	03 Feb 2023 17:22:36	U	SMPL	
O1386-03A - 1		49157	11.9166	-7471B	PPB	03 Feb 2023 17:24:52	U	SMPL	
O1418-03LX5 - 1		18892	4.5507	-7471B	PPB	03 Feb 2023 17:27:08	U	SMPL	
O1418-03A - 1		51809	12.5621	-7471B	PPB	03 Feb 2023 17:29:24	U	SMPL	
CCV84 - 1	CCV84	20493	4.9404	-7471B	PPB	03 Feb 2023 17:45:20	U	SMPL	
CCB84 - 1	CCB84	-77	-0.0660	-7471B	PPB	03 Feb 2023 17:47:36	U	SMPL	
O1386-07DLX10 - 1		11261	2.6935	-7471B	PPB	03 Feb 2023 17:49:53	U	SMPL	
O1403-01DLX2 - 1		25948	6.2680	-7471B	PPB	03 Feb 2023 17:52:09	U	SMPL	
CCV85 - 1	CCV85	20892	5.0375	-7471B	PPB	03 Feb 2023 17:54:25	U	SMPL	
CCB85 - 1	CCB85	-64	-0.0628	-7471B	PPB	03 Feb 2023 17:56:42	U	SMPL	