



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB123952

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	02/06/2023	12:35
2	O1419-01		1	12	100.00	>212.0	>212.0	02/06/2023	13:05
3	O1419-01DUP		1	12	100.00	>212.0	>212.0	02/06/2023	13:35
4	O1437-01		1	14	100.00	>212.0	>212.0	02/06/2023	14:05
5	O1437-02		1	13	100.00	>212.0	>212.0	02/06/2023	14:35
6	O1437-03		1	12	100.00	>212.0	>212.0	02/06/2023	15:05
7	O1437-04		1	13	100.00	>212.0	>212.0	02/06/2023	15:35
8	O1440-01		1	14	100.00	>212.0	>212.0	02/06/2023	16:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66123952

WorkList Name : fp2-6

WorkList ID : 167124

Department : Wet-Chemistry

Date : 02-06-2023 08:15:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/08/2023	Water	O1419-01	Flash Point	Cool 4 deg C	PSEG03	E52	DRUM-COMP	02/03/2023	1010B
02/10/2023	Water	O1437-01	Flash Point	Cool 4 deg C	PSEG03	F11	31881	02/03/2023	1010B
02/10/2023	Water	O1437-02	Flash Point	Cool 4 deg C	PSEG03	F11	31962	02/03/2023	1010B
02/10/2023	Water	O1437-03	Flash Point	Cool 4 deg C	PSEG03	F11	31956	02/03/2023	1010B
02/10/2023	Water	O1437-04	Flash Point	Cool 4 deg C	PSEG03	F11	2423	02/03/2023	1010B
02/08/2023	Water	O1440-01	Flash Point	Cool 4 deg C	PSEG03	F11	MH-1	02/03/2023	1010B

Date/Time 02/06/2023 12:30
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: JH SW

Date/Time 02/06/2023 16:45
 Raw Sample Received by: CBC
 Raw Sample Relinquished by: RM CWC