



PERCENT SOLID

Supervisor: Iwona
Analyst: jaswal
Date: 2/17/2023

OVENTEMP IN Celsius(°C): 108
Time IN: 03:00
In Date: 02/07/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 2.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 108
Time OUT: 12:00
Out Date: 02/08/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB123958

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O1451-01	MH0DK5	1	1.15	8.83	9.98	9.72	97.1	
O1451-02	MH0DK6	2	1.14	8.76	9.9	9.6	96.6	
O1451-03	MH0DK7	3	1.13	8.84	9.97	9.58	95.6	
O1451-04	MH0DL0	4	1.15	8.76	9.91	9.5	95.3	
O1451-05	MH0DL1	5	1.13	8.82	9.95	9.43	94.1	
O1451-06	MH0DL2	6	1.13	8.84	9.97	9.73	97.3	
O1451-07	MH0DL5	7	1.15	8.72	9.87	8.89	88.8	
O1451-08	MH0DL6	8	1.14	8.81	9.95	9.12	90.6	
O1451-09	MH0DL7	9	1.12	8.74	9.86	9.33	93.9	
O1451-10	MH0DM5	10	1.14	8.79	9.93	9.59	96.1	
O1451-11	MH0DM6	11	1.15	8.81	9.96	9.6	95.9	
O1451-12	MH0DM7	12	1.15	8.70	9.85	9.49	95.9	
O1451-13	MH0DN0	13	1.13	8.74	9.87	9.67	97.7	
O1451-14	MH0DN1	14	1.15	8.75	9.9	9.74	98.2	
O1451-15	MH0DN5	15	1.14	8.82	9.96	9.68	96.8	
O1451-16	MH0DN6	16	1.13	8.85	9.98	9.62	95.9	
O1451-17	MH0DN7	17	1.15	8.72	9.87	9.55	96.3	
O1451-18	MH0DP0	18	1.14	8.72	9.86	9.48	95.6	
O1451-19	MH0DP0D	19	1.14	8.72	9.86	9.48	95.6	
O1451-20	MH0DP0S	20	1.14	8.72	9.86	9.48	95.6	
O1451-21	MH0DP1	21	1.13	8.73	9.86	9.69	98.1	
O1451-22	MH0DP2	22	1.14	8.75	9.89	9.71	97.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$