



PERCENT SOLID

Supervisor: Iwona
Analyst: jaswal
Date: 2/17/2023

OVENTEMP IN Celsius(°C): 108
Time IN: 03:15
In Date: 02/07/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 2.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 108
Time OUT: 12:15
Out Date: 02/08/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB123959

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O1452-01	MH0DP5	1	1.13	8.84	9.97	9.73	97.3	
O1452-02	MH0DP6	2	1.14	8.81	9.95	9.88	99.2	
O1452-03	MH0DQ0	3	1.13	8.85	9.98	9.58	95.5	
O1452-04	MH0DQ1	4	1.14	8.72	9.86	9.49	95.8	
O1452-05	MH0DQ5	5	1.13	8.83	9.96	9.26	92.1	
O1452-06	MH0DQ6	6	1.15	8.67	9.82	8.47	84.4	
O1452-07	MH0DW0	7	1.12	8.85	9.97	9.41	93.7	
O1452-08	MH0DW1	8	1.15	8.77	9.92	9.31	93.0	
O1452-09	MH0DW2	9	1.13	8.75	9.88	9.63	97.1	
O1452-10	MH0E81	10	1.12	8.85	9.97	9.78	97.9	
O1452-11	MH0E82	11	1.12	8.69	9.81	9.61	97.7	
O1452-12	MH0E83	12	1.14	8.82	9.96	9.74	97.5	
O1452-13	MH0E86	13	1.12	8.76	9.88	9.67	97.6	
O1452-14	MH0E87	14	1.13	8.77	9.9	9.65	97.1	
O1452-15	MH0E88	15	1.14	8.71	9.85	9.72	98.5	
O1452-16	MH0E91	16	1.14	8.84	9.98	9.83	98.3	
O1452-17	MH0E91D	17	1.14	8.84	9.98	9.83	98.3	
O1452-18	MH0E91S	18	1.14	8.84	9.98	9.83	98.3	
O1452-19	MH0E92	19	1.14	8.84	9.98	9.83	98.3	
O1452-20	MH0E93	20	1.14	8.83	9.97	9.8	98.1	
O1452-21	MH0E96	21	1.12	8.83	9.95	9.85	98.9	
O1452-22	MH0E97	22	1.14	8.84	9.98	9.86	98.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$