

LB123966


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

52172



A= 0.0000e+000

B= 1.9228e-004

C= 2.5680e-002

Rho= 0.9998955

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01)
0.0	0.000	0.075	0.075	258	0.000	258					-19
0.2	0.200	0.163	-0.037	712	0.0 %	712					-1
2.5	2.500	2.468	-0.032	12701	0.0 %	12701					0
5.0	5.000	5.012	0.012	25932	0.0 %	25932					-1
7.5	7.500	7.425	-0.075	38485	0.0 %	38485					1
10.0	10.000	10.057	0.057	52172	0.0 %	52172					

LB123966 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 07 Feb 2023 14:45:21

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	258	-	0.000	7471B	PPB	07 Feb 2023 14:48:15	S	Std
0.2 - 1	SD	712	-	0.200	7471B	PPB	07 Feb 2023 14:50:33	S	Std
2.5 - 1	SD	12701	-	2.500	7471B	PPB	07 Feb 2023 14:52:50	S	Std
5.0 - 1	SD	25932	-	5.000	7471B	PPB	07 Feb 2023 14:57:14	S	Std
7.5 - 1	SD	38485	-	7.500	7471B	PPB	07 Feb 2023 14:59:31	S	Std
10.0 - 1	SD	52172	-	10.000	7471B	PPB	07 Feb 2023 15:07:06	S	Std
ICV36 - 1	ICV36	22181	4.2906	-	7471B	PPB	07 Feb 2023 15:09:56	U	SMPL
ICB36 - 1	ICB36	12	0.0280	-	7471B	PPB	07 Feb 2023 15:12:11	U	SMPL
CCV95 - 1	CCV95	25864	4.9987	-	7471B	PPB	07 Feb 2023 15:14:25	U	SMPL
CCB95 - 1	CCB95	-8	0.0241	-	7471B	PPB	07 Feb 2023 15:16:41	U	SMPL
CRA - 1	CRA	714	0.1630	-	7471B	PPB	07 Feb 2023 15:18:55	U	SMPL
HighStd - 1	HighStd	51870	9.9991	-	7471B	PPB	07 Feb 2023 15:21:11	U	SMPL
ChkStd - 1	ChkStd	36007	6.9490	-	7471B	PPB	07 Feb 2023 15:23:26	U	SMPL
PB150702BL - 1	PB150702BL	-37	0.0186	-	7471B	PPB	07 Feb 2023 15:25:42	U	SMPL
PB150702BS - 1	PB150702BS	22155	4.2856	-	7471B	PPB	07 Feb 2023 15:27:59	U	SMPL
O1428-02 - 1	1574-1577	581	0.1374	-	7471B	PPB	07 Feb 2023 15:30:15	U	SMPL
O1428-03 - 1	348	5591	1.1007	-	7471B	PPB	07 Feb 2023 15:32:31	U	SMPL
O1432-01 - 1	DOMINICK-STREET-UTLITY-RELOCATION-EXCAVATION	25402	4.9099	-	7471B	PPB	07 Feb 2023 15:34:49	U	SMPL
O1442-01 - 1	B-01-SB02	1174	0.2514	-	7471B	PPB	07 Feb 2023 15:37:06	U	SMPL
O1445-03 - 1	JPP-26-020323	15918	3.0863	-	7471B	PPB	07 Feb 2023 15:39:21	U	SMPL
CCV96 - 1	CCV96	25817	4.9897	-	7471B	PPB	07 Feb 2023 15:41:37	U	SMPL
CCB96 - 1	CCB96	-90	0.0084	-	7471B	PPB	07 Feb 2023 15:43:53	U	SMPL
O1445-07 - 1	JPP-29-020323	32235	6.2237	-	7471B	PPB	07 Feb 2023 15:46:08	U	SMPL
O1445-11 - 1	JPP-34-020323	6143	1.2068	-	7471B	PPB	07 Feb 2023 15:48:24	U	SMPL
O1460-01 - 1	OR-2-020623	1470	0.3083	-	7471B	PPB	07 Feb 2023 15:50:41	U	SMPL
O1462-01 - 1	COMP-1	801	0.1797	-	7471B	PPB	07 Feb 2023 15:52:56	U	SMPL
O1462-08 - 1	COMP-2	1027	0.2231	-	7471B	PPB	07 Feb 2023 15:55:13	U	SMPL
O1462-15 - 1	COMP-3	1134	0.2437	-	7471B	PPB	07 Feb 2023 15:57:29	U	SMPL
O1462-22 - 1	COMP-4	921	0.2028	-	7471B	PPB	07 Feb 2023 15:59:45	U	SMPL
O1463-05 - 1	COMP-5	908	0.2003	-	7471B	PPB	07 Feb 2023 16:02:03	U	SMPL
O1463-12 - 1	COMP-6	2127	0.4347	-	7471B	PPB	07 Feb 2023 16:04:20	U	SMPL
O1463-19 - 1	COMP-7	924	0.2033	-	7471B	PPB	07 Feb 2023 16:06:35	U	SMPL
CCV97 - 1	CCV97	25971	5.0193	-	7471B	PPB	07 Feb 2023 16:08:51	U	SMPL
CCB97 - 1	CCB97	-47	0.0166	-	7471B	PPB	07 Feb 2023 16:11:07	U	SMPL
O1464-03 - 1	COMP-8	764	0.1726	-	7471B	PPB	07 Feb 2023 16:13:23	U	SMPL
O1465-01 - 1	COMP-1	34536	6.6661	-	7471B	PPB	07 Feb 2023 16:15:39	U	SMPL
O1466-01 - 1	A-1	790	0.1776	-	7471B	PPB	07 Feb 2023 16:17:54	U	SMPL
O1466-02 - 1	A-2	2550	0.5160	-	7471B	PPB	07 Feb 2023 16:20:10	U	SMPL
O1466-03 - 1	A-3	672	0.1549	-	7471B	PPB	07 Feb 2023 16:22:26	U	SMPL
O1466-03DUP - 1	A-3DUP	738	0.1676	-	7471B	PPB	07 Feb 2023 16:24:43	U	SMPL
O1466-03MS - 1	A-3MS	22301	4.3136	-	7471B	PPB	07 Feb 2023 16:27:00	U	SMPL
O1466-03MSD - 1	A-3MSD	21746	4.2069	-	7471B	PPB	07 Feb 2023 16:29:17	U	SMPL
O1464-03LX5 - 1		77	0.0405	-	7471B	PPB	07 Feb 2023 16:31:34	U	SMPL
O1464-03A - 1		22276	4.3088	-	7471B	PPB	07 Feb 2023 16:33:50	U	SMPL
CCV98 - 1	CCV98	25951	5.0154	-	7471B	PPB	07 Feb 2023 16:36:06	U	SMPL
CCB98 - 1	CCB98	-57	0.0147	-	7471B	PPB	07 Feb 2023 16:38:22	U	SMPL