



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Iwona

Parameter: Corrosivity

Supervisor Review By : Sohil

Run Number: LB123976

Slope : 99.2

BalanceID: WC PH METER-1

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	02/07/2023	14:30
2	CAL2	1	Water	NA	NA	20.8	7.01	02/07/2023	14:31
3	CAL3	1	Water	NA	NA	20.7	10.02	02/07/2023	14:32
4	ICV	1	Water	NA	NA	20.8	6.99	02/07/2023	14:34
5	CCV1	1	Water	NA	NA	20.9	2.02	02/07/2023	14:35
6	O1445-04	1	Solid	20	20	21.5	9.72	02/07/2023	14:37
7	O1445-04DUP	1	Solid	20	20	21.4	9.71	02/07/2023	14:38
8	O1445-08	1	Solid	20	20	21.8	8.75	02/07/2023	14:40
9	O1445-12	1	Solid	20	20	21.7	10.95	02/07/2023	14:43
10	O1462-02	1	Solid	20	20	22.4	8.90	02/07/2023	14:45
11	O1462-09	1	Solid	20	20	22.1	8.93	02/07/2023	14:48
12	O1462-16	1	Solid	20	20	22.3	8.84	02/07/2023	14:50
13	O1462-23	1	Solid	20	20	22.3	8.90	02/07/2023	14:53
14	O1463-06	1	Solid	20	20	22.6	8.83	02/07/2023	14:55
15	O1463-13	1	Solid	20	20	22.5	8.89	02/07/2023	14:57
16	CCV2	1	Water	NA	NA	20.8	12.02	02/07/2023	14:59
17	O1463-20	1	Solid	20	20	22.6	8.85	02/07/2023	15:02
18	O1464-04	1	Solid	20	20	22.6	8.75	02/07/2023	15:05
19	O1465-02	1	Solid	20	20	21.6	8.22	02/07/2023	15:08
20	CCV3	1	Water	NA	NA	20.8	2.01	02/07/2023	15:10

WORKLIST(Hardcopy Internal Chain)

CB123476

WorkList Name : CORR-020723

WorkList ID : 167170

Department : Wet-Chemistry

Date : 02-07-2023 09:57:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/03/2023	Solid	O1445-04	Corrosivity	Cool 4 deg C	RUTW01	F11	JPP-26-020323	02/03/2023	9045D
02/03/2023	Solid	O1445-08	Corrosivity	Cool 4 deg C	RUTW01	F11	JPP-29-020323	02/03/2023	9045D
02/03/2023	Solid	O1445-12	Corrosivity	Cool 4 deg C	RUTW01	F11	JPP-34-020323	02/03/2023	9045D
02/06/2023	Solid	O1462-02	Corrosivity	Cool 4 deg C	SOUN03	F12	COMP-1	02/06/2023	9045D
02/06/2023	Solid	O1462-09	Corrosivity	Cool 4 deg C	SOUN03	F12	COMP-2	02/06/2023	9045D
02/06/2023	Solid	O1462-16	Corrosivity	Cool 4 deg C	SOUN03	F12	COMP-3	02/06/2023	9045D
02/06/2023	Solid	O1462-23	Corrosivity	Cool 4 deg C	SOUN03	F12	COMP-4	02/06/2023	9045D
02/06/2023	Solid	O1463-06	Corrosivity	Cool 4 deg C	SOUN03	F12	COMP-5	02/06/2023	9045D
02/06/2023	Solid	O1463-13	Corrosivity	Cool 4 deg C	SOUN03	F12	COMP-6	02/06/2023	9045D
02/06/2023	Solid	O1463-20	Corrosivity	Cool 4 deg C	SOUN03	F12	COMP-7	02/06/2023	9045D
02/06/2023	Solid	O1464-04	Corrosivity	Cool 4 deg C	SOUN03	F12	COMP-8	02/06/2023	9045D
02/06/2023	Solid	O1465-02	Corrosivity	Cool 4 deg C	SOUN03	F12	COMP-1	02/06/2023	9045D

Date/Time 02/07/23 14:10
 Raw Sample Received by: 12/07/23
 Raw Sample Relinquished by: OLS SFR

Date/Time 02/07/23 15:20
 Raw Sample Received by: OLS SFR
 Raw Sample Relinquished by: 12/07/23