



Analytical Summary Report

Analysis Method: 9040C
Parameter: Corrosivity
Run Number: LB123977

Analyst By : Iwona
Supervisor Review By : Sohil
Slope : 99.2
pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	02/07/2023	14:30
2	CAL2	1	Water	NA	NA	20.8	7.01	02/07/2023	14:31
3	CAL3	1	Water	NA	NA	20.7	10.02	02/07/2023	14:32
4	ICV	1	Water	NA	NA	20.8	6.99	02/07/2023	14:34
5	CCV1	1	Water	NA	NA	20.7	2.02	02/07/2023	15:15
6	O1419-01	1	Water	NA	NA	20.9	9.49	02/07/2023	15:18
7	O1419-01DUP	1	Water	NA	NA	20.9	9.48	02/07/2023	15:20
8	CCV2	1	Water	NA	NA	20.8	12.01	02/07/2023	15:23

WORKLIST(Hardcopy Internal Chain)

L13123877

WorkList Name : CORR-020723-2

WorkList ID : 167171

Department : Wet-Chemistry

Date : 02-07-2023 09:57:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/03/2023	Water	O1419-01	Corrosivity	Cool 4 deg C	PSEG03	E52	DRUM-COMP	02/03/2023	9040C

Date/Time 02/07/23 15:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 02/07/23 15:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]