



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Iwona

Parameter: pH

Supervisor Review By : Sohil

Run Number: LB123978

Slope : 99.2

BalanceID: WC PH METER-1

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	02/07/2023	14:30
2	CAL2	1	Water	NA	NA	20.8	7.01	02/07/2023	14:31
3	CAL3	1	Water	NA	NA	20.7	10.02	02/07/2023	14:32
4	ICV	1	Water	NA	NA	20.8	6.99	02/07/2023	14:34
5	CCV1	1	Water	NA	NA	20.8	2.01	02/07/2023	16:45
6	O1449-01	1	Solid	20	20	22.6	9.17	02/07/2023	16:47
7	O1449-01DUP	1	Solid	20	20	22.6	9.16	02/07/2023	16:50
8	O1470-01	1	Solid	20	20	22.7	7.06	02/07/2023	16:52
9	CCV2	1	Water	NA	NA	20.08	12.01	02/07/2023	16:54

WORKLIST(Hardcopy Internal Chain)

LB123078

WorkList Name : pH-020723-2 WorkList ID : 167169 Department : Wet-Chemistry Date : 02-07-2023 09:56:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/06/2023	Solid	O1449-01	pH	Cool 4 deg C	PSEG03	F12	34530	02/06/2023	9045D
02/07/2023	Solid	O1470-01	pH	Cool 4 deg C	PSEG03	F21	NWB-1823	02/07/2023	9045D

Date/Time 02/07/23 16:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 02/07/23 17:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]