



Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB123979

Analyst By : Iwona
Supervisor Review By : Sohil
Slope : 99.2
pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	02/07/2023	14:30
2	CAL2	1	Water	NA	NA	20.8	7.01	02/07/2023	14:31
3	CAL3	1	Water	NA	NA	20.7	10.02	02/07/2023	14:32
4	ICV	1	Water	NA	NA	20.8	6.99	02/07/2023	14:34
5	CCV1	1	Water	NA	NA	20.8	2.02	02/07/2023	17:00
6	O1437-01	1	Water	NA	NA	22.2	8.01	02/07/2023	17:02
7	O1437-01DUP	1	Water	NA	NA	22.2	8.02	02/07/2023	17:03
8	O1437-02	1	Water	NA	NA	22.3	7.53	02/07/2023	17:05
9	O1437-03	1	Water	NA	NA	22.5	8.38	02/07/2023	17:07
10	O1437-04	1	Water	NA	NA	22.4	7.03	02/07/2023	17:09
11	O1440-01	1	Water	NA	NA	21.2	7.06	02/07/2023	17:11
12	O1441-01	1	Water	NA	NA	22.4	7.36	02/07/2023	17:14
13	O1447-01	1	Water	NA	NA	20.9	7.24	02/07/2023	17:16
14	O1456-01	1	Water	NA	NA	22.5	6.03	02/07/2023	17:18
15	O1456-02	1	Water	NA	NA	22.6	5.39	02/07/2023	17:20
16	CCV2	1	Water	NA	NA	20.8	12.02	02/07/2023	17:22
17	O1467-01	1	Water	NA	NA	22.6	6.09	02/07/2023	17:24
18	O1468-01	1	Water	NA	NA	21.2	5.05	02/07/2023	17:26
19	CCV3	1	Water	NA	NA	20.8	2.02	02/07/2023	17:28

WORKLIST(Hardcopy Internal Chain)

L13123879

WorkList Name : pH-020723 WorkList ID : 167168 Department : Wet-Chemistry Date : 02-07-2023 09:56:46

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/07/2023	Water	O1468-01	pH	Cool 4 deg C	STER05	F41	POOL-WATER	02/07/2023	9040C
02/03/2023	Water	O1437-01	pH	Cool 4 deg C	PSEG03	F11	31881	02/03/2023	9040C
02/03/2023	Water	O1437-02	pH	Cool 4 deg C	PSEG03	F11	31962	02/03/2023	9040C
02/03/2023	Water	O1437-03	pH	Cool 4 deg C	PSEG03	F11	31956	02/03/2023	9040C
02/03/2023	Water	O1437-04	pH	Cool 4 deg C	PSEG03	F11	2423	02/03/2023	9040C
02/03/2023	Water	O1440-01	pH	Cool 4 deg C	PSEG03	F11	MH-1	02/03/2023	9040C
02/03/2023	Water	O1441-01	pH	Cool 4 deg C	PSEG03	F11	CMS-OWS-OUTLET	02/03/2023	9040C
02/06/2023	Water	O1447-01	pH	Cool 4 deg C	PSEG03	F12	T2	02/06/2023	9040C
02/06/2023	Water	O1456-01	pH	Cool 4 deg C	PSEG03	F12	34150	02/06/2023	9040C
02/06/2023	Water	O1456-02	pH	Cool 4 deg C	PSEG03	F12	34151	02/06/2023	9040C
02/07/2023	Water	O1467-01	pH	Cool 4 deg C	PSEG03	F21	33906	02/07/2023	9040C

Date/Time 02/07/23 16:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 02/07/23 17:40
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]