

LB123980


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

29777



A= 0.0000e+000

B= 3.3049e-004

C= -8.6370e-003

Rho= 0.9995662

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.0	0.000	0.004	0.004	38	0.000	38					10
0.05	0.050	0.055	0.005	193	0.0 %	193	0				-5
0.2	0.200	0.190	-0.010	601	0.0 %	601					-3
2.5	2.500	2.417	-0.083	7339	0.0 %	7339					1
5.0	5.000	5.034	0.034	15259	0.0 %	15259					3
7.5	7.500	7.718	0.218	23378	0.0 %	23378					-2
10.0	10.000	9.832	-0.168	29777	0.0 %	29777					

LB123980 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 08 Feb 2023 12:11:12

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	38	-	0.0000	245.1	PPB	08 Feb 2023 12:13:53	S	Std
0.05 - 1	50.05	193	-	0.0500	245.1	PPB	08 Feb 2023 12:16:11	S	Std
0.2 - 1	50.2	601	-	0.2000	245.1	PPB	08 Feb 2023 12:18:28	S	Std
2.5 - 1	52.5	7339	-	2.5000	245.1	PPB	08 Feb 2023 12:20:45	S	Std
5.0 - 1	55	15259	-	5.0000	245.1	PPB	08 Feb 2023 12:23:01	S	Std
7.5 - 1	57.5	23378	-	7.5000	245.1	PPB	08 Feb 2023 12:25:17	S	Std
10.0 - 1	510	29777	-	10.0000	245.1	PPB	08 Feb 2023 12:35:07	S	Std
ICV37 - 1	ICV37	12494	4.1205	-245.1	PPB	PPB	08 Feb 2023 12:40:32	U	SMPL
ICB37 - 1	ICB37	18	-0.0027	-245.1	PPB	PPB	08 Feb 2023 12:42:47	U	SMPL
CCV01 - 1	CCV01	15234	5.0260	-245.1	PPB	PPB	08 Feb 2023 12:45:02	U	SMPL
CCB01 - 1	CCB01	-8	-0.0113	-245.1	PPB	PPB	08 Feb 2023 12:47:17	U	SMPL
CRA - 1	CRA	164	0.0456	-245.1	PPB	PPB	08 Feb 2023 12:49:32	U	SMPL
HighStd - 1	HighStd	29722	9.8141	-245.1	PPB	PPB	08 Feb 2023 12:51:47	U	SMPL
ChkStd - 1	ChkStd	20990	6.9283	-245.1	PPB	PPB	08 Feb 2023 12:54:03	U	SMPL
PB150736BL - 1	PBW	-35	-0.0202	-245.1	PPB	PPB	08 Feb 2023 12:56:20	U	SMPL
PB150736BS - 1	LCSW	11643	3.8392	-245.1	PPB	PPB	08 Feb 2023 12:58:36	U	SMPL
O1375-01 - 1	MANHOLE	13	-0.0043	-245.1	PPB	PPB	08 Feb 2023 13:00:52	U	SMPL
O1375-01DUP - 1	MANHOLEDUP	73	0.0155	-245.1	PPB	PPB	08 Feb 2023 13:03:08	U	SMPL
O1375-01MS - 1	MANHOLEMS	10651	3.5114	-245.1	PPB	PPB	08 Feb 2023 13:05:26	U	SMPL
O1375-01MSD - 1	MANHOLEMSD	10735	3.5392	-245.1	PPB	PPB	08 Feb 2023 13:07:43	U	SMPL
O1384-01 - 1	EFF-DAY-1-METALS	15	-0.0037	-245.1	PPB	PPB	08 Feb 2023 13:09:58	U	SMPL
CCV02 - 1	CCV02	14965	4.9371	-245.1	PPB	PPB	08 Feb 2023 13:12:14	U	SMPL
CCB02 - 1	CCB02	-3	-0.0096	-245.1	PPB	PPB	08 Feb 2023 13:14:29	U	SMPL
O1457-07 - 1	14B-(1-4)-COMP	2184	0.7131	-245.1	PPB	PPB	08 Feb 2023 13:16:45	U	SMPL
O1469-07 - 1	14B-(1-4)-COMP	2173	0.7095	-245.1	PPB	PPB	08 Feb 2023 13:19:02	U	SMPL
O1471-02 - 1	EFFLUENT-METALS	495	0.1550	-245.1	PPB	PPB	08 Feb 2023 13:21:18	U	SMPL
O1471-07 - 1	EFFLUENT-METALS	479	0.1497	-245.1	PPB	PPB	08 Feb 2023 13:23:33	U	SMPL
O1245-05 - 1	PT-HG-WS	4880	1.6041	-245.1	PPB	PPB	08 Feb 2023 13:28:06	U	SMPL
O1375-01LX5 - 1		-25	-0.0169	-245.1	PPB	PPB	08 Feb 2023 13:34:57	U	SMPL
O1375-01A - 1		12716	4.1939	-245.1	PPB	PPB	08 Feb 2023 13:40:10	U	SMPL
CCV03 - 1	CCV03	15083	4.9761	-245.1	PPB	PPB	08 Feb 2023 13:42:25	U	SMPL
CCB03 - 1	CCB03	8	-0.0060	-245.1	PPB	PPB	08 Feb 2023 13:44:42	U	SMPL