

LB 123 986


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

32334



A= 0.0000e+000

B= 3.0238e-004

C= 9.5575e-002

Rho= 0.9991007

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	6/7/8/9
0.0	0.000	0.099	0.099	11	0.000	11					
0.2	0.200	0.246	0.046	496	0.0 %	496					23
2.5	2.500	2.232	-0.268	7064	0.0 %	7064					-11
5.0	5.000	5.039	0.039	16347	0.0 %	16347					1
7.5	7.500	7.713	0.213	25191	0.0 %	25191					3
10.0	10.000	9.873	-0.127	32334	0.0 %	32334					-1

LB123986 INSTRUMENT IDF : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 08 Feb 2023 15:03:22

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	11	-	0.0000	7470A	PPB	08 Feb 2023 15:08:11	S	Std
0.2 - 1	50.2	496	-	0.2000	7470A	PPB	08 Feb 2023 15:10:28	S	Std
2.5 - 1	52.5	7064	-	2.5000	7470A	PPB	08 Feb 2023 15:12:45	S	Std
5.0 - 1	55	16347	-	5.0000	7470A	PPB	08 Feb 2023 15:15:02	S	Std
7.5 - 1	57.5	25191	-	7.5000	7470A	PPB	08 Feb 2023 15:17:19	S	Std
10.0 - 1	59	32334	-	10.0000	7470A	PPB	08 Feb 2023 15:22:11	S	Std
ICV38 - 1	ICV38	13125	4.0643	-	7470A	PPB	08 Feb 2023 15:27:33	U	SMPL
ICB38 - 1	ICB38	14	0.0998	-	7470A	PPB	08 Feb 2023 15:29:48	U	SMPL
CCV04 - 1	CCV04	16276	5.0171	-	7470A	PPB	08 Feb 2023 15:32:03	U	SMPL
CCB04 - 1	CCB04	1	0.0959	-	7470A	PPB	08 Feb 2023 15:37:40	U	SMPL
CRA - 1	CRA	447	0.2307	-	7470A	PPB	08 Feb 2023 15:39:55	U	SMPL
HighStd - 1	HighStd	32347	9.8766	-	7470A	PPB	08 Feb 2023 15:42:10	U	SMPL
ChkStd - 1	ChkStd	22980	7.0442	-	7470A	PPB	08 Feb 2023 15:44:25	U	SMPL
PB150747BL - 1	PBW	-53	0.0795	-	7470A	PPB	08 Feb 2023 15:46:41	U	SMPL
PB150747BS - 1	LCSW	11204	3.4834	-	7470A	PPB	08 Feb 2023 15:48:57	U	SMPL
O1419-01 - 1	DRUM-COMP	884	0.3629	-	7470A	PPB	08 Feb 2023 15:51:14	U	SMPL
O1419-01DUP - 1	DRUM-COMPDUP	924	0.3750	-	7470A	PPB	08 Feb 2023 15:53:30	U	SMPL
O1419-01MS - 1	DRUM-COMPMS	15818	4.8786	-	7470A	PPB	08 Feb 2023 15:55:49	U	SMPL
O1419-01MSD - 1	DRUM-COMPMSD	13655	4.2245	-	7470A	PPB	08 Feb 2023 15:58:06	U	SMPL
CCV05 - 1	CCV05	16205	4.9956	-	7470A	PPB	08 Feb 2023 16:00:22	U	SMPL
CCB05 - 1	CCB05	12	0.0992	-	7470A	PPB	08 Feb 2023 16:05:53	U	SMPL
O1419-01LX5 - 1		42	0.1083	-	7470A	PPB	08 Feb 2023 16:16:36	U	SMPL
O1419-01A - 1		14427	4.4580	-	7470A	PPB	08 Feb 2023 16:18:51	U	SMPL
PB150718TB - 1	PB150718TB	-41	0.0832	-	7470A	PPB	08 Feb 2023 16:21:06	U	SMPL
CCV06 - 1	CCV06	16274	5.0165	-	7470A	PPB	08 Feb 2023 16:23:21	U	SMPL
CCB06 - 1	CCB06	-35	0.0850	-	7470A	PPB	08 Feb 2023 16:25:37	U	SMPL