



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB123992

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	02/08/2023	08:40
2	O1447-01		1	12	100.00	>212.0	>212.0	02/08/2023	09:10
3	O1447-01DUP		1	12	100.00	>212.0	>212.0	02/08/2023	09:40
4	O1456-01		1	14	100.00	>212.0	>212.0	02/08/2023	10:10
5	O1456-02		1	13	100.00	>212.0	>212.0	02/08/2023	10:40
6	O1467-01		1	13	100.00	>212.0	>212.0	02/08/2023	11:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6123992

WorkList Name : fp2-7

WorkList ID : 167157

Department : Wet-Chemistry

Date : 02-07-2023 08:27:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/09/2023	Water	O1447-01	Flash Point	Cool 4 deg C	PSEG03	F12	T2	02/06/2023	1010B
02/13/2023	Water	O1456-01	Flash Point	Cool 4 deg C	PSEG03	F12	34150	02/06/2023	1010B
02/13/2023	Water	O1456-02	Flash Point	Cool 4 deg C	PSEG03	F12	34151	02/06/2023	1010B
02/14/2023	Water	O1467-01	Flash Point	Cool 4 deg C	PSEG03	F21	33906	02/07/2023	1010B

Date/Time 02/08/2023 08:130
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: CWC

Date/Time 02/08/2023 11:45
 Raw Sample Received by: CWC
 Raw Sample Relinquished by: RM CWC