



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB124008

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	02/09/2023	10:30
2	O1491-02		1	12	100.00	>212.0	>212.0	02/09/2023	11:00
3	O1491-02DUP		1	12	100.00	>212.0	>212.0	02/09/2023	11:30
4	O1491-03		1	14	100.00	>212.0	>212.0	02/09/2023	12:00
5	O1503-04		1	13	100.00	>212.0	>212.0	02/09/2023	12:30
6	O1507-03		1	14	100.00	>212.0	>212.0	02/09/2023	13:00
7	O1507-04		1	12	100.00	>212.0	>212.0	02/09/2023	13:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6124008

WorkList Name : fp w 2-9

WorkList ID : 167242

Department : Wet-Chemistry

Date : 02-09-2023 08:42:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/15/2023	Water	O1491-02	Flash Point	Cool 4 deg C	PSEG03	F22	2762	02/08/2023	1010B
02/15/2023	Water	O1491-03	Flash Point	Cool 4 deg C	PSEG03	F22	OW-TOTE-0208	02/08/2023	1010B
02/15/2023	Water	O1503-04	Flash Point	Cool 4 deg C	PSEG03	G11	D2889	02/08/2023	1010B
02/13/2023	Water	O1507-03	Flash Point	Cool 4 deg C	PSEG03	G11	189-190-191	02/08/2023	1010B
02/13/2023	Water	O1507-04	Flash Point	Cool 4 deg C	PSEG03	G11	HA-WATER-0208	02/08/2023	1010B

Date/Time 02/09/2023
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: RJ C Ext-1064

Date/Time 02/09/2023
 Raw Sample Received by: RJ C Ext-1064
 Raw Sample Relinquished by: RM CWC