

LB124017


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37071



A= 0.0000e+000

B= 2.6907e-004 slope

C= -5.4864e-002 intercept

Rho= 0.9999037

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%
0.00	0.000	-0.010	-0.010	168	0.000	168					100
0.05	0.050					185					-10
0.20	0.200	0.181	-0.019	878	0.0 %	878	0				0
2.50	2.500	2.488	-0.012	9451	0.0 %	9451					1
5.00	5.000	5.033	0.033	18910	0.0 %	18910					1
7.50	7.500	7.587	0.087	28402	0.0 %	28402					1
10.0	10.000	9.920	-0.080	37071	0.0 %	37071					-1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	168	-		0 SFAM01.1	PPB	2/10/2023 13:57	Std
	0.2 S01	878	-		0.2 SFAM01.1	PPB	2/10/2023 13:59	Std
	2.5 S02	9451	-		2.5 SFAM01.1	PPB	2/10/2023 14:02	Std
	5 S03	18910	-		5 SFAM01.1	PPB	2/10/2023 14:04	Std
	7.5 S04	28402	-		7.5 SFAM01.1	PPB	2/10/2023 14:06	Std
	10 S05	37071	-		10 SFAM01.1	PPB	2/10/2023 14:09	Std
ICV042	ICV042	15586	4.1388 -		SFAM01.1	PPB	2/10/2023 14:14	SMPL
ICB042	ICB042	190	-0.0037 -		SFAM01.1	PPB	2/10/2023 14:16	SMPL
CCV015	CCV015	18626	4.9568 -		SFAM01.1	PPB	2/10/2023 14:19	SMPL
CCB015	CCB015	162	-0.0113 -		SFAM01.1	PPB	2/10/2023 14:21	SMPL
PB150783BL	PBW783	437	0.0627 -		SFAM01.1	PPB	2/10/2023 14:23	SMPL
O1304-06	MC0AX7	424	0.0592 -		SFAM01.1	PPB	2/10/2023 14:25	SMPL
O1304-07	MC0AX8	406	0.0544 -		SFAM01.1	PPB	2/10/2023 14:28	SMPL
O1304-08	MC0AX9	421	0.0584 -		SFAM01.1	PPB	2/10/2023 14:30	SMPL
O1304-09	MC0AY0	402	0.0533 -		SFAM01.1	PPB	2/10/2023 14:32	SMPL
PB150793BL	PBW793	-21	-0.0605 -		SFAM01.1	PPB	2/10/2023 14:35	SMPL
O1317-01	A44H3	10988	2.9017 -		SFAM01.1	PPB	2/10/2023 14:37	SMPL
O1317-03	A44G5	424	0.0592 -		SFAM01.1	PPB	2/10/2023 14:39	SMPL
O1317-04	A44G6	354	0.0404 -		SFAM01.1	PPB	2/10/2023 14:41	SMPL
O1317-05	A44G7	452	0.0668 -		SFAM01.1	PPB	2/10/2023 14:44	SMPL
O1317-06	A44G7D	470	0.0716 -		SFAM01.1	PPB	2/10/2023 14:46	SMPL
O1317-07	A44G7S	4288	1.0989 -		SFAM01.1	PPB	2/10/2023 14:48	SMPL
O1317-08	A44G8	512	0.0829 -		SFAM01.1	PPB	2/10/2023 14:50	SMPL
O1317-09	A44G9	415	0.0568 -		SFAM01.1	PPB	2/10/2023 14:53	SMPL
O1317-10	A44H0	418	0.0576 -		SFAM01.1	PPB	2/10/2023 14:55	SMPL
O1317-11	A44H1	432	0.0614 -		SFAM01.1	PPB	2/10/2023 14:57	SMPL
O1317-12	A44H8	406	0.0544 -		SFAM01.1	PPB	2/10/2023 14:59	SMPL
O1317-13	A44J2	427	0.06 -		SFAM01.1	PPB	2/10/2023 15:02	SMPL
O1317-14	A44J3	399	0.0525 -		SFAM01.1	PPB	2/10/2023 15:04	SMPL
PB150794BL	PBW794	413	0.0563 -		SFAM01.1	PPB	2/10/2023 15:06	SMPL
CCV016	CCV016	18344	4.8809 -		SFAM01.1	PPB	2/10/2023 15:09	SMPL
CCB016	CCB016	158	-0.0124 -		SFAM01.1	PPB	2/10/2023 15:11	SMPL
O1336-01	A44L2	4871	1.2558 -		SFAM01.1	PPB	2/10/2023 15:13	SMPL
O1336-02	A44L3	4764	1.227 -		SFAM01.1	PPB	2/10/2023 15:15	SMPL
O1336-03	A44L4	2118	0.515 -		SFAM01.1	PPB	2/10/2023 15:18	SMPL
O1336-04	A44L7	2194	0.5355 -		SFAM01.1	PPB	2/10/2023 15:20	SMPL
O1336-05	A44L8	2804	0.6996 -		SFAM01.1	PPB	2/10/2023 15:22	SMPL
O1336-06	A44N9	471	0.0719 -		SFAM01.1	PPB	2/10/2023 15:24	SMPL
O1336-07	A44P0	533	0.0885 -		SFAM01.1	PPB	2/10/2023 15:27	SMPL
O1336-08	A44P9	11047	2.9175 -		SFAM01.1	PPB	2/10/2023 15:29	SMPL
O1336-10	A44K0	434	0.0619 -		SFAM01.1	PPB	2/10/2023 15:31	SMPL
O1336-11	A44K8	2691	0.6692 -		SFAM01.1	PPB	2/10/2023 15:34	SMPL
O1336-12	A44L1	17166	4.564 -		SFAM01.1	PPB	2/10/2023 15:36	SMPL
O1336-13	A44L5	2475	0.6111 -		SFAM01.1	PPB	2/10/2023 15:38	SMPL
O1336-14	A44L5D	2489	0.6148 -		SFAM01.1	PPB	2/10/2023 15:40	SMPL
O1336-15	A44L5S	6155	1.6013 -		SFAM01.1	PPB	2/10/2023 15:43	SMPL

LB124017

INSTRUMENT ID : CV1

O1336-16	A44L6	599	0.1063 -	SFAM01.1	PPB	2/10/2023 15:45 SMPL
O1336-17	A44Q7	1853	0.4437 -	SFAM01.1	PPB	2/10/2023 15:47 SMPL
O1336-18	A44K7	4925	1.2703 -	SFAM01.1	PPB	2/10/2023 15:49 SMPL
O1336-19	A44M1	684	0.1292 -	SFAM01.1	PPB	2/10/2023 15:52 SMPL
O1336-20	A44M2	743	0.1451 -	SFAM01.1	PPB	2/10/2023 15:54 SMPL
O1336-21	A44N5	515	0.0837 -	SFAM01.1	PPB	2/10/2023 15:56 SMPL
CCV017	CCV017	18134	4.8244 -	SFAM01.1	PPB	2/10/2023 15:58 SMPL
CCB017	CCB017	140	-0.0172 -	SFAM01.1	PPB	2/10/2023 16:01 SMPL
O1336-22	A44Q6	2083	0.5056 -	SFAM01.1	PPB	2/10/2023 16:03 SMPL
CCV018	CCV018	18027	4.7956 -	SFAM01.1	PPB	2/10/2023 16:14 SMPL
CCB018	CCB018	170	-0.0091 -	SFAM01.1	PPB	2/10/2023 16:17 SMPL