



Analytical Summary Report

Analysis Method: 9040C
Parameter: Corrosivity
Run Number: LB124021

Analyst By : Iwona
Supervisor Review By : Sohil
Slope : 99.1
pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.00	02/10/2023	10:30
2	CAL2	1	Water	NA	NA	20.7	7.01	02/10/2023	10:32
3	CAL3	1	Water	NA	NA	20.6	10.03	02/10/2023	10:33
4	ICV	1	Water	NA	NA	20.6	6.98	02/10/2023	10:35
5	CCV1	1	Water	NA	NA	20.7	2.02	02/10/2023	11:20
6	O1491-02	1	Water	NA	NA	22.5	8.96	02/10/2023	11:22
7	O1491-02DUP	1	Water	NA	NA	22.4	8.95	02/10/2023	11:24
8	O1491-03	1	Water	NA	NA	22.5	8.96	02/10/2023	11:25
9	O1503-04	1	Water	NA	NA	22.4	7.48	02/10/2023	11:28
10	O1507-03	1	Water	NA	NA	22.1	7.01	02/10/2023	11:30
11	O1507-04	1	Water	NA	NA	22.7	7.00	02/10/2023	11:32
12	CCV2	1	Water	NA	NA	20.6	12.02	02/10/2023	11:34

WORKLIST(Hardcopy Internal Chain)

C3124021

WorkList Name : CORR-021023-2 WorkList ID : 167293 Department : Wet-Chemistry Date : 02-10-2023 09:42:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/08/2023	Water	O1491-02	Corrosivity	Cool 4 deg C	PSEG03	F22	2762	02/08/2023	9040C
02/08/2023	Water	O1491-03	Corrosivity	Cool 4 deg C	PSEG03	F22	OW-TOTE-0208	02/08/2023	9040C
02/08/2023	Water	O1503-04	Corrosivity	Cool 4 deg C	PSEG03	G11	D2889	02/08/2023	9040C
02/08/2023	Water	O1507-03	Corrosivity	Cool 4 deg C	PSEG03	G11	189-190-191	02/08/2023	9040C
02/08/2023	Water	O1507-04	Corrosivity	Cool 4 deg C	PSEG03	G11	HA-WATER-0208	02/08/2023	9040C

Date/Time 02/10/23 10:50
 Raw Sample Received by: 12501
 Raw Sample Relinquished by: CP SM

Date/Time 02/10/23 11:50
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: 12501