



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB124032

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	02/10/2023	11:20
2	O1487-01		1	12	100.00	>212.0	>212.0	02/10/2023	11:45
3	O1487-01DUP		1	12	100.00	>212.0	>212.0	02/10/2023	12:15
4	O1491-01		1	14	100.00	>212.0	>212.0	02/10/2023	12:45
5	O1503-03		1	13	100.00	>212.0	>212.0	02/10/2023	13:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name: FPS 2-8

WorkList ID: 167240

Department: Wet-Chemistry

Date: 02-08-2023 17:22:42

124032

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/15/2023	Solid	O1487-01	Flash Point	Cool 4 deg C	PSEG03	F22	340-341	02/08/2023	1010B
02/15/2023	Solid	O1491-01	Flash Point	Cool 4 deg C	PSEG03	F22	2742	02/08/2023	1010B
02/15/2023	Solid	O1503-03	Flash Point	Cool 4 deg C	PSEG03	G11	D2896	02/08/2023	1010B

Date/Time 02/10/2023 11:15
 Raw Sample Received by: RTH CP SW
 Raw Sample Relinquished by: CP SW

Date/Time 02/10/2023 13:50
 Raw Sample Received by: CP SW
 Raw Sample Relinquished by: RTH CWC