

LB124037


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

3950E



A= 0.0000e+000

B= 2.5274e-004 slope

C= 3.2848e-002 y-intercept

Rho= 0.9998662

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%d
0.00	0.000	0.064	0.064	125	0.000	125					
0.05	0.050					185					
0.20	0.200	0.229	0.029	776	0.0 %	776	0				
2.50	2.500	2.372	-0.128	9255	0.0 %	9255					
5.00	5.000	5.004	0.004	19670	0.0 %	19670					
7.50	7.500	7.513	0.013	29595	0.0 %	29595					
10.0	10.000	10.017	0.017	39505	0.0 %	39505					

0.00511

LB124037

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	125	-		0 SFAM01.1	PPB	2/12/2023 15:10	Std
	0.2 S01	776	-		0.2 SFAM01.1	PPB	2/12/2023 15:13	Std
	2.5 S02	9255	-		2.5 SFAM01.1	PPB	2/12/2023 15:15	Std
	5 S03	19670	-		5 SFAM01.1	PPB	2/12/2023 15:17	Std
	7.5 S04	29595	-		7.5 SFAM01.1	PPB	2/12/2023 15:19	Std
	10 S05	39505	-		10 SFAM01.1	PPB	2/12/2023 15:22	Std
ICV043	ICV043	15733	4.0093 -		SFAM01.1	PPB	2/12/2023 15:27	SMPL
ICB043	ICB043	109	0.0604 -		SFAM01.1	PPB	2/12/2023 15:29	SMPL
CCV019	CCV019	19321	4.9161 -		SFAM01.1	PPB	2/12/2023 15:31	SMPL
CCB019	CCB019	69	0.0503 -		SFAM01.1	PPB	2/12/2023 15:33	SMPL
PB150801BL	PBS801	112	0.0612 -		SFAM01.1	PPB	2/12/2023 15:36	SMPL
O1304-01	MCOAX4	942	0.2709 -		SFAM01.1	PPB	2/12/2023 15:38	SMPL
O1304-02	MCOAX5	1002	0.2861 -		SFAM01.1	PPB	2/12/2023 15:40	SMPL
O1304-03	MCOAX6	885	0.2565 -		SFAM01.1	PPB	2/12/2023 15:42	SMPL
O1304-04	MCOAX6D	919	0.2651 -		SFAM01.1	PPB	2/12/2023 15:45	SMPL
O1304-05	MCOAX6S	10519	2.6915 -		SFAM01.1	PPB	2/12/2023 15:47	SMPL
CCV020	CCV020	19214	4.8891 -		SFAM01.1	PPB	2/12/2023 15:49	SMPL
CCB020	CCB020	61	0.0483 -		SFAM01.1	PPB	2/12/2023 15:52	SMPL