

LB124089


7470A

INSTRUMENT ID: CV4

Linear

 μ Abs.:

32711



A= 0.0000e+000

B= 3.0420e-004

C= -3.0669e-002

Rho= 0.9998551

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%1)
0.0	0.000	-0.028	-0.028	10	0.000	10					-
0.2	0.200	0.226	0.026	845	0.0 %	845					13
2.5	2.500	2.432	-0.068	8095	0.0 %	8095					-3
5.0	5.000	5.072	0.072	16773	0.0 %	16773					-1
7.5	7.500	7.578	0.078	25011	0.0 %	25011					-1
10.0	10.000	9.920	-0.080	32711	0.0 %	32711					-1

LB124039 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Feb 2023 10:01:53

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc/Method	Units	Date	Type	Type
0.0 - 1	50	10	-	0.00007470A	PPB	13 Feb 2023 10:16:12	S	Std
0.2 - 1	50-2	845	-	0.20007470A	PPB	13 Feb 2023 10:18:29	S	Std
2.5 - 1	52.5	8095	-	2.50007470A	PPB	13 Feb 2023 10:23:24	S	Std
5.0 - 1	55	16773	-	5.00007470A	PPB	13 Feb 2023 10:25:41	S	Std
7.5 - 1	57.5	25011	-	7.50007470A	PPB	13 Feb 2023 10:27:57	S	Std
10.0 - 1	50	32711	-	10.00007470A	PPB	13 Feb 2023 10:30:13	S	Std
ICV44 - 1	ICV44	12750	3.8479	-7470A	PPB	13 Feb 2023 10:32:57	U	SMPL
ICB44 - 1	ICB44	18	-0.0252	-7470A	PPB	13 Feb 2023 10:35:11	U	SMPL
CCV21 - 1	CCV21	15367	4.6440	-7470A	PPB	13 Feb 2023 10:37:26	U	SMPL
CCB21 - 1	CCB21	-16	-0.0355	-7470A	PPB	13 Feb 2023 10:39:42	U	SMPL
CRA - 1	CRA	678	0.1756	-7470A	PPB	13 Feb 2023 10:41:57	U	SMPL
HighStd - 1	HighStd	32595	9.8847	-7470A	PPB	13 Feb 2023 10:44:12	U	SMPL
ChkStd - 1	ChkStd	22014	6.6660	-7470A	PPB	13 Feb 2023 10:46:27	U	SMPL
PB150799BL - 1	PBW	-44	-0.0441	-7470A	PPB	13 Feb 2023 10:48:43	U	SMPL
PB150799BS - 1	LCSW	12506	3.7737	-7470A	PPB	13 Feb 2023 10:53:55	U	SMPL
O1481-04 - 1	JPP-39-020723	-12	-0.0343	-7470A	PPB	13 Feb 2023 10:56:11	U	SMPL
O1481-08 - 1	JPP-42-020723	5	-0.0291	-7470A	PPB	13 Feb 2023 10:58:28	U	SMPL
O1481-12 - 1	JPP-46-020723	41	-0.0182	-7470A	PPB	13 Feb 2023 11:00:45	U	SMPL
O1481-16 - 1	JPP-48-020723	61	-0.0121	-7470A	PPB	13 Feb 2023 11:03:02	U	SMPL
O1485-01 - 1	EA-FILTERBAGS-0208	77	-0.0072	-7470A	PPB	13 Feb 2023 11:05:17	U	SMPL
CCV22 - 1	CCV22	15198	4.5926	-7470A	PPB	13 Feb 2023 11:07:33	U	SMPL
CCB22 - 1	CCB22	-38	-0.0422	-7470A	PPB	13 Feb 2023 11:09:48	U	SMPL
O1487-01 - 1	340-341	32	-0.0209	-7470A	PPB	13 Feb 2023 11:12:04	U	SMPL
O1487-02 - 1	342-343	41	-0.0182	-7470A	PPB	13 Feb 2023 11:14:19	U	SMPL
O1489-01 - 1	BARN-EAST-END-DRAIN	38	-0.0191	-7470A	PPB	13 Feb 2023 11:16:34	U	SMPL
O1491-01 - 1	2742	2	-0.0301	-7470A	PPB	13 Feb 2023 11:18:51	U	SMPL
O1503-01 - 1	D2880	13	-0.0267	-7470A	PPB	13 Feb 2023 11:21:07	U	SMPL
O1503-02 - 1	D2895	17	-0.0255	-7470A	PPB	13 Feb 2023 11:23:23	U	SMPL
O1503-02DUP - 1	D2895DUP	28	-0.0222	-7470A	PPB	13 Feb 2023 11:25:40	U	SMPL
O1503-02MS - 1	D2895MS	10945	3.2988	-7470A	PPB	13 Feb 2023 11:27:57	U	SMPL
O1503-02MSD - 1	D2895MSD	11185	3.3718	-7470A	PPB	13 Feb 2023 11:32:49	U	SMPL
O1503-03 - 1	D2896	338	0.0722	-7470A	PPB	13 Feb 2023 11:35:04	U	SMPL
CCV23 - 1	CCV23	16189	4.8940	-7470A	PPB	13 Feb 2023 11:37:20	U	SMPL
CCB23 - 1	CCB23	-26	-0.0386	-7470A	PPB	13 Feb 2023 11:39:36	U	SMPL
O1504-02 - 1	COMP-1	29	-0.0218	-7470A	PPB	13 Feb 2023 11:41:52	U	SMPL
O1510-04 - 1	TWM-TP14-020823	35	-0.0200	-7470A	PPB	13 Feb 2023 11:44:07	U	SMPL
O1510-08 - 1	JPP-51B-020823	44	-0.0173	-7470A	PPB	13 Feb 2023 11:46:23	U	SMPL
O1510-12 - 1	JPP-55-020823	18	-0.0252	-7470A	PPB	13 Feb 2023 11:48:39	U	SMPL
O1510-16 - 1	JPP-59-020823	4	-0.0295	-7470A	PPB	13 Feb 2023 11:50:55	U	SMPL
PB150800BL - 1	PBW	2	-0.0301	-7470A	PPB	13 Feb 2023 11:53:11	U	SMPL
PB150800BS - 1	LCSW	11003	3.3164	-7470A	PPB	13 Feb 2023 12:00:49	U	SMPL
O1515-01 - 1	B-01-GW01	6862	2.0568	-7470A	PPB	13 Feb 2023 12:08:03	U	SMPL
O1515-01DUP - 1	B-01-GW01DUP	6516	1.9515	-7470A	PPB	13 Feb 2023 12:10:20	U	SMPL
O1515-01MS - 1	B-01-GW01MS	17516	5.2977	-7470A	PPB	13 Feb 2023 12:12:35	U	SMPL
CCV24 - 1	CCV24	15208	4.5956	-7470A	PPB	13 Feb 2023 12:14:51	U	SMPL
CCB24 - 1	CCB24	-20	-0.0368	-7470A	PPB	13 Feb 2023 12:17:07	U	SMPL
O1515-01MSD - 1	B-01-GW01MSD	18106	5.4772	-7470A	PPB	13 Feb 2023 12:19:23	U	SMPL
O1527-05 - 1	34533	139	0.0116	-7470A	PPB	13 Feb 2023 12:21:38	U	SMPL
O1527-06 - 1	34534	73	-0.0085	-7470A	PPB	13 Feb 2023 12:23:54	U	SMPL
O1527-07 - 1	34535	61	-0.0121	-7470A	PPB	13 Feb 2023 12:26:09	U	SMPL
PB150766TB - 1	PB150766TB	-8	-0.0331	-7470A	PPB	13 Feb 2023 12:28:26	U	SMPL
O1503-02LX5 - 1		-33	-0.0407	-7470A	PPB	13 Feb 2023 12:30:42	U	SMPL
O1503-02A - 1		13448	4.0602	-7470A	PPB	13 Feb 2023 12:32:59	U	SMPL
O1515-01LX5 - 1		1366	0.3849	-7470A	PPB	13 Feb 2023 12:35:16	U	SMPL
O1515-01A - 1		19513	5.9052	-7470A	PPB	13 Feb 2023 12:37:33	U	SMPL
CCV25 - 1	CCV25	15991	4.8338	-7470A	PPB	13 Feb 2023 12:39:49	U	SMPL
CCB25 - 1	CCB25	-35	-0.0413	-7470A	PPB	13 Feb 2023 12:42:04	U	SMPL