

LB124055


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

31372



A= 0.0000e+000

B= 3.1987e-004

C= -9.1629e-002

Rho= 0.9998332

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	O/D
0.0	0.000	-0.090	-0.090	5	0.000	5					-
0.2	0.200	0.198	-0.002	906	0.0 %	906					-1
2.5	2.500	2.560	0.060	8291	0.0 %	8291					2
5.0	5.000	5.109	0.109	16258	0.0 %	16258					2
7.5	7.500	7.479	-0.021	23668	0.0 %	23668					0
10.0	10.000	9.943	-0.057	31372	0.0 %	31372					-1

LB124055 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Feb 2023 14:30:33

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	5	-	0.0000	7470A	PPB	13 Feb 2023 14:36:33	S	Std
0.2 - 1	30.2	906	-	0.2000	7470A	PPB	13 Feb 2023 14:38:51	S	Std
2.5 - 1	52.5	8291	-	2.5000	7470A	PPB	13 Feb 2023 14:41:23	S	Std
5.0 - 1	55	16258	-	5.0000	7470A	PPB	13 Feb 2023 14:43:40	S	Std
7.5 - 1	57.5	23668	-	7.5000	7470A	PPB	13 Feb 2023 14:50:49	S	Std
10.0 - 1	510	31372	-	10.0000	7470A	PPB	13 Feb 2023 14:53:08	S	Std
ICV45 - 1	ICV45	13350	4.1787	-	7470A	PPB	13 Feb 2023 14:55:59	U	SMPL
ICB45 - 1	ICB45	68	-0.0699	-	7470A	PPB	13 Feb 2023 15:00:49	U	SMPL
CCV26 - 1	CCV26	15807	4.9646	-	7470A	PPB	13 Feb 2023 15:03:04	U	SMPL
CCB26 - 1	CCB26	5	-0.0900	-	7470A	PPB	13 Feb 2023 15:05:19	U	SMPL
CRA - 1	CRA	860	0.1835	-	7470A	PPB	13 Feb 2023 15:07:40	U	SMPL
HighStd - 1	HighStd	30692	9.7259	-	7470A	PPB	13 Feb 2023 15:09:55	U	SMPL
ChkStd - 1	ChkStd	21860	6.9008	-	7470A	PPB	13 Feb 2023 15:12:10	U	SMPL
PB150815BL - 1	PBW	62	-0.0718	-	7470A	PPB	13 Feb 2023 15:17:16	U	SMPL
PB150815BS - 1	LCSW	13025	4.0747	-	7470A	PPB	13 Feb 2023 15:19:31	U	SMPL
O1491-02 - 1	2762	21	-0.0849	-	7470A	PPB	13 Feb 2023 15:21:48	U	SMPL
O1491-03 - 1	OW-TOTE-0208	-12	-0.0955	-	7470A	PPB	13 Feb 2023 15:24:05	U	SMPL
O1503-04 - 1	D2889	47	-0.0766	-	7470A	PPB	13 Feb 2023 15:26:22	U	SMPL
O1507-03 - 1	189-190-191	-11	-0.0951	-	7470A	PPB	13 Feb 2023 15:28:39	U	SMPL
O1507-04 - 1	HA-WATER-0208	-13	-0.0958	-	7470A	PPB	13 Feb 2023 15:30:54	U	SMPL
CCV27 - 1	CCV27	16075	5.0503	-	7470A	PPB	13 Feb 2023 15:33:10	U	SMPL
CCB27 - 1	CCB27	87	-0.0638	-	7470A	PPB	13 Feb 2023 15:35:26	U	SMPL
O1507-04DUP - 1	HA-WATER-0208DUP	7	-0.0894	-	7470A	PPB	13 Feb 2023 15:37:41	U	SMPL
O1507-04MS - 1	HA-WATER-0208MS	12672	3.9618	-	7470A	PPB	13 Feb 2023 15:39:57	U	SMPL
O1507-04MSD - 1	HA-WATER-0208MSD	12567	3.9282	-	7470A	PPB	13 Feb 2023 15:42:13	U	SMPL
O1507-04LX5 - 1		6	-0.0897	-	7470A	PPB	13 Feb 2023 15:46:45	U	SMPL
PB150798TB - 1		130	-0.0500	-	7470A	PPB	13 Feb 2023 15:50:16	U	SMPL
O1507-04A - 1		12808	4.0053	-	7470A	PPB	13 Feb 2023 15:52:49	U	SMPL
CCV28 - 1	CCV28	16519	5.1924	-	7470A	PPB	13 Feb 2023 16:01:10	U	SMPL
CCB28 - 1	CCB28	116	-0.0545	-	7470A	PPB	13 Feb 2023 16:03:27	U	SMPL