



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB124058

Analyst By : Iwona

Supervisor Review By : Sohil

Slope : 99.2

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	02/13/2023	15:20
2	CAL2	1	Water	NA	NA	20.8	7.02	02/13/2023	15:23
3	CAL3	1	Water	NA	NA	20.8	10.02	02/13/2023	15:25
4	ICV	1	Water	NA	NA	20.9	6.99	02/13/2023	15:28
5	CCV1	1	Water	NA	NA	20.9	2.01	02/13/2023	15:30
6	O1486-01	1	Water	NA	NA	22.4	4.58	02/13/2023	15:35
7	O1486-01DUP	1	Water	NA	NA	22.4	4.57	02/13/2023	15:37
8	O1524-02	1	Water	NA	NA	21.8	4.47	02/13/2023	15:39
9	O1527-05	1	Water	NA	NA	22.00	5.49	02/13/2023	15:42
10	O1527-06	1	Water	NA	NA	22.1	5.36	02/13/2023	15:44
11	O1527-07	1	Water	NA	NA	22.2	4.96	02/13/2023	15:46
12	O1544-01	1	Water	NA	NA	20.9	5.46	02/13/2023	15:48
13	CCV2	1	Water	NA	NA	20.8	12.02	02/13/2023	15:50

WORKLIST(Hardcopy Internal Chain)

LB124058

WorkList Name : PH-021023

WorkList ID : 167290

Department : Wet-Chemistry

Date : 02-10-2023 09:42:37

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/06/2023	Water	O1486-01	pH	Cool 4 deg C	STER04	F22	POOL-WATER	02/06/2023	9040C
02/10/2023	Water	O1524-02	pH	Cool 4 deg C	PSEG03	G12	34152	02/10/2023	9040C
02/10/2023	Water	O1527-05	pH	Cool 4 deg C	PSEG03	G12	34533	02/10/2023	9040C
02/10/2023	Water	O1527-06	pH	Cool 4 deg C	PSEG03	G12	34534	02/10/2023	9040C
02/10/2023	Water	O1527-07	pH	Cool 4 deg C	PSEG03	G12	34535	02/10/2023	9040C
02/07/2023	Water	O1544-01	pH	Cool 4 deg C	STER01	F51	POOL-WATER	02/07/2023	9040C

Date/Time 02/13/23 15:00
 Raw Sample Received by: 12/20/23
 Raw Sample Relinquished by: CP San

Date/Time 02/13/23 16:10
 Raw Sample Received by: CP San
 Raw Sample Relinquished by: 12/20/23