



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Iwona

Parameter: pH

Supervisor Review By : Sohil

Run Number: LB124059

Slope : 99.2

BalanceID: WC PH METER-1

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	02/13/2023	15:20
2	CAL2	1	Water	NA	NA	20.8	7.02	02/13/2023	15:23
3	CAL3	1	Water	NA	NA	20.8	10.02	02/13/2023	15:25
4	ICV	1	Water	NA	NA	20.9	6.99	02/13/2023	15:28
5	CCV1	1	Water	NA	NA	20.9	2.02	02/13/2023	15:55
6	O1489-01	1	Solid	20	20	21.4	7.73	02/13/2023	15:53
7	O1489-01DUP	1	Solid	20	20	21.5	7.72	02/13/2023	15:55
8	CCV2	1	Water	NA	NA	20.8	12.02	02/13/2023	15:58

WORKLIST(Hardcopy Internal Chain)

L B124054

WorkList Name : PH-021023-2 WorkList ID : 167291 Department : Wet-Chemistry Date : 02-10-2023 09:42:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/07/2023	Solid	O1489-01	pH	Cool 4 deg C	ATCE02	F22	BARN-EAST-END-DRAIN	02/07/2023	9045D

Date/Time 02/13/23 15:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 02/13/23 16:40
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]