



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB124067

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: FLASH POINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	02/14/2023	09:00
2	O1524-02		1	12	100.00	>212.0	>212.0	02/14/2023	09:30
3	O1524-02DUP		1	12	100.00	>212.0	>212.0	02/14/2023	10:00
4	O1527-05		1	14	100.00	>212.0	>212.0	02/14/2023	10:30
5	O1527-06		1	13	100.00	>212.0	>212.0	02/14/2023	11:00
6	O1527-07		1	14	100.00	>212.0	>212.0	02/14/2023	11:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6124067

WorkList Name : FP 2-13

WorkList ID : 167327

Department : Wet-Chemistry

Date : 02-13-2023 08:33:16

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/17/2023	Water	O1524-02	Flash Point	Cool 4 deg C	PSEG03	G12	34152	02/10/2023	1010B
02/17/2023	Water	O1527-05	Flash Point	Cool 4 deg C	PSEG03	G12	34533	02/10/2023	1010B
02/17/2023	Water	O1527-06	Flash Point	Cool 4 deg C	PSEG03	G12	34534	02/10/2023	1010B
02/17/2023	Water	O1527-07	Flash Point	Cool 4 deg C	PSEG03	G12	34535	02/10/2023	1010B

Date/Time 02/14/2023 08:16
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: CP Sm

Date/Time 02/14/2023 12:15
 Raw Sample Received by: CP Sm
 Raw Sample Relinquished by: RM CWC