

LB124077


7471B

INSTRUMENT ID: CX1

Linear

μ Abs.:

46308



A= 0.0000e+000

B= 2.1492e-004

C= -2.6469e-002

Rho= 0.9999009

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.026	-0.026	4	0.000	4					1
0.2	0.200	0.178	-0.022	953	0.0 %	953					-11
2.5	2.500	2.541	0.041	11944	0.0 %	11944					2
5.0	5.000	4.992	-0.008	23348	0.0 %	23348					0
7.5	7.500	7.589	0.089	35433	0.0 %	35433					-11
10.0	10.000	9.926	-0.074	46308	0.0 %	46308					1

LB124077 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 14 Feb 2023 15:09:25

Sample ID	Extended ID	μ Abs	Conc	Strnd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	4	-	0.0000	7471B	PPB	14 Feb 2023 15:11:33	S	Std
0.2 - 1	50.2	953	-	0.2000	7471B	PPB	14 Feb 2023 15:13:51	S	Std
2.5 - 1	52.5	11944	-	2.5000	7471B	PPB	14 Feb 2023 15:16:08	S	Std
5.0 - 1	55	23348	-	5.0000	7471B	PPB	14 Feb 2023 15:18:25	S	Std
7.5 - 1	57.5	35433	-	7.5000	7471B	PPB	14 Feb 2023 15:20:42	S	Std
10.0 - 1	510	46308	-	10.0000	7471B	PPB	14 Feb 2023 15:22:58	S	Std
ICV46 - 1	ICV46	18598	3.9707	-	7471B	PPB	14 Feb 2023 15:30:59	U	SMPL
ICB46 - 1	ICB46	-19	-0.0306	-	7471B	PPB	14 Feb 2023 15:33:14	U	SMPL
CCV29 - 1	CCV29	23070	4.9318	-	7471B	PPB	14 Feb 2023 15:35:30	U	SMPL
CCB29 - 1	CCB29	-56	-0.0385	-	7471B	PPB	14 Feb 2023 15:37:45	U	SMPL
CRA - 1	CRA	963	0.1805	-	7471B	PPB	14 Feb 2023 15:40:00	U	SMPL
HighStd - 1	HighStd	46681	10.0064	-	7471B	PPB	14 Feb 2023 15:42:15	U	SMPL
ChkStd - 1	ChkStd	33563	7.1870	-	7471B	PPB	14 Feb 2023 15:44:30	U	SMPL
PB150835BL - 1	PBS	-86	-0.0450	-	7471B	PPB	14 Feb 2023 15:46:47	U	SMPL
PB150835BS - 1	LCSS	16961	3.6188	-	7471B	PPB	14 Feb 2023 15:49:02	U	SMPL
O1526-01 - 1	PL-01-021023	1491	0.2940	-	7471B	PPB	14 Feb 2023 15:51:19	U	SMPL
O1526-01DUP - 1	PL-01-021023DUP	1366	0.2650	-	7471B	PPB	14 Feb 2023 15:53:36	U	SMPL
O1526-01MS - 1	PL-01-021023MS	19308	4.1233	-	7471B	PPB	14 Feb 2023 15:55:53	U	SMPL
O1526-01MSD - 1	PL-01-021023MSD	17761	3.7908	-	7471B	PPB	14 Feb 2023 15:58:10	U	SMPL
O1528-01 - 1	EO-03-021023	1287	0.2501	-	7471B	PPB	14 Feb 2023 16:00:26	U	SMPL
CCV30 - 1	CCV30	23038	4.9249	-	7471B	PPB	14 Feb 2023 16:02:41	U	SMPL
CCB30 - 1	CCB30	-57	-0.0387	-	7471B	PPB	14 Feb 2023 16:04:56	U	SMPL
O1529-01 - 1	SB-06-0-2.0	23639	5.0541	-	7471B	PPB	14 Feb 2023 16:07:11	U	SMPL
O1529-02 - 1	SB-06-5.0-7.0	30776	6.5880	-	7471B	PPB	14 Feb 2023 16:09:27	U	SMPL
O1529-04 - 1	SB-01-0-2.0	5720	1.2029	-	7471B	PPB	14 Feb 2023 16:11:42	U	SMPL
O1529-05 - 1	SB-01-5.0-7.0	12172	2.5896	-	7471B	PPB	14 Feb 2023 16:13:58	U	SMPL
O1529-07 - 1	SB-05-0-2.0	338850	72.8003	-	7471B	PPB	14 Feb 2023 16:16:14	U	SMPL
O1529-08 - 1	SB-05-5.0-7.0	41013	8.7882	-	7471B	PPB	14 Feb 2023 16:18:31	U	SMPL
O1537-01 - 1	TEST-PIT-1-(TP1)-COMP-1	3326	0.6884	-	7471B	PPB	14 Feb 2023 16:20:51	U	SMPL
O1537-09 - 1	TEST-PIT-2-(TP2)-COMP-2	5421	1.1386	-	7471B	PPB	14 Feb 2023 16:23:09	U	SMPL
O1537-16 - 1	TEST-PIT-6-(TP6)-COMP-3	3409	0.7062	-	7471B	PPB	14 Feb 2023 16:25:26	U	SMPL
O1538-01 - 1	TEST-PIT-10-(TP10)-COMP-4	5641	1.1859	-	7471B	PPB	14 Feb 2023 16:27:42	U	SMPL
CCV31 - 1	CCV31	22932	4.9022	-	7471B	PPB	14 Feb 2023 16:29:57	U	SMPL
CCB31 - 1	CCB31	-85	-0.0447	-	7471B	PPB	14 Feb 2023 16:32:13	U	SMPL
O1538-08 - 1	TEST-PIT-5-(TP5)-COMP-5	6545	1.3802	-	7471B	PPB	14 Feb 2023 16:34:29	U	SMPL
O1538-15 - 1	STOCKPILE-COMP-6	157	0.0073	-	7471B	PPB	14 Feb 2023 16:36:44	U	SMPL
O1539-01 - 1	COMP-7	132	0.0019	-	7471B	PPB	14 Feb 2023 16:39:00	U	SMPL
O1539-09 - 1	COMP-8	645	0.1122	-	7471B	PPB	14 Feb 2023 16:41:16	U	SMPL
O1541-02 - 1	ASH	13597	2.8958	-	7471B	PPB	14 Feb 2023 16:43:32	U	SMPL
O1542-01 - 1	SOIL	-56	-0.0385	-	7471B	PPB	14 Feb 2023 16:45:48	U	SMPL
O1526-01LX5 - 1		262	0.0298	-	7471B	PPB	14 Feb 2023 16:48:05	U	SMPL
O1526-01A - 1		19809	4.2309	-	7471B	PPB	14 Feb 2023 16:50:22	U	SMPL
O1529-07DLX10 - 1		40839	8.7508	-	7471B	PPB	14 Feb 2023 16:52:40	U	SMPL
CCV32 - 1	CCV32	22863	4.8873	-	7471B	PPB	14 Feb 2023 16:54:56	U	SMPL
CCB32 - 1	CCB32	-213	-0.0722	-	7471B	PPB	14 Feb 2023 16:57:11	U	SMPL