

L8124089


245.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

30944



A= 0.0000e+000

B= 3.1865e-004

C= -1.6897e-002

Rho= 0.9995523

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	(%)
0.0	0.000	0.018	0.018	109	0.000	109					
0.05	0.050	0.042	-0.008	184	0.0 %	184	0				~16
0.2	0.200	0.209	0.009	708	0.0 %	708					4
2.5	2.500	2.358	-0.142	7454	0.0 %	7454					~6
5.0	5.000	5.107	0.107	16080	0.0 %	16080					2
7.5	7.500	7.673	0.173	24132	0.0 %	24132					2
10.0	10.000	9.843	-0.157	30944	0.0 %	30944					-2

LB124089 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 15 Feb 2023 12:23:04

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	109	-	0.0000	245.1	PPB	15 Feb 2023 13:08:51	S	Std
0.05 - 1	50.05	184	-	0.0500	245.1	PPB	15 Feb 2023 13:16:21	S	Std
0.2 - 1	50.2	708	-	0.2000	245.1	PPB	15 Feb 2023 13:18:38	S	Std
2.5 - 1	50.5	7454	-	2.5000	245.1	PPB	15 Feb 2023 13:20:55	S	Std
5.0 - 1	55	16080	-	5.0000	245.1	PPB	15 Feb 2023 13:23:12	S	Std
7.5 - 1	57.5	24132	-	7.5000	245.1	PPB	15 Feb 2023 13:25:27	S	Std
10.0 - 1	510	30944	-	10.0000	245.1	PPB	15 Feb 2023 13:47:33	S	Std
ICV47 - 1	ICV47	12540	3.9790	-245.1		PPB	15 Feb 2023 13:57:34	U	SMPL
ICB47 - 1	ICB47	31	-0.0070	-245.1		PPB	15 Feb 2023 13:59:49	U	SMPL
CCV33 - 1	CCV33	15480	4.9159	-245.1		PPB	15 Feb 2023 14:02:05	U	SMPL
CCB33 - 1	CCB33	24	-0.0092	-245.1		PPB	15 Feb 2023 14:04:20	U	SMPL
CRA - 1	CRA	167	0.0363	-245.1		PPB	15 Feb 2023 14:09:11	U	SMPL
HighStd - 1	HighStd	30252	9.6230	-245.1		PPB	15 Feb 2023 14:11:26	U	SMPL
ChkStd - 1	ChkStd	21490	6.8310	-245.1		PPB	15 Feb 2023 14:13:41	U	SMPL
PB150865BL - 1	PBW	-11	-0.0204	-245.1		PPB	15 Feb 2023 14:15:59	U	SMPL
PB150865BS - 1	LCSW	11443	3.6294	-245.1		PPB	15 Feb 2023 14:18:16	U	SMPL
O1493-01 - 1	402	117	0.0204	-245.1		PPB	15 Feb 2023 14:20:32	U	SMPL
O1535-07 - 1	14B-(1-4)-COMP	1330	0.4069	-245.1		PPB	15 Feb 2023 14:22:49	U	SMPL
O1547-07 - 1	14B-(1-4)-COMP	1560	0.4802	-245.1		PPB	15 Feb 2023 14:25:06	U	SMPL
O1553-05 - 1	TWP-06	864	0.2584	-245.1		PPB	15 Feb 2023 14:27:23	U	SMPL
O1553-05DUP - 1	TWP-06DUP	834	0.2489	-245.1		PPB	15 Feb 2023 14:29:38	U	SMPL
CCV34 - 1	CCV34	14654	4.6526	-245.1		PPB	15 Feb 2023 14:31:54	U	SMPL
CCB34 - 1	CCB34	-21	-0.0236	-245.1		PPB	15 Feb 2023 14:34:10	U	SMPL
O1553-05MS - 1	TWP-06MS	12422	3.9414	-245.1		PPB	15 Feb 2023 14:36:28	U	SMPL
O1553-05MSD - 1	TWP-06MSD	11918	3.7808	-245.1		PPB	15 Feb 2023 14:38:44	U	SMPL
O1553-05LX5 - 1		176	0.0392	-245.1		PPB	15 Feb 2023 14:41:00	U	SMPL
O1553-05A - 1		12017	3.8124	-245.1		PPB	15 Feb 2023 14:43:17	U	SMPL
CCV35 - 1	CCV35	15156	4.8126	-245.1		PPB	15 Feb 2023 14:45:34	U	SMPL
CCB35 - 1	CCB35	-67	-0.0382	-245.1		PPB	15 Feb 2023 14:47:51	U	SMPL