

LB124098


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

29350



A= 0.0000e+000

B= 3.4083e-004

C= 3.8046e-002

Rho= 0.9998903

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9/10
0.0	0.000	0.048	0.048	29	0.000	29					1
0.2	0.200	0.249	0.049	620	0.0 %	620					25
2.5	2.500	2.404	-0.096	6942	0.0 %	6942					4
5.0	5.000	4.952	-0.048	14419	0.0 %	14419					1
7.5	7.500	7.505	0.005	21907	0.0 %	21907					0
10.0	10.000	10.041	0.041	29350	0.0 %	29350					0

LB124098 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Feb 2023 15:42:38

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	5.2	29	-	0.0000	7470A	PPB	15 Feb 2023 15:43:00	S	Std
0.2 - 1	50.2	620	-	0.2000	7470A	PPB	15 Feb 2023 15:45:17	S	Std
2.5 - 1	52.5	6942	-	2.5000	7470A	PPB	15 Feb 2023 15:47:35	S	Std
5.0 - 1	55	14419	-	5.0000	7470A	PPB	15 Feb 2023 15:49:52	S	Std
7.5 - 1	55.5	21907	-	7.5000	7470A	PPB	15 Feb 2023 15:57:44	S	Std
10.0 - 1	57.5	29350	-	10.0000	7470A	PPB	15 Feb 2023 16:00:02	S	Std
ICV48 - 1	ICV48	11402	3.9242	-	7470A	PPB	15 Feb 2023 16:03:47	U	SMPL
ICB48 - 1	ICB48	-91	0.0070	-	7470A	PPB	15 Feb 2023 16:08:02	U	SMPL
CCV36 - 1	CCV36	13512	4.6434	-	7470A	PPB	15 Feb 2023 16:08:19	U	SMPL
CCB36 - 1	CCB36	-38	0.0251	-	7470A	PPB	15 Feb 2023 16:10:35	U	SMPL
CRA - 1	CRA	591	0.2395	-	7470A	PPB	15 Feb 2023 16:12:53	U	SMPL
HighStd - 1	HighStd	33212	11.3578	-	7470A	PPB	15 Feb 2023 16:23:22	U	SMPL
ChkStd - 1	ChkStd	20106	6.8908	-	7470A	PPB	15 Feb 2023 16:25:38	U	SMPL
PB150851BL - 1	PBW	-137	-0.0086	-	7470A	PPB	15 Feb 2023 16:27:59	U	SMPL
PB150851BS - 1	LCSW	10805	3.7207	-	7470A	PPB	15 Feb 2023 16:30:15	U	SMPL
O1432-08 - 1	DOMINICK-STREET-UTILITY-RELOCATION-EXCAVATION	-37	0.0254	-	7470A	PPB	15 Feb 2023 16:32:31	U	SMPL
O1529-03 - 1	SB-06-COMP	5	0.0398	-	7470A	PPB	15 Feb 2023 16:34:49	U	SMPL
O1529-06 - 1	SB-01-COMP	0	0.0380	-	7470A	PPB	15 Feb 2023 16:37:06	U	SMPL
O1529-09 - 1	SB-05-COMP	19	0.0445	-	7470A	PPB	15 Feb 2023 16:39:23	U	SMPL
O1541-01 - 1	ASH	-2	0.0374	-	7470A	PPB	15 Feb 2023 16:41:39	U	SMPL
CCV37 - 1	CCV37	13921	4.7828	-	7470A	PPB	15 Feb 2023 16:43:54	U	SMPL
CCB37 - 1	CCB37	-141	-0.0100	-	7470A	PPB	15 Feb 2023 16:46:10	U	SMPL
O1541-01DUP - 1	ASHDUP	12	0.0421	-	7470A	PPB	15 Feb 2023 16:48:27	U	SMPL
O1541-01MS - 1	ASHMS	12342	4.2446	-	7470A	PPB	15 Feb 2023 16:50:43	U	SMPL
O1541-01MSD - 1	ASHMSD	11967	4.1168	-	7470A	PPB	15 Feb 2023 16:52:59	U	SMPL
PB150826TB - 1		-202	-0.0308	-	7470A	PPB	15 Feb 2023 16:55:16	U	SMPL
O1541-01LX5 - 1		5	0.0398	-	7470A	PPB	15 Feb 2023 16:57:32	U	SMPL
O1541-01A - 1		11007	3.7896	-	7470A	PPB	15 Feb 2023 16:59:49	U	SMPL
CCV38 - 1	CCV38	14544	4.9951	-	7470A	PPB	15 Feb 2023 17:04:45	U	SMPL
CCB38 - 1	CCB38	-45	0.0227	-	7470A	PPB	15 Feb 2023 17:07:04	U	SMPL