



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB124100

Analyst By : Iwona

Supervisor Review By : Sohil

Slope : 99.2

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.02	02/15/2023	16:00
2	CAL2	1	Water	NA	NA	20.8	7.0	02/15/2023	16:01
3	CAL3	1	Water	NA	NA	20.9	10.02	02/15/2023	16:02
4	ICV	1	Water	NA	NA	20.9	7.01	02/15/2023	16:04
5	CCV1	1	Water	NA	NA	20.9	2.02	02/15/2023	16:05
6	O1545-01	1	Water	NA	NA	20.4	7.65	02/15/2023	16:07
7	O1545-01DUP	1	Water	NA	NA	20.5	7.66	02/15/2023	16:09
8	O1546-01	1	Water	NA	NA	21.6	7.01	02/15/2023	16:11
9	O1546-02	1	Water	NA	NA	21.8	7.61	02/15/2023	16:14
10	O1549-11	1	Water	NA	NA	21.8	6.96	02/15/2023	16:16
11	O1572-07	1	Water	NA	NA	20.2	7.08	02/15/2023	16:20
12	CCV2	1	Water	NA	NA	20.9	12.03	02/15/2023	16:23

WORKLIST(Hardcopy Internal Chain)

LB124100

WorkList Name : PH-021523

WorkList ID : 167409

Department : Wet-Chemistry

Date : 02-15-2023 15:38:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Collect Date	Method
						Storage	Customer Sample		
02/13/2023	Water	O1545-01	pH	Cool 4 deg C	ATCE02	G22	MH-M	02/13/2023	9040C
02/14/2023	Water	O1546-01	pH	Cool 4 deg C	PSEG03	G22	33910	02/14/2023	9040C
02/14/2023	Water	O1546-02	pH	Cool 4 deg C	PSEG03	G22	40033	02/14/2023	9040C
02/14/2023	Water	O1549-11	pH	Cool 4 deg C	PSEG03	G22	1929-1930	02/14/2023	9040C
02/15/2023	Water	O1572-07	pH	Cool 4 deg C	PSEG03	G32	TOTE	02/15/2023	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

02/15/23 15:45
 12/20/23
 CP 8hr

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

02/15/23 16:30
 CP 8hr
 12/20/23