



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB124101

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	02/15/2023	11:10
2	O1545-01		1	12	100.00	>212.0	>212.0	02/15/2023	11:40
3	O1545-01DUP		1	14	100.00	>212.0	>212.0	02/15/2023	12:10
4	O1546-01		1	12	100.00	>212.0	>212.0	02/15/2023	12:40
5	O1546-02		1	14	100.00	>212.0	>212.0	02/15/2023	13:10
6	O1549-06		1	12	100.00	>212.0	>212.0	02/15/2023	13:40
7	O1549-07		1	12	100.00	>212.0	>212.0	02/15/2023	14:10
8	O1549-08		1	14	100.00	>212.0	>212.0	02/15/2023	14:40
9	O1549-09		1	14	100.00	>212.0	>212.0	02/15/2023	15:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6124101

WorkList Name: FP2-15

WorkList ID: 167405

Department: Wet-Chemistry

Date: 02-15-2023 11:01:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/16/2023	Water	O1545-01	Flash Point	Cool 4 deg C	ATCE02	G22	MH-M	02/13/2023	1010B
02/21/2023	Water	O1546-01	Flash Point	Cool 4 deg C	PSEG03	G22	33910	02/14/2023	1010B
02/21/2023	Water	O1546-02	Flash Point	Cool 4 deg C	PSEG03	G22	40033	02/14/2023	1010B
02/21/2023	Water	O1549-06	Flash Point	Cool 4 deg C	PSEG03	G22	40189	02/14/2023	1010B
02/21/2023	Water	O1549-07	Flash Point	Cool 4 deg C	PSEG03	G22	1831	02/14/2023	1010B
02/21/2023	Water	O1549-08	Flash Point	Cool 4 deg C	PSEG03	G22	1696	02/14/2023	1010B
02/21/2023	Water	O1549-09	Flash Point	Cool 4 deg C	PSEG03	G22	1850	02/14/2023	1010B

Date/Time 02/15/2023 10:45
 Raw Sample Received by: RM CWCJ
 Raw Sample Relinquished by: RS CEXT-1006J

Date/Time 02/15/2023 15:45
 Raw Sample Received by: RS CEXT-1006J
 Raw Sample Relinquished by: RM CWCJ