


LB124107

7470A

INSTRUMENT ID: CU1

Linear



A= 0.0000e+000

B= 3.2788e-004

C= -5.7637e-002

Rho= 0.9995588

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	6/10
0.0	0.000	0.007	0.007	197	0.000	197					-
0.2	0.200	0.195	-0.005	772	0.0 %	772					-3
2.5	2.500	2.488	-0.012	7764	0.0 %	7764					0
5.0	5.000	4.917	-0.083	15172	0.0 %	15172					-2
7.5	7.500	7.720	0.220	23722	0.0 %	23722					3
10.0	10.000	9.872	-0.128	30284	0.0 %	30284					-1

LB124107 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 16 Feb 2023 11:11:07

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	197	-	0.0000	7470A	PPB	16 Feb 2023 11:20:13	S	Std
0.2 - 1	51.2	772	-	0.2000	7470A	PPB	16 Feb 2023 11:22:30	S	Std
2.5 - 1	52.5	7764	-	2.5000	7470A	PPB	16 Feb 2023 11:24:48	S	Std
5.0 - 1	53	15172	-	5.0000	7470A	PPB	16 Feb 2023 11:27:05	S	Std
7.5 - 1	53.5	23722	-	7.5000	7470A	PPB	16 Feb 2023 11:31:56	S	Std
10.0 - 1	54	30284	-	10.0000	7470A	PPB	16 Feb 2023 11:42:29	S	Std
ICV49 - 1	ICV49	12634	4.0849	-	7470A	PPB	16 Feb 2023 11:48:29	U	SMPL
ICB49 - 1	ICB49	219	0.0142	-	7470A	PPB	16 Feb 2023 11:50:44	U	SMPL
CCV39 - 1	CCV39	16008	5.1911	-	7470A	PPB	16 Feb 2023 11:52:59	U	SMPL
CCB39 - 1	CCB39	178	0.0007	-	7470A	PPB	16 Feb 2023 11:55:14	U	SMPL
CRA - 1	CRA	764	0.1929	-	7470A	PPB	16 Feb 2023 11:57:29	U	SMPL
HighStd - 1	HighStd	30132	9.8222	-	7470A	PPB	16 Feb 2023 11:58:44	U	SMPL
ChkStd - 1	ChkStd	23434	7.6260	-	7470A	PPB	16 Feb 2023 12:06:17	U	SMPL
PB150852BL - 1	PBW	-64	-0.0786	-	7470A	PPB	16 Feb 2023 12:08:32	U	SMPL
PB150852BS - 1	LCSW	11950	3.8606	-	7470A	PPB	16 Feb 2023 12:10:48	U	SMPL
O1545-01 - 1	MH-M	-29	-0.0671	-	7470A	PPB	16 Feb 2023 12:13:06	U	SMPL
O1545-01DUP - 1	MH-MDUP	-3	-0.0586	-	7470A	PPB	16 Feb 2023 12:15:22	U	SMPL
O1545-01MS - 1	MH-MMS	10454	3.3701	-	7470A	PPB	16 Feb 2023 12:17:39	U	SMPL
O1545-01MSD - 1	MH-MMSD	12200	3.9426	-	7470A	PPB	16 Feb 2023 12:19:56	U	SMPL
PB150857BL - 1	PBW	-33	-0.0685	-	7470A	PPB	16 Feb 2023 12:22:12	U	SMPL
CCV40 - 1	CCV40	15897	5.1547	-	7470A	PPB	16 Feb 2023 12:24:27	U	SMPL
CCB40 - 1	CCB40	185	0.0030	-	7470A	PPB	16 Feb 2023 12:26:42	U	SMPL
PB150857BS - 1	LCSW	10270	3.3097	-	7470A	PPB	16 Feb 2023 12:28:58	U	SMPL
O1549-09 - 1	1850	116	-0.0196	-	7470A	PPB	16 Feb 2023 12:31:13	U	SMPL
O1549-10 - 1	1737	117	-0.0193	-	7470A	PPB	16 Feb 2023 12:33:29	U	SMPL
O1549-11 - 1	1929-1930	148	-0.0091	-	7470A	PPB	16 Feb 2023 12:35:45	U	SMPL
O1553-01 - 1	TWP-01	241	0.0214	-	7470A	PPB	16 Feb 2023 12:38:01	U	SMPL
O1553-01DUP - 1	TWP-01DUP	192	0.0053	-	7470A	PPB	16 Feb 2023 12:40:17	U	SMPL
O1553-01MS - 1	TWP-01MS	12187	3.9383	-	7470A	PPB	16 Feb 2023 12:42:34	U	SMPL
O1553-01MSD - 1	TWP-01MSD	12254	3.9603	-	7470A	PPB	16 Feb 2023 12:44:52	U	SMPL
O1553-03 - 1	TWP-01-DUP	132	-0.0144	-	7470A	PPB	16 Feb 2023 12:47:09	U	SMPL
PB150885BL - 1	PBW	116	-0.0196	-	7470A	PPB	16 Feb 2023 12:49:25	U	SMPL
CCV41 - 1	CCV41	15922	5.1629	-	7470A	PPB	16 Feb 2023 12:51:41	U	SMPL
CCB41 - 1	CCB41	-47	-0.0730	-	7470A	PPB	16 Feb 2023 12:53:56	U	SMPL
O1537-02 - 1	TEST-PIT-1-(TP1)-COMP-1	63	-0.0370	-	7470A	PPB	16 Feb 2023 12:58:27	U	SMPL
O1537-10 - 1	TEST-PIT-2-(TP2)-COMP-2	54	-0.0398	-	7470A	PPB	16 Feb 2023 13:00:43	U	SMPL
O1537-17 - 1	TEST-PIT-6-(TP6)-COMP-3	134	-0.0137	-	7470A	PPB	16 Feb 2023 13:02:58	U	SMPL
O1538-02 - 1	TEST-PIT-10-(TP10)-COMP-4	134	-0.0137	-	7470A	PPB	16 Feb 2023 13:05:14	U	SMPL
O1538-09 - 1	TEST-PIT-5-(TP5)-COMP-5	81	-0.0311	-	7470A	PPB	16 Feb 2023 13:07:31	U	SMPL
O1538-16 - 1	STOCKPILE-COMP-6	138	-0.0124	-	7470A	PPB	16 Feb 2023 13:09:48	U	SMPL
O1539-02 - 1	COMP-7	127	-0.0160	-	7470A	PPB	16 Feb 2023 13:12:05	U	SMPL
O1539-10 - 1	COMP-8	0	-0.0576	-	7470A	PPB	16 Feb 2023 13:14:22	U	SMPL
O1539-10 - 1	COMP-8	141	-0.0114	-	7470A	PPB	16 Feb 2023 13:16:59	U	SMPL
O1539-10DUP - 1	COMP-8DUP	154	-0.0071	-	7470A	PPB	16 Feb 2023 13:19:14	U	SMPL
CCV42 - 1	CCV42	15095	4.8918	-	7470A	PPB	16 Feb 2023 13:21:30	U	SMPL
CCB42 - 1	CCB42	-33	-0.0685	-	7470A	PPB	16 Feb 2023 13:23:46	U	SMPL
O1539-10MS - 1	COMP-8MS	12338	3.9878	-	7470A	PPB	16 Feb 2023 13:28:43	U	SMPL
O1539-10MSD - 1	COMP-8MSD	10886	3.5117	-	7470A	PPB	16 Feb 2023 13:30:59	U	SMPL
PB150843TB - 1	PB150843TB	160	-0.0052	-	7470A	PPB	16 Feb 2023 13:33:14	U	SMPL
PB150847TB - 1	PB150847TB	254	0.0256	-	7470A	PPB	16 Feb 2023 13:35:30	U	SMPL
O1545-01LX5 - 1		1	-0.0573	-	7470A	PPB	16 Feb 2023 13:37:46	U	SMPL
O1545-01A - 1		11983	3.8714	-	7470A	PPB	16 Feb 2023 13:40:03	U	SMPL
O1553-01LX5 - 1		11	-0.0540	-	7470A	PPB	16 Feb 2023 13:42:19	U	SMPL
O1553-01A - 1		12893	4.1698	-	7470A	PPB	16 Feb 2023 13:44:36	U	SMPL
O1539-10LX5 - 1		-10	-0.0609	-	7470A	PPB	16 Feb 2023 13:46:53	U	SMPL
O1539-10A - 1		12384	4.0029	-	7470A	PPB	16 Feb 2023 13:49:09	U	SMPL
CCV43 - 1	CCV43	15985	5.1836	-	7470A	PPB	16 Feb 2023 13:51:24	U	SMPL
CCB43 - 1	CCB43	95	-0.0265	-	7470A	PPB	16 Feb 2023 13:53:40	U	SMPL
PB150885BS - 1	LCSW	13006	4.2068	-	7470A	PPB	16 Feb 2023 13:55:56	U	SMPL
CCV44 - 1	CCV44	14831	4.8052	-	7470A	PPB	16 Feb 2023 13:58:12	U	SMPL
CCB44 - 1	CCB44	147	-0.0094	-	7470A	PPB	16 Feb 2023 14:00:28	U	SMPL

2/16/23